

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021622\
 Data File : BG052408.D
 Acq On : 16 Feb 2022 10:02
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 16 12:43:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	0.00
2	1,4-Dioxane	40.000	39.877	0.3	96	0.00
3	Pyridine	40.000	39.626	0.9	89	0.00
4	n-Nitrosodimethylamine	40.000	37.059	7.4	86	0.00
5 S	2-Fluorophenol	80.000	78.982	1.3	91	0.00
6	Aniline	40.000	37.720	5.7	84	0.00
7 S	Phenol-d6	80.000	75.888	5.1	85	0.00
8	2-Chlorophenol	40.000	37.998	5.0	89	0.00
9	Benzaldehyde	40.000	38.241	4.4	86	0.00
10 C	Phenol	40.000	38.399	4.0	87	0.00
11	bis(2-Chloroethyl)ether	40.000	36.433	8.9	86	0.00
12	1,3-Dichlorobenzene	40.000	38.451	3.9	89	0.00
13 C	1,4-Dichlorobenzene	40.000	38.648	3.4	89	0.00
14	1,2-Dichlorobenzene	40.000	36.994	7.5	86	0.00
15	Benzyl Alcohol	40.000	37.514	6.2	82	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.386	9.0	86	0.01
17	2-Methylphenol	40.000	38.448	3.9	87	0.00
18	Hexachloroethane	40.000	39.640	0.9	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.697	10.8	82	0.00
20	3+4-Methylphenols	40.000	37.914	5.2	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	37.619	6.0	84	0.00
23 S	Nitrobenzene-d5	80.000	75.841	5.2	84	0.00
24	Nitrobenzene	40.000	38.243	4.4	83	0.00
25	Isophorone	40.000	37.695	5.8	83	0.00
26 C	2-Nitrophenol	40.000	39.896	0.3	85	0.00
27	2,4-Dimethylphenol	40.000	39.646	0.9	86	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.182	4.5	84	0.00
29 C	2,4-Dichlorophenol	40.000	38.643	3.4	85	0.00
30	1,2,4-Trichlorobenzene	40.000	38.354	4.1	86	0.00
31	Naphthalene	40.000	38.762	3.1	86	0.00
32	Benzoic acid	40.000	34.102	14.7	76	0.01
33	4-Chloroaniline	40.000	39.056	2.4	84	0.00
34 C	Hexachlorobutadiene	40.000	38.448	3.9	87	0.00
35	Caprolactam	40.000	40.021	-0.1	86	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.620	3.5	83	0.00
37	2-Methylnaphthalene	40.000	38.878	2.8	87	0.00
38	1-Methylnaphthalene	40.000	38.908	2.7	85	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.338	1.7	85	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.816	8.0	81	0.00
42 S	2,4,6-Tribromophenol	80.000	82.658	-3.3	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.432	1.4	83	0.00
44	2,4,5-Trichlorophenol	40.000	39.767	0.6	85	0.00
45 S	2-Fluorobiphenyl	80.000	79.473	0.7	86	0.00
46	1,1'-Biphenyl	40.000	39.761	0.6	86	0.00
47	2-Chloronaphthalene	40.000	40.308	-0.8	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021622\
 Data File : BG052408.D
 Acq On : 16 Feb 2022 10:02
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 16 12:43:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.045	-2.6	87	0.00
49	Acenaphthylene	40.000	40.076	-0.2	87	0.00
50	Dimethylphthalate	40.000	39.421	1.4	86	0.00
51	2,6-Dinitrotoluene	40.000	40.901	-2.3	87	0.00
52 C	Acenaphthene	40.000	37.393	6.5	80	0.00
53	3-Nitroaniline	40.000	41.918	-4.8	88	0.00
54 P	2,4-Dinitrophenol	40.000	37.195	7.0	80	0.00
55	Dibenzofuran	40.000	40.036	-0.1	88	0.00
56 P	4-Nitrophenol	40.000	40.770	-1.9	84	0.00
57	2,4-Dinitrotoluene	40.000	41.019	-2.5	86	0.00
58	Fluorene	40.000	40.851	-2.1	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.204	-0.5	86	0.00
60	Diethylphthalate	40.000	40.859	-2.1	88	0.00
61	4-Chlorophenyl-phenylether	40.000	40.009	-0.0	88	0.00
62	4-Nitroaniline	40.000	43.840	-9.6	92	0.00
63	Azobenzene	40.000	40.048	-0.1	87	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	92	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.082	2.3	86	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.315	4.2	88	0.00
67	4-Bromophenyl-phenylether	40.000	38.156	4.6	88	0.00
68	Hexachlorobenzene	40.000	36.780	8.0	86	0.00
69	Atrazine	40.000	39.030	2.4	89	0.00
70 C	Pentachlorophenol	40.000	37.021	7.4	82	0.00
71	Phenanthrene	40.000	38.973	2.6	91	0.00
72	Anthracene	40.000	39.202	2.0	91	0.00
73	Carbazole	40.000	39.643	0.9	93	0.00
74	Di-n-butylphthalate	40.000	40.971	-2.4	95	0.00
75 C	Fluoranthene	40.000	39.922	0.2	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	43.277	-8.2	93	0.00
78	Pyrene	40.000	39.177	2.1	95	0.00
79 S	Terphenyl-d14	80.000	76.546	4.3	95	0.00
80	Butylbenzylphthalate	40.000	41.118	-2.8	98	0.00
81	Benzo(a)anthracene	40.000	39.133	2.2	96	0.00
82	3,3'-Dichlorobenzidine	40.000	40.076	-0.2	96	0.00
83	Chrysene	40.000	38.800	3.0	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.381	-3.5	97	-0.01
85 c	Di-n-octyl phthalate	40.000	41.698	-4.2	99	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.750	0.6	98	0.00
88	Benzo(b)fluoranthene	40.000	39.198	2.0	97	0.00
89	Benzo(k)fluoranthene	40.000	40.109	-0.3	98	0.00
90 C	Benzo(a)pyrene	40.000	39.857	0.4	98	0.00
91	Dibenzo(a,h)anthracene	40.000	39.524	1.2	98	-0.01
92	Benzo(g,h,i)perylene	40.000	39.950	0.1	99	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021622\
Data File : BG052408.D
Acq On : 16 Feb 2022 10:02
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 16 12:43:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 08 01:21:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0