

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021622\
 Data File : BG052412.D
 Acq On : 16 Feb 2022 18:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 17 00:42:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	84	0.00
2	1,4-Dioxane	40.000	40.227	-0.6	90	0.00
3	Pyridine	40.000	39.931	0.2	83	0.00
4	n-Nitrosodimethylamine	40.000	36.870	7.8	79	0.00
5 S	2-Fluorophenol	80.000	80.559	-0.7	86	0.00
6	Aniline	40.000	38.041	4.9	78	0.00
7 S	Phenol-d6	80.000	77.026	3.7	80	0.00
8	2-Chlorophenol	40.000	38.932	2.7	85	0.00
9	Benzaldehyde	40.000	36.495	8.8	77	0.00
10 C	Phenol	40.000	39.143	2.1	82	0.00
11	bis(2-Chloroethyl)ether	40.000	37.675	5.8	82	0.00
12	1,3-Dichlorobenzene	40.000	39.694	0.8	85	0.00
13 C	1,4-Dichlorobenzene	40.000	38.585	3.5	83	0.00
14	1,2-Dichlorobenzene	40.000	37.975	5.1	82	0.00
15	Benzyl Alcohol	40.000	38.204	4.5	78	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.848	7.9	81	0.00
17	2-Methylphenol	40.000	38.554	3.6	81	0.00
18	Hexachloroethane	40.000	39.745	0.6	88	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.934	7.7	79	0.00
20	3+4-Methylphenols	40.000	38.986	2.5	81	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	83	0.00
22	Acetophenone	40.000	38.200	4.5	81	0.00
23 S	Nitrobenzene-d5	80.000	76.650	4.2	81	0.00
24	Nitrobenzene	40.000	37.588	6.0	78	0.00
25	Isophorone	40.000	38.773	3.1	81	0.00
26 C	2-Nitrophenol	40.000	40.887	-2.2	83	0.00
27	2,4-Dimethylphenol	40.000	39.540	1.2	82	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.911	2.7	82	0.00
29 C	2,4-Dichlorophenol	40.000	38.958	2.6	82	0.00
30	1,2,4-Trichlorobenzene	40.000	38.262	4.3	82	0.00
31	Naphthalene	40.000	38.354	4.1	81	0.00
32	Benzoic acid	40.000	32.972	17.6	70	0.00
33	4-Chloroaniline	40.000	39.871	0.3	82	0.00
34 C	Hexachlorobutadiene	40.000	38.226	4.4	83	0.00
35	Caprolactam	40.000	39.920	0.2	82	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.331	-0.8	82	0.00
37	2-Methylnaphthalene	40.000	38.908	2.7	83	0.00
38	1-Methylnaphthalene	40.000	38.748	3.1	80	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.092	7.3	81	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.049	14.9	74	0.00
42 S	2,4,6-Tribromophenol	80.000	78.838	1.5	84	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.765	5.6	80	0.00
44	2,4,5-Trichlorophenol	40.000	37.289	6.8	80	0.00
45 S	2-Fluorobiphenyl	80.000	74.989	6.3	81	0.00
46	1,1'-Biphenyl	40.000	37.618	6.0	82	0.00
47	2-Chloronaphthalene	40.000	37.619	6.0	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021622\
 Data File : BG052412.D
 Acq On : 16 Feb 2022 18:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 17 00:42:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.747	5.6	80	0.00
49	Acenaphthylene	40.000	38.183	4.5	83	0.00
50	Dimethylphthalate	40.000	38.240	4.4	83	0.00
51	2,6-Dinitrotoluene	40.000	39.009	2.5	83	0.00
52 C	Acenaphthene	40.000	35.830	10.4	77	0.00
53	3-Nitroaniline	40.000	39.844	0.4	84	0.00
54 P	2,4-Dinitrophenol	40.000	35.458	11.4	76	0.00
55	Dibenzofuran	40.000	38.059	4.9	84	0.00
56 P	4-Nitrophenol	40.000	39.202	2.0	81	0.00
57	2,4-Dinitrotoluene	40.000	39.414	1.5	83	0.00
58	Fluorene	40.000	39.062	2.3	85	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.894	2.8	84	0.00
60	Diethylphthalate	40.000	38.898	2.8	85	0.00
61	4-Chlorophenyl-phenylether	40.000	38.106	4.7	84	0.00
62	4-Nitroaniline	40.000	41.739	-4.3	89	0.00
63	Azobenzene	40.000	38.202	4.5	84	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.840	0.4	86	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.957	7.6	83	0.00
67	4-Bromophenyl-phenylether	40.000	37.364	6.6	85	0.00
68	Hexachlorobenzene	40.000	36.669	8.3	84	0.00
69	Atrazine	40.000	38.693	3.3	86	0.00
70 C	Pentachlorophenol	40.000	35.469	11.3	78	0.00
71	Phenanthrene	40.000	37.754	5.6	87	0.00
72	Anthracene	40.000	38.115	4.7	87	0.00
73	Carbazole	40.000	38.789	3.0	89	0.00
74	Di-n-butylphthalate	40.000	39.329	1.7	89	0.00
75 C	Fluoranthene	40.000	38.429	3.9	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	37.636	5.9	78	0.00
78	Pyrene	40.000	38.092	4.8	89	0.00
79 S	Terphenyl-d14	80.000	76.092	4.9	91	0.00
80	Butylbenzylphthalate	40.000	41.140	-2.9	95	0.00
81	Benzo(a)anthracene	40.000	38.952	2.6	92	0.00
82	3,3'-Dichlorobenzidine	40.000	38.828	2.9	90	0.00
83	Chrysene	40.000	38.173	4.6	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.471	-1.2	91	-0.02
85 c	Di-n-octyl phthalate	40.000	41.414	-3.5	95	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	96	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	38.958	2.6	93	0.00
88	Benzo(b)fluoranthene	40.000	38.615	3.5	93	0.00
89	Benzo(k)fluoranthene	40.000	39.337	1.7	93	0.00
90 C	Benzo(a)pyrene	40.000	38.486	3.8	92	-0.02
91	Dibenzo(a,h)anthracene	40.000	38.699	3.3	93	-0.02
92	Benzo(g,h,i)perylene	40.000	38.631	3.4	92	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021622\
Data File : BG052412.D
Acq On : 16 Feb 2022 18:56
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 17 00:42:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 08 01:21:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0