

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021722\
 Data File : BG052418.D
 Acq On : 17 Feb 2022 14:26
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 17 15:36:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 17 11:30:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.02
2	1,4-Dioxane	40.000	40.003	-0.0	97	0.00
3	Pyridine	40.000	41.348	-3.4	93	0.01
4	n-Nitrosodimethylamine	40.000	36.987	7.5	86	0.01
5 S	2-Fluorophenol	80.000	84.635	-5.8	98	0.02
6	Aniline	40.000	39.900	0.3	89	0.02
7 S	Phenol-d6	80.000	79.781	0.3	90	0.01
8	2-Chlorophenol	40.000	40.207	-0.5	95	0.01
9	Benzaldehyde	40.000	35.215	12.0	85	0.01
10 C	Phenol	40.000	39.904	0.2	91	0.02
11	bis(2-Chloroethyl)ether	40.000	38.831	2.9	92	0.01
12	1,3-Dichlorobenzene	40.000	40.239	-0.6	94	0.01
13 C	1,4-Dichlorobenzene	40.000	39.857	0.4	93	0.01
14	1,2-Dichlorobenzene	40.000	38.564	3.6	90	0.01
15	Benzyl Alcohol	40.000	38.460	3.8	85	0.02
16	2,2'-oxybis(1-Chloropropane	40.000	37.343	6.6	89	0.02
17	2-Methylphenol	40.000	40.867	-2.2	93	0.02
18	Hexachloroethane	40.000	40.124	-0.3	96	0.02
19 P	n-Nitroso-di-n-propylamine	40.000	38.340	4.1	89	0.02
20	3+4-Methylphenols	40.000	39.726	0.7	90	0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	93	0.01
22	Acetophenone	40.000	37.117	7.2	89	0.02
23 S	Nitrobenzene-d5	80.000	74.319	7.1	88	0.01
24	Nitrobenzene	40.000	36.926	7.7	86	0.01
25	Isophorone	40.000	37.698	5.8	88	0.01
26 C	2-Nitrophenol	40.000	39.230	1.9	89	0.01
27	2,4-Dimethylphenol	40.000	38.891	2.8	90	0.01
28	bis(2-Chloroethoxy)methane	40.000	38.580	3.6	91	0.01
29 C	2,4-Dichlorophenol	40.000	38.955	2.6	92	0.01
30	1,2,4-Trichlorobenzene	40.000	37.684	5.8	90	0.01
31	Naphthalene	40.000	38.141	4.6	91	0.02
32	Benzoic acid	40.000	33.406	16.5	77	0.02
33	4-Chloroaniline	40.000	39.234	1.9	90	0.02
34 C	Hexachlorobutadiene	40.000	37.411	6.5	91	0.01
35	Caprolactam	40.000	40.571	-1.4	94	0.02
36 C	4-Chloro-3-methylphenol	40.000	39.211	2.0	90	0.01
37	2-Methylnaphthalene	40.000	37.824	5.4	90	0.01
38	1-Methylnaphthalene	40.000	38.452	3.9	90	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.989	7.5	89	0.01
41 P	Hexachlorocyclopentadiene	40.000	36.113	9.7	84	0.01
42 S	2,4,6-Tribromophenol	80.000	78.440	2.0	92	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.363	6.6	87	0.01
44	2,4,5-Trichlorophenol	40.000	37.382	6.5	88	0.01
45 S	2-Fluorobiphenyl	80.000	75.859	5.2	90	0.01
46	1,1'-Biphenyl	40.000	37.920	5.2	90	0.01
47	2-Chloronaphthalene	40.000	37.871	5.3	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021722\
 Data File : BG052418.D
 Acq On : 17 Feb 2022 14:26
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 17 15:36:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 17 11:30:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.644	5.9	88	0.01
49	Acenaphthylene	40.000	37.976	5.1	91	0.00
50	Dimethylphthalate	40.000	37.857	5.4	91	0.00
51	2,6-Dinitrotoluene	40.000	39.257	1.9	92	0.01
52 C	Acenaphthene	40.000	38.087	4.8	90	0.00
53	3-Nitroaniline	40.000	40.182	-0.5	93	0.00
54 P	2,4-Dinitrophenol	40.000	38.128	4.7	86	0.01
55	Dibenzofuran	40.000	37.666	5.8	91	0.00
56 P	4-Nitrophenol	40.000	39.225	1.9	89	0.01
57	2,4-Dinitrotoluene	40.000	39.393	1.5	91	0.01
58	Fluorene	40.000	38.659	3.4	92	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.446	1.4	93	0.00
60	Diethylphthalate	40.000	37.877	5.3	90	0.00
61	4-Chlorophenyl-phenylether	40.000	37.484	6.3	91	0.00
62	4-Nitroaniline	40.000	40.477	-1.2	94	0.00
63	Azobenzene	40.000	37.445	6.4	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.308	-0.8	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.666	5.8	92	0.00
67	4-Bromophenyl-phenylether	40.000	37.412	6.5	92	0.00
68	Hexachlorobenzene	40.000	37.276	6.8	92	0.00
69	Atrazine	40.000	37.883	5.3	92	0.00
70 C	Pentachlorophenol	40.000	35.850	10.4	85	0.01
71	Phenanthrene	40.000	37.465	6.3	94	0.00
72	Anthracene	40.000	37.705	5.7	94	0.00
73	Carbazole	40.000	38.473	3.8	96	0.00
74	Di-n-butylphthalate	40.000	39.386	1.5	97	0.00
75 C	Fluoranthene	40.000	38.396	4.0	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	38.699	3.3	86	0.00
78	Pyrene	40.000	38.258	4.4	96	0.00
79 S	Terphenyl-d14	80.000	75.062	6.2	97	0.00
80	Butylbenzylphthalate	40.000	40.881	-2.2	102	0.00
81	Benzo(a)anthracene	40.000	38.123	4.7	97	0.00
82	3,3'-Dichlorobenzidine	40.000	38.307	4.2	95	0.00
83	Chrysene	40.000	37.450	6.4	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.975	0.1	97	0.00
85 c	Di-n-octyl phthalate	40.000	40.626	-1.6	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.178	4.6	98	0.02
88	Benzo(b)fluoranthene	40.000	38.483	3.8	99	0.00
89	Benzo(k)fluoranthene	40.000	37.971	5.1	97	0.00
90 C	Benzo(a)pyrene	40.000	38.062	4.8	97	0.00
91	Dibenzo(a,h)anthracene	40.000	37.602	6.0	97	0.00
92	Benzo(g,h,i)perylene	40.000	38.020	4.9	98	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG021722\
Data File : BG052418.D
Acq On : 17 Feb 2022 14:26
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 17 15:36:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 17 11:30:33 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0