

Data Path : Z:\HPCHEM1\BNA G\DATA\BG021816\
 Data File : BG020967.D
 Acq On : 19 Feb 2016 11:30
 Operator : SJ/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 19 12:24:42 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 11 15:26:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2	1,4-Dioxane	0.432	0.404	6.5	121	-0.01
3	Pyridine	1.277	1.219	4.5	116	-0.01
4	n-Nitrosodimethylamine	0.334	0.320	4.2	118	-0.01
5 S	2-Fluorophenol	1.188	1.180	0.7	125	0.00
6	Aniline	2.164	2.141	1.1	123	0.00
7 S	Phenol-d6	1.736	1.745	-0.5	125	0.00
8	2-Chlorophenol	1.484	1.514	-2.0	129	-0.01
9	Benzaldehyde	0.869	0.813	6.4	117	0.00
10 C	Phenol	1.837	1.817	1.1	124	0.00
11	bis(2-Chloroethyl)ether	1.467	1.453	1.0	122	-0.01
12	1,3-Dichlorobenzene	1.569	1.598	-1.8	129	-0.01
13 C	1,4-Dichlorobenzene	1.590	1.640	-3.1	129	-0.01
14	1,2-Dichlorobenzene	1.547	1.586	-2.5	128	-0.01
15	Benzyl Alcohol	1.094	1.060	3.1	121	0.00
16	2,2'-oxybis(1-Chloropropane	1.530	1.457	4.8	120	0.00
17	2-Methylphenol	1.323	1.314	0.7	125	0.00
18	Hexachloroethane	0.547	0.566	-3.5	130	-0.01
19 P	n-Nitroso-di-n-propylamine	1.120	1.068	4.6	120	-0.01
20	3+4-Methylphenols	1.844	1.828	0.9	126	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	126	-0.01
22	Acetophenone	0.482	0.479	0.6	123	0.00
23 S	Nitrobenzene-d5	0.304	0.301	1.0	122	0.00
24	Nitrobenzene	0.317	0.314	0.9	124	0.00
25	Isophorone	0.641	0.620	3.3	120	0.00
26 C	2-Nitrophenol	0.167	0.180	-7.8	128	-0.01
27	2,4-Dimethylphenol	0.320	0.312	2.5	119	-0.01
28	bis(2-Chloroethoxy)methane	0.395	0.389	1.5	123	-0.01
29 C	2,4-Dichlorophenol	0.287	0.293	-2.1	129	0.00
30	1,2,4-Trichlorobenzene	0.289	0.301	-4.2	130	0.00
31	Naphthalene	1.035	1.034	0.1	125	-0.01
32	Benzoic acid	0.256	0.246	3.9	122	0.00
33	4-Chloroaniline	0.445	0.441	0.9	124	0.00
34 C	Hexachlorobutadiene	0.147	0.155	-5.4	132	-0.01
35	Caprolactam	0.110	0.106	3.6	122	0.00
36 C	4-Chloro-3-methylphenol	0.328	0.320	2.4	123	0.00
37	2-Methylnaphthalene	0.726	0.719	1.0	125	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	125	-0.01
39	1,2,4,5-Tetrachlorobenzene	0.587	0.606	-3.2	128	0.00
40 P	Hexachlorocyclopentadiene	0.258	0.218	15.5	101	-0.01
41 S	2,4,6-Tribromophenol	0.197	0.204	-3.6	129	0.00
42 C	2,4,6-Trichlorophenol	0.394	0.409	-3.8	125	-0.01
43	2,4,5-Trichlorophenol	0.414	0.426	-2.9	126	0.00
44 S	2-Fluorobiphenyl	1.424	1.447	-1.6	125	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG021816\
 Data File : BG020967.D
 Acq On : 19 Feb 2016 11:30
 Operator : SJ/UM
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 19 12:24:42 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 11 15:26:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.764	1.770	-0.3	124	-0.01
46	2-Chloronaphthalene	1.263	1.268	-0.4	125	-0.01
47	2-Nitroaniline	0.303	0.301	0.7	122	0.00
48	Acenaphthylene	2.196	2.196	0.0	124	-0.01
49	Dimethylphthalate	1.564	1.558	0.4	123	-0.01
50	2,6-Dinitrotoluene	0.335	0.342	-2.1	126	-0.01
51 C	Acenaphthene	1.332	1.327	0.4	124	-0.01
52	3-Nitroaniline	0.389	0.392	-0.8	124	0.00
53 P	2,4-Dinitrophenol	0.116	0.097	16.4	103	0.00
54	Dibenzofuran	1.762	1.753	0.5	125	-0.01
55 P	4-Nitrophenol	0.294	0.299	-1.7	122	0.00
56	2,4-Dinitrotoluene	0.433	0.450	-3.9	126	-0.01
57	Fluorene	1.622	1.598	1.5	123	-0.01
58	2,3,4,6-Tetrachlorophenol	0.316	0.319	-0.9	125	0.00
59	Diethylphthalate	1.608	1.591	1.1	124	-0.01
60	4-Chlorophenyl-phenylether	0.720	0.714	0.8	126	-0.01
61	4-Nitroaniline	0.391	0.392	-0.3	123	0.00
62	Azobenzene	1.264	1.197	5.3	118	-0.01
63 I	Phenanthrene-d10	1.000	1.000	0.0	123	-0.01
64	4,6-Dinitro-2-methylphenol	0.094	0.087	7.4	111	0.00
65 c	n-Nitrosodiphenylamine	0.646	0.667	-3.3	124	-0.01
66	4-Bromophenyl-phenylether	0.211	0.219	-3.8	125	-0.02
67	Hexachlorobenzene	0.225	0.235	-4.4	128	-0.01
68	Atrazine	0.196	0.200	-2.0	125	-0.01
69 C	Pentachlorophenol	0.129	0.134	-3.9	126	-0.01
70	Phenanthrene	1.139	1.154	-1.3	124	-0.01
71	Anthracene	1.156	1.177	-1.8	124	-0.01
72	Carbazole	1.037	1.047	-1.0	123	-0.01
73	Di-n-butylphthalate	1.330	1.353	-1.7	124	-0.01
74 C	Fluoranthene	1.230	1.249	-1.5	124	-0.01
75 I	Chrysene-d12	1.000	1.000	0.0	122	-0.02
76	Benzidine	0.647	0.597	7.7	109	-0.01
77	Pyrene	1.319	1.349	-2.3	125	-0.01
78 S	Terphenyl-d14	0.857	0.874	-2.0	124	-0.02
79	Butylbenzylphthalate	0.632	0.650	-2.8	123	-0.02
80	Benzo(a)anthracene	1.194	1.205	-0.9	124	-0.02
81	3,3'-Dichlorobenzidine	0.443	0.446	-0.7	122	-0.02
82	Chrysene	1.123	1.130	-0.6	123	-0.02
83	Bis(2-ethylhexyl)phthalate	0.936	0.948	-1.3	122	-0.02
84 c	Di-n-octyl phthalate	1.537	1.576	-2.5	123	-0.04
85	Indeno(1,2,3-cd)pyrene	1.291	1.356	-5.0	127	-0.06
86 I	Perylene-d12	1.000	1.000	0.0	125	-0.04
87	Benzo(b)fluoranthene	1.224	1.233	-0.7	125	-0.03

Data Path : Z:\HPCHEM1\BNA G\DATA\BG021816\
Data File : BG020967.D
Acq On : 19 Feb 2016 11:30
Operator : SJ/UM
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 19 12:24:42 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 11 15:26:53 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.199	1.183	1.3	123	-0.03
89 C	Benzo(a)pyrene	1.166	1.172	-0.5	125	-0.04
90	Dibenzo(a,h)anthracene	1.134	1.146	-1.1	126	-0.05
91	Benzo(a,h,i)perylene	1.126	1.139	-1.2	126	-0.07

(#) = Out of Range

SPCC's out = 0 CCC's out = 0