

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG021819\
 Data File : BG039548.D
 Acq On : 18 Feb 2019 20:41
 Operator : JU/SJ
 Sample : K1508-04MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD

Quant Time: Feb 19 00:45:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 29 15:41:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	59054	20.00	ng	-0.02
21) Naphthalene-d8	10.99	136	262276	20.00	ng	-0.02
39) Acenaphthene-d10	14.79	164	184677	20.00	ng	-0.02
64) Phenanthrene-d10	17.54	188	456609	20.00	ng	-0.02
76) Chrysene-d12	21.83	240	443125	20.00	ng	-0.02
87) Perylene-d12	25.20	264	448001	20.00	ng	-0.04

System Monitoring Compounds

5) 2-Fluorophenol	5.71	112	422695	122.51	ng	0.00
7) Phenol-d6	7.31	99	569244	112.08	ng	0.00
23) Nitrobenzene-d5	9.35	82	392588	93.51	ng	-0.02
42) 2,4,6-Tribromophenol	16.28	330	270620	136.08	ng	-0.01
45) 2-Fluorobiphenyl	13.42	172	944163	73.34	ng	-0.02
79) Terphenyl-d14	20.13	244	1497231	71.52	ng	-0.02

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.61	88	46223	30.626	ng	# 32
3) Pyridine	4.01	79	93219	23.328	ng	95
4) n-Nitrosodimethylamine	3.93	42	69715	34.884	ng	83
6) Aniline	7.50	93	182441	28.678	ng	95
8) 2-Chlorophenol	7.73	128	158007	41.894	ng	95
9) Benzaldehyde	7.31	77	117404	49.498	ng	98
10) Phenol	7.34	94	188678	35.638	ng	95
11) bis(2-Chloroethyl)ether	7.59	93	152962	36.563	ng	99
12) 1,3-Dichlorobenzene	8.05	146	166371	36.413	ng	97
13) 1,4-Dichlorobenzene	8.21	146	167086	36.690	ng	96
14) 1,2-Dichlorobenzene	8.52	146	161655	36.668	ng	98
15) Benzyl Alcohol	8.41	79	157266	39.441	ng	95
16) 2,2'-oxybis(1-Chloropropan	8.68	45	292285	30.938	ng	99
17) 2-Methylphenol	8.60	107	138226	40.345	ng	98
18) Hexachloroethane	9.25	117	57549	36.731	ng	90
19) n-Nitroso-di-n-propylamine	8.97	70	127653	35.412	ng	92
20) 3+4-Methylphenols	8.93	107	190812	40.444	ng	97
22) Acetophenone	9.00	105	248850	35.322	ng	# 94
24) Nitrobenzene	9.39	77	193349	42.629	ng	96
25) Isophorone	9.90	82	370131	39.784	ng	100
26) 2-Nitrophenol	10.10	139	93852	51.593	ng	94
27) 2,4-Dimethylphenol	10.14	122	164568	43.728	ng	92
28) bis(2-Chloroethoxy)methane	10.38	93	222326	38.798	ng	97
29) 2,4-Dichlorophenol	10.63	162	180535	43.891	ng	94
30) 1,2,4-Trichlorobenzene	10.85	180	176829	38.292	ng	99
31) Naphthalene	11.04	128	518029	39.814	ng	98
32) Benzoic acid	10.26	122	92712	39.571	ng	98
33) 4-Chloroaniline	11.15	127	139615	23.142	ng	98
34) Hexachlorobutadiene	11.30	225	111564	36.328	ng	97
35) Caprolactam	11.93	113	49969	29.358	ng	97
36) 4-Chloro-3-methylphenol	12.25	107	196739	41.358	ng	96
37) 2-Methylnaphthalene	12.63	142	395167	41.006	ng	98
38) 1-Methylnaphthalene	12.85	142	378524	40.747	ng	100
40) 1,2,4,5-Tetrachlorobenzene	12.99	216	218501	37.186	ng	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG021819\
 Data File : BG039548.D
 Acq On : 18 Feb 2019 20:41
 Operator : JU/SJ
 Sample : K1508-04MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD

Quant Time: Feb 19 00:45:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 29 15:41:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.96	237	258434	80.714	ng	99
43) 2,4,6-Trichlorophenol	13.23	196	150286	41.598	ng	97
44) 2,4,5-Trichlorophenol	13.30	196	167729	44.296	ng	# 98
46) 1,1'-Biphenyl	13.63	154	536698	34.929	ng	99
47) 2-Chloronaphthalene	13.68	162	419665	36.306	ng	99
48) 2-Nitroaniline	13.88	65	149074	45.431	ng	96
49) Acenaphthylene	14.52	152	676485	38.785	ng	97
50) Dimethylphthalate	14.24	163	560817	37.600	ng	99
51) 2,6-Dinitrotoluene	14.37	165	127835	46.151	ng	95
52) Acenaphthene	14.86	154	416580	36.794	ng	99
53) 3-Nitroaniline	14.70	138	113142	38.869	ng	# 91
54) 2,4-Dinitrophenol	14.90	184	99397	80.410	ng	96
55) Dibenzofuran	15.19	168	685736	37.776	ng	98
56) 4-Nitrophenol	15.00	139	170094	64.298	ng	87
57) 2,4-Dinitrotoluene	15.15	165	184849	51.153	ng	87
58) Fluorene	15.84	166	543233	40.144	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.41	232	162390	45.804	ng	97
60) Diethylphthalate	15.59	149	572757	37.141	ng	97
61) 4-Chlorophenyl-phenylether	15.82	204	288176	37.610	ng	93
62) 4-Nitroaniline	15.87	138	133335	43.101	ng	95
63) Azobenzene	16.11	77	513131	34.043	ng	97
65) 4,6-Dinitro-2-methylphenol	15.91	198	89907	49.395	ng	94
66) n-Nitrosodiphenylamine	16.04	169	502068	36.162	ng	97
67) 4-Bromophenyl-phenylether	16.72	248	186579	36.833	ng	96
68) Hexachlorobenzene	16.83	284	196910	38.744	ng	93
69) Atrazine	16.98	200	178546	35.190	ng	97
70) Pentachlorophenol	17.18	266	228151	64.590	ng	100
71) Phenanthrene	17.58	178	899868	38.077	ng	99
72) Anthracene	17.67	178	912094	39.182	ng	99
73) Carbazole	17.94	167	788702	33.950	ng	98
74) Di-n-butylphthalate	18.47	149	954988	35.236	ng	99
75) Fluoranthene	19.58	202	1055611	38.484	ng	100
77) Benzidine	19.76	184	219516	15.110	ng	98
78) Pyrene	19.94	202	1059870	38.421	ng	98
80) Butylbenzylphthalate	20.81	149	440588	37.862	ng	94
81) Benzo(a)anthracene	21.81	228	1031167	39.382	ng	100
82) 3,3'-Dichlorobenzidine	21.72	252	314775	32.421	ng	100
83) Chrysene	21.88	228	964731	38.705	ng	99
84) Bis(2-ethylhexyl)phthalate	21.67	149	616064	37.109	ng	100
85) Di-n-octyl phthalate	22.93	149	1049864	38.278	ng	95
86) Indeno(1,2,3-cd)pyrene	29.09	276	998821	37.349	ng	96
88) Benzo(b)fluoranthene	24.12	252	993940	40.682	ng	98
89) Benzo(k)fluoranthene	24.20	252	1005403	40.988	ng	99
90) Benzo(a)pyrene	25.04	252	964645	40.997	ng	99
91) Dibenzo(a,h)anthracene	29.16	278	814184	36.949	ng	98
92) Benzo(g,h,i)perylene	30.32	276	780319	35.528	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.17 min Scan# 864
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Instrument :
BNA_G
ClientSampleId :
M-12-SMSD

Tgt Ion:152 Resp: 59054
Ion Ratio Lower Upper
152 100
150 156.1 123.7 185.5
115 58.5 45.9 68.9




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 30.626 ng
RT: 3.61 min Scan# 88
Delta R.T. -0.01 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion: 88 Resp: 46223
Ion Ratio Lower Upper
88 100
58 94.8 81.1 121.7
43 177.5 31.4 47.0#



Abundance Ion 88.00 (87.70 to 88.70): BGC
Ion 58.00 (57.70 to 58.70): BGC
Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 23.328 ng

RT: 4.01 min Scan# 157

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion: 79 Resp: 93219

Ion Ratio Lower Upper




79 100

52 97.0 82.6 124.0

51 46.3 37.5 56.3



Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 52.00 (51.70 to 52.70): BGC
Ion 51.00 (50.70 to 51.70): BGC



#4

n-Nitrosodimethylamine

Concen: 34.884 ng

RT: 3.93 min Scan# 142

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion: 42 Resp: 69715

Ion Ratio Lower Upper



42 100


74 123.2 83.6 125.4

44 9.0 6.9 10.3



Abundance Ion 42.00 (41.70 to 42.70): BGC
Ion 74.00 (73.70 to 74.70): BGC
Ion 44.00 (43.70 to 44.70): BGC





#5

2-Fluorophenol

Concen: 122.506 ng

RT: 5.71 min Scan# 446

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion: 112 Resp: 422695

Ion Ratio Lower Upper




112 100

64 70.1 57.8 86.8

63 36.6 29.8 44.6



Abundance Ion 112.00 (111.70 to 112.70): BGC
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 28.678 ng

RT: 7.50 min Scan# 750

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion: 93 Resp: 182441

Ion Ratio Lower Upper



93 100


66 38.9 34.2 51.4

65 21.2 17.7 26.5



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 66.00 (65.70 to 66.70): BGC
Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 112.080 ng

RT: 7.31 min Scan# 719

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion: 99 Resp: 569244

Ion Ratio Lower Upper




99 100

42 19.9 18.2 27.2

71 35.1 28.0 42.0



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 41.894 ng

RT: 7.73 min Scan# 790

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion: 128 Resp: 158007

Ion Ratio Lower Upper



128 100


130 32.8 12.9 52.9

64 53.0 38.7 78.7



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 49.498 ng

RT: 7.31 min Scan# 718

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD





Tgt Ion: 77 Resp: 117404

Ion Ratio Lower Upper

77 100

105 93.5 74.5 114.5

106 90.5 68.0 108.0



#10

Phenol

Concen: 35.638 ng

RT: 7.34 min Scan# 724

Delta R.T. 0.00 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41




Tgt Ion: 94 Resp: 188678


Ion Ratio Lower Upper

94 100

65 31.0 11.7 51.7

66 38.7 23.7 63.7









#11
 bis(2-Chloroethyl)ether
 Concen: 36.563 ng
 RT: 7.59 min Scan# 765
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD


Tgt Ion: 93 Resp: 152962
 Ion Ratio Lower Upper
 93 100
 63 87.7 69.5 109.5
 95 32.5 12.7 52.7



#12
 1,3-Dichlorobenzene
 Concen: 36.413 ng
 RT: 8.05 min Scan# 845
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion: 146 Resp: 166371
 Ion Ratio Lower Upper
 146 100
 148 63.5 52.6 79.0
 75 30.3 25.8 38.6









#13
 1,4-Dichlorobenzene
 Concen: 36.690 ng
 RT: 8.21 min Scan# 871
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD

Tgt Ion:146 Resp: 167086
 Ion Ratio Lower Upper
 146 100
 148 61.2 51.1 76.7
 111 37.9 32.6 48.8




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#14
 1,2-Dichlorobenzene
 Concen: 36.668 ng
 RT: 8.52 min Scan# 925
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion:146 Resp: 161655
 Ion Ratio Lower Upper
 146 100
 148 63.5 51.5 77.3
 111 40.2 33.8 50.6



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 39.441 ng

RT: 8.41 min Scan# 905

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion: 79 Resp: 157266

Ion Ratio Lower Upper




79 100

108 78.4 57.8 86.6

77 63.0 51.1 76.7



Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 108.00 (107.70 to 108.70): BGC
Ion 77.00 (76.70 to 77.70): BGC



#16

2,2'-oxybis(1-chloropropane)

Concen: 30.938 ng

RT: 8.68 min Scan# 952

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion: 45 Resp: 292285

Ion Ratio Lower Upper



45 100


77 9.7 0.0 30.0

79 7.2 0.0 27.5



Abundance Ion 45.00 (44.70 to 45.70): BGC
Ion 77.00 (76.70 to 77.70): BGC
Ion 79.00 (78.70 to 79.70): BGC









#17
2-Methylphenol
Concen: 40.345 ng
RT: 8.60 min Scan# 938
Delta R.T. -0.01 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Instrument :
BNA_G
ClientSampleId :
M-12-SMSD

Tgt Ion:	107	Resp:	138226
Ion	Ratio	Lower	Upper
107	100		
108	114.7	90.8	136.2
77	50.3	40.7	61.1
79	47.9	40.6	60.8




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BG
Ion 79.00 (78.70 to 79.70): BG








#18
Hexachloroethane
Concen: 36.731 ng
RT: 9.25 min Scan# 1048
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:	117	Resp:	57549
Ion	Ratio	Lower	Upper
117	100		
119	89.9	76.3	114.5
201	99.8	91.1	136.7



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 35.412 ng

RT: 8.97 min Scan# 1001

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD

Tgt Ion: 70 Resp: 127653





Ion Ratio Lower Upper

70 100

42 68.0 60.4 90.6

101 10.5 7.5 11.3

130 24.1 16.9 25.3



#20

3+4-Methylphenols

Concen: 40.444 ng

RT: 8.93 min Scan# 994

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Tgt Ion: 107 Resp: 190812


Ion Ratio Lower Upper


107 100

108 90.1 67.9 107.9

77 33.9 15.4 55.4

79 30.8 13.7 53.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.99 min Scan# 1345
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41




Instrument :
BNA_G
ClientSampleId :
M-12-SMSD


Tgt Ion:	136	Resp:	262276
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.1	13.7
54	10.8	9.4	14.2
68	5.5	4.2	6.4



#22
Acetophenone
Concen: 35.322 ng
RT: 9.00 min Scan# 1006
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:	105	Resp:	248850
Ion	Ratio	Lower	Upper
105	100		
71	0.3	0.8	1.2#
51	33.0	30.2	45.2
120	23.9	18.0	27.0









#23
 Nitrobenzene-d5
 Concen: 93.511 ng
 RT: 9.35 min Scan# 1065
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleID :
 M-12-SMSD


Tgt Ion	Ratio	Lower	Upper
82	100		
128	43.7	31.3	46.9
54	58.8	50.9	76.3



#24
 Nitrobenzene
 Concen: 42.629 ng
 RT: 9.39 min Scan# 1072
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
77	100		
123	45.8	34.1	51.1
65	14.9	12.3	18.5





#25

Isophorone

Concen: 39.784 ng

RT: 9.90 min Scan# 1159

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion: 82 Resp: 370131

Ion Ratio Lower Upper




82 100

95 7.8 6.2 9.2

138 16.3 13.0 19.4



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#26

2-Nitrophenol

Concen: 51.593 ng

RT: 10.10 min Scan# 1193

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion: 139 Resp: 93852

Ion Ratio Lower Upper



139 100


109 33.6 23.4 35.0

65 52.3 44.3 66.5



Abundance Ion 139.00 (138.70 to 139.70): BGC
Ion 109.00 (108.70 to 109.70): BGC
Ion 65.00 (64.70 to 65.70): BGC









#27
 2,4-Dimethylphenol
 Concen: 43.728 ng
 RT: 10.14 min Scan# 1200
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD


Tgt Ion	Ratio	Lower	Upper
122	100		
107	125.8	91.7	137.5
121	61.3	46.4	69.6



#28
 bis(2-Chloroethoxy)methane
 Concen: 38.798 ng
 RT: 10.38 min Scan# 1241
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
93	100		
95	33.0	25.0	37.4
123	11.4	8.6	12.8









#29
 2,4-Dichlorophenol
 Concen: 43.891 ng
 RT: 10.63 min Scan# 1283
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD


Tgt Ion	Ratio	Lower	Upper
162	100		
164	62.6	48.1	88.1
98	35.5	18.1	58.1



#30
 1,2,4-Trichlorobenzene
 Concen: 38.292 ng
 RT: 10.85 min Scan# 1320
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
180	100		
182	94.2	76.0	114.0
145	27.5	22.7	34.1









#31
Naphthalene
Concen: 39.814 ng
RT: 11.04 min Scan# 1353
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41




Instrument :
BNA_G
ClientSampleId :
M-12-SMSD


Tgt Ion:128 Resp: 518029
Ion Ratio Lower Upper
128 100
129 11.6 8.7 13.1
127 14.3 10.8 16.2



#32
Benzoic acid
Concen: 39.571 ng
RT: 10.26 min Scan# 1221
Delta R.T. 0.00 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:122 Resp: 92712
Ion Ratio Lower Upper
122 100
105 118.4 101.8 141.8
77 96.1 76.3 116.3









#33
4-Chloroaniline
Concen: 23.142 ng
RT: 11.15 min Scan# 1372
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Instrument :
BNA_G
ClientSampleId :
M-12-SMSD

Tgt Ion:	127	Resp:	139615
Ion	Ratio	Lower	Upper
127	100		
129	32.0	27.0	40.4
65	34.6	28.8	43.2
92	20.0	16.6	25.0




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BG
Ion 92.00 (91.70 to 92.70): BG






#34
Hexachlorobutadiene
Concen: 36.328 ng
RT: 11.30 min Scan# 1397
Delta R.T. -0.03 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:	225	Resp:	111564
Ion	Ratio	Lower	Upper
225	100		
223	58.1	49.0	73.6
227	56.6	46.3	69.5



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 29.358 ng
 RT: 11.93 min Scan# 1505
 Delta R.T. -0.03 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD


Tgt Ion:113 Resp: 49969
 Ion Ratio Lower Upper
 113 100
 55 228.6 216.9 256.9
 56 173.1 152.6 192.6



#36
 4-Chloro-3-methylphenol
 Concen: 41.358 ng
 RT: 12.25 min Scan# 1559
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion:107 Resp: 196739
 Ion Ratio Lower Upper
 107 100
 144 26.0 20.2 30.2
 142 80.8 61.4 92.2





#37

2-Methylnaphthalene

Concen: 41.006 ng

RT: 12.63 min Scan# 1624

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:142 Resp: 395167

Ion Ratio Lower Upper

142 100

141 89.6 73.5 110.3




115 37.2 30.8 46.2



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

1-Methylnaphthalene

Concen: 40.747 ng

RT: 12.85 min Scan# 1661

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:142 Resp: 378524

Ion Ratio Lower Upper

142 100

141 92.9 74.2 111.4



116 4.1 3.0 4.6




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.79 min Scan# 1992

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:164 Resp: 184677

Ion Ratio Lower Upper

164 100

162 100.2 81.6 122.4

160 42.6 34.2 51.2






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 37.186 ng

RT: 12.99 min Scan# 1685

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Tgt Ion:216 Resp: 218501


Ion Ratio Lower Upper

216 100

214 79.4 62.6 94.0

179 20.9 16.6 25.0

108 17.7 14.0 21.0





Abundance


Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 80.714 ng

RT: 12.96 min Scan# 1680

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:237 Resp: 258434

Ion Ratio Lower Upper

237 100

235 62.8 43.3 83.3




272 12.8 0.0 32.3



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 136.082 ng

RT: 16.28 min Scan# 2245

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:330 Resp: 270620

Ion Ratio Lower Upper

330 100

332 96.9 75.7 113.5



141 40.3 35.6 53.4




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 41.598 ng
 RT: 13.23 min Scan# 1726
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD


Tgt Ion	Ratio	Lower	Upper
196	100		
198	105.0	80.6	121.0
200	31.9	26.1	39.1



#44
 2,4,5-Trichlorophenol
 Concen: 44.296 ng
 RT: 13.30 min Scan# 1738
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
196	100		
198	98.4	78.5	117.7
97	48.1	38.4	57.6
132	29.2	19.2	28.8









#45
2-Fluorobiphenyl
Concen: 73.342 ng
RT: 13.42 min Scan# 1758
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41




Instrument :
BNA_G
ClientSampleId :
M-12-SMSD


Tgt Ion	Ratio	Lower	Upper
172	100		
171	38.6	30.8	46.2
170	25.9	20.2	30.4



#46
1,1'-Biphenyl
Concen: 34.929 ng
RT: 13.63 min Scan# 1794
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
154	100		
153	42.5	22.1	62.1
76	12.3	0.0	32.4









#47
 2-Chloronaphthalene
 Concen: 36.306 ng
 RT: 13.68 min Scan# 1802
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD

Tgt Ion: 162 Resp: 419665
 Ion Ratio Lower Upper
 162 100
 127 38.2 31.0 46.6
 164 33.2 25.7 38.5




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E






#48
 2-Nitroaniline
 Concen: 45.431 ng
 RT: 13.88 min Scan# 1837
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion: 65 Resp: 149074
 Ion Ratio Lower Upper
 65 100
 92 64.8 51.4 77.2
 138 94.7 71.4 107.2



Abundance Ion 65.00 (64.70 to 65.70): BGC
 Ion 92.00 (91.70 to 92.70): BGC
 Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 38.785 ng

RT: 14.52 min Scan# 1946

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:152 Resp: 676485

Ion Ratio Lower Upper




152 100

151 22.6 16.8 25.2

153 13.8 11.7 17.5



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 37.600 ng

RT: 14.24 min Scan# 1898

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:163 Resp: 560817

Ion Ratio Lower Upper



163 100


194 4.6 3.6 5.4

164 10.7 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E





#51

2,6-Dinitrotoluene

Concen: 46.151 ng

RT: 14.37 min Scan# 1920

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:165 Resp: 127835

Ion Ratio Lower Upper

165 100

63 55.3 47.0 70.6




89 54.0 45.8 68.8



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 89.00 (88.70 to 89.70): BGC



#52

Acenaphthene

Concen: 36.794 ng

RT: 14.86 min Scan# 2003

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:154 Resp: 416580

Ion Ratio Lower Upper

154 100

153 112.9 90.1 135.1



152 51.8 40.9 61.3




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 38.869 ng
RT: 14.70 min Scan# 1977
Delta R.T. -0.01 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41




Instrument :
BNA_G
ClientSampleId :
M-12-SMSD


Tgt Ion:138 Resp: 113142
Ion Ratio Lower Upper
138 100
108 11.1 13.8 20.8#
92 120.6 103.7 155.5



#54
2,4-Dinitrophenol
Concen: 80.410 ng
RT: 14.90 min Scan# 2011
Delta R.T. -0.01 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:184 Resp: 99397
Ion Ratio Lower Upper
184 100
63 61.1 50.9 76.3
154 63.2 53.0 79.6





#55

Dibenzenofuran

Concen: 37.776 ng

RT: 15.19 min Scan# 2060

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:168 Resp: 685736

Ion Ratio Lower Upper

168 100

139 38.9 30.1 45.1




169 14.3 10.6 16.0



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 64.298 ng

RT: 15.00 min Scan# 2027

Delta R.T. 0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:139 Resp: 170094

Ion Ratio Lower Upper

139 100

109 75.1 61.9 101.9



65 101.0 99.1 139.1




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BGC





#57

2,4-Dinitrotoluene

Concen: 51.153 ng

RT: 15.15 min Scan# 2053

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD

Tgt Ion:165 Resp: 184849


Ion Ratio Lower Upper

165 100

63 43.7 41.0 61.6

89 67.6 64.6 96.8

182 2.8 2.0 3.0






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 89.00 (88.70 to 89.70): BGC

Ion 182.00 (181.70 to 182.70): E



#58

Fluorene

Concen: 40.144 ng

RT: 15.84 min Scan# 2170

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:166 Resp: 543233

Ion Ratio Lower Upper

166 100

165 99.4 77.9 116.9



167 14.0 10.8 16.2




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E






#59
2,3,4,6-Tetrachlorophenol
Concen: 45.804 ng
RT: 15.41 min Scan# 2097
Delta R.T. -0.01 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Instrument :
BNA_G
ClientSampleId :
M-12-SMSD

Tgt Ion	Ratio	Lower	Upper
232	100		
131	41.0	34.9	52.3
130	2.4	2.0	3.0
166	29.1	24.1	36.1






Abundance

Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E


Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#60
Diethylphthalate
Concen: 37.141 ng
RT: 15.59 min Scan# 2128
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
149	100		
177	24.5	18.3	27.5
150	13.3	10.2	15.2






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E









#61
 4-Chlorophenyl-phenylether
 Concen: 37.610 ng
 RT: 15.82 min Scan# 2167
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD


Tgt Ion	Ratio	Lower	Upper
204	100		
206	32.9	29.2	43.8
141	54.9	49.0	73.4



#62
 4-Nitroaniline
 Concen: 43.101 ng
 RT: 15.87 min Scan# 2175
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
138	100		
92	56.4	38.7	78.7
108	89.1	74.6	114.6









#63
Azobenzene
Concen: 34.043 ng
RT: 16.11 min Scan# 2217
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Instrument :
BNA_G
ClientSampleId :
M-12-SMSD

Tgt Ion:	77	Resp:	513131
Ion	Ratio	Lower	Upper
77	100		
182	28.2	6.4	46.4
105	20.2	0.0	39.8
51	31.6	14.1	54.1




Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 182.00 (181.70 to 182.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 51.00 (50.70 to 51.70): BGC






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.54 min Scan# 2459
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:	188	Resp:	456609
Ion	Ratio	Lower	Upper
188	100		
94	7.9	6.9	10.3
80	9.1	8.1	12.1



Abundance Ion 188.00 (187.70 to 188.70): BGC
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC





#65

4,6-Dinitro-2-methylphenol

Concen: 49.395 ng

RT: 15.91 min Scan# 2182

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:198 Resp: 89907

Ion Ratio Lower Upper

198 100

51 41.5 27.4 67.4




105 41.2 19.7 59.7



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BG

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 36.162 ng

RT: 16.04 min Scan# 2204

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:169 Resp: 502068

Ion Ratio Lower Upper

169 100

168 66.8 56.6 85.0



167 35.7 28.8 43.2




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
 4-Bromophenyl-phenylether
 Concen: 36.833 ng
 RT: 16.72 min Scan# 2320
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD

Tgt Ion:248 Resp: 186579
 Ion Ratio Lower Upper
 248 100
 250 97.5 77.6 116.4
 141 72.0 63.9 95.9




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#68
 Hexachlorobenzene
 Concen: 38.744 ng
 RT: 16.83 min Scan# 2339
 Delta R.T. -0.01 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion:284 Resp: 196910
 Ion Ratio Lower Upper
 284 100
 142 32.2 30.6 45.8
 249 33.2 28.2 42.2



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E





#69

Atrazine

Concen: 35.190 ng

RT: 16.98 min Scan# 2364

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampled :

M-12-SMSD


Tgt Ion:200 Resp: 178546

Ion Ratio Lower Upper

200 100

173 24.5 3.9 43.9

215 48.3 30.9 70.9






Abundance

Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 64.590 ng

RT: 17.18 min Scan# 2398

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:266 Resp: 228151

Ion Ratio Lower Upper

266 100

268 64.0 51.1 76.7

264 65.3 52.1 78.1






Abundance

Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#71
Phenanthrene
Concen: 38.077 ng
RT: 17.58 min Scan# 2467
Delta R.T. -0.01 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Instrument :
BNA_G
ClientSampleId :
M-12-SMSD

Tgt Ion:178 Resp: 899868
Ion Ratio Lower Upper
178 100
176 20.3 16.2 24.4
179 16.4 12.6 19.0




Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E






#72
Anthracene
Concen: 39.182 ng
RT: 17.67 min Scan# 2482
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:178 Resp: 912094
Ion Ratio Lower Upper
178 100
176 19.9 15.8 23.6
179 16.0 12.5 18.7



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E









#73
 Carbazole
 Concen: 33.950 ng
 RT: 17.94 min Scan# 2528
 Delta R.T. -0.02 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41




Instrument :
 BNA_G
 ClientSampleId :
 M-12-SMSD


Tgt Ion	Ratio	Lower	Upper
167	100		
166	23.2	19.4	29.2
139	13.7	10.8	16.2



#74
 Di-n-butylphthalate
 Concen: 35.236 ng
 RT: 18.47 min Scan# 2618
 Delta R.T. -0.03 min
 Lab File: BG039548.D
 Acq: 18 Feb 2019 20:41

Tgt Ion	Ratio	Lower	Upper
149	100		
150	10.3	7.7	11.5
104	5.6	4.5	6.7





#75

Fluoranthene

Concen: 38.484 ng

RT: 19.58 min Scan# 2807

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:202 Resp: 1055611

Ion Ratio Lower Upper

202 100

101 9.5 0.0 30.0




203 18.9 0.0 38.9



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.83 min Scan# 3190

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:240 Resp: 443125

Ion Ratio Lower Upper

240 100

120 9.0 7.0 10.6



236 26.1 21.0 31.4




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 15.110 ng

RT: 19.76 min Scan# 2838

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:184 Resp: 219516

Ion Ratio Lower Upper

184 100

185 14.4 12.2 18.2




183 14.1 10.6 15.8



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 38.421 ng

RT: 19.94 min Scan# 2869

Delta R.T. -0.01 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:202 Resp: 1059870

Ion Ratio Lower Upper

202 100

200 23.5 17.8 26.8



203 18.6 15.3 22.9




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 71.519 ng

RT: 20.13 min Scan# 2900

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:244 Resp: 1497231

Ion Ratio Lower Upper

244 100

212 9.5 7.3 10.9




122 11.0 8.4 12.6



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 37.862 ng

RT: 20.81 min Scan# 3016

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:149 Resp: 440588

Ion Ratio Lower Upper

149 100

91 68.7 59.5 89.3



206 24.9 18.6 27.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E









#81
Benzo(a)anthracene
Concen: 39.382 ng
RT: 21.81 min Scan# 3186
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41




Instrument :
BNA_G
ClientSampleId :
M-12-SMSD


Tgt Ion:228 Resp: 1031167
Ion Ratio Lower Upper
228 100
226 28.7 23.0 34.6
229 21.4 16.9 25.3



#82
3,3'-Dichlorobenzidine
Concen: 32.421 ng
RT: 21.72 min Scan# 3171
Delta R.T. -0.02 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:252 Resp: 314775
Ion Ratio Lower Upper
252 100
254 64.2 51.3 76.9
126 10.2 8.0 12.0





#83

Chrysene

Concen: 38.705 ng

RT: 21.88 min Scan# 3198

Delta R.T. -0.02 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD





Tgt Ion:228 Resp: 964731

Ion Ratio Lower Upper

228 100

226 32.5 25.8 38.6

229 22.0 16.6 25.0



#84

Bis(2-ethylhexyl)phthalate

Concen: 37.109 ng

RT: 21.67 min Scan# 3163

Delta R.T. -0.04 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41




Tgt Ion:149 Resp: 616064


Ion Ratio Lower Upper

149 100

167 28.2 22.6 34.0

279 4.9 3.5 5.3





#85

Di-n-octyl phthalate

Concen: 38.278 ng

RT: 22.93 min Scan# 3378

Delta R.T. -0.06 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion:149 Resp: 1049864

Ion Ratio Lower Upper

149 100

167 1.7 1.3 1.9




43 10.5 10.2 15.4



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BGC



#86

Indeno(1,2,3-cd)pyrene

Concen: 37.349 ng

RT: 29.09 min Scan# 4426

Delta R.T. -0.06 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion:276 Resp: 998821

Ion Ratio Lower Upper

276 100

138 12.7 12.6 19.0



277 24.7 20.8 31.2




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Perylene-d12
Concen: 20.000 ng
RT: 25.20 min Scan# 3764
Delta R.T. -0.04 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41




Instrument :
BNA_G
ClientSampleId :
M-12-SMSD


Tgt Ion:	264	Resp:	448001
Ion	Ratio	Lower	Upper
264	100		
260	22.9	18.2	27.4
265	22.2	18.5	27.7



#88
Benzo(b)fluoranthene
Concen: 40.682 ng
RT: 24.12 min Scan# 3580
Delta R.T. -0.04 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion:	252	Resp:	993940
Ion	Ratio	Lower	Upper
252	100		
253	23.1	17.8	26.8
125	8.9	7.8	11.6









#89
Benzo(k)fluoranthene
Concen: 40.988 ng
RT: 24.20 min Scan# 3593
Delta R.T. -0.04 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41




Instrument :
BNA_G
ClientSampleId :
M-12-SMSD


Tgt Ion: 252 Resp: 1005403
Ion Ratio Lower Upper
252 100
253 21.9 18.2 27.2
125 9.3 7.4 11.2



#90
Benzo(a)pyrene
Concen: 40.997 ng
RT: 25.04 min Scan# 3737
Delta R.T. -0.04 min
Lab File: BG039548.D
Acq: 18 Feb 2019 20:41

Tgt Ion: 252 Resp: 964645
Ion Ratio Lower Upper
252 100
253 22.7 17.9 26.9
125 9.7 8.3 12.5





#91

Dibenzo(a,h)anthracene

Concen: 36.949 ng

RT: 29.16 min Scan# 4437

Delta R.T. -0.08 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41

Instrument :

BNA_G

ClientSampleId :

M-12-SMSD


Tgt Ion: 278 Resp: 814184

Ion Ratio Lower Upper

278 100

139 13.5 11.7 17.5




279 23.3 19.1 28.7



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 35.528 ng

RT: 30.32 min Scan# 4635

Delta R.T. -0.06 min

Lab File: BG039548.D

Acq: 18 Feb 2019 20:41


Tgt Ion: 276 Resp: 780319

Ion Ratio Lower Upper

276 100

277 24.3 19.6 29.4

138 16.5 14.7 22.1



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

