

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021915\
 Data File : BG016019.D
 Acq On : 20 Feb 2015 2:19
 Operator : TP/IZ
 Sample : MDL-04-S
 Misc : MDL-SOIL-3PPM-SOM0
 ALS Vial : 19 Sample Multiplier: 1

Quant Time: Feb 20 07:27:26 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-MA-BG021915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 20 07:16:53 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.81	152	97607	20.00	ng/ul	0.00
16) Naphthalene-d8	10.60	136	435497	20.00	ng/ul	0.00
33) Acenaphthene-d10	14.44	164	263868	20.00	ng/ul	0.00
59) Phenanthrene-d10	17.18	188	550853	20.00	ng/ul	0.00
75) Chrysene-d12	21.35	240	558444	20.00	ng/ul	0.00
83) Perylene-d12	23.65	264	555106	20.00	ng/ul	0.00

System Monitoring Compounds

3) Phenol-d5	6.97	99	335957	36.18	ng/ul	0.00
5) Bis-(2-Chloroethyl)ether-d	7.15	67	206727	37.69	ng/ul	0.00
7) 2-Chlorophenol-d4	7.34	132	258237	37.59	ng/ul	0.00
11) 4-Methylphenol-d8	8.50	113	255173	37.13	ng/ul	0.00
17) Nitrobenzene-d5	8.96	128	129842	42.18	ng/ul	0.00
20) 2-Nitrophenol-d4	9.69	143	120353	42.39	ng/ul	0.00
24) 2,4-Dichlorophenol-d3	10.21	165	220443	35.79	ng/ul	0.00
27) 4-Chloroaniline-d4	10.73	131	398919	58.80	ng/ul	0.00
41) Dimethylphthalate-d6	13.85	166	722302	40.94	ng/ul	0.00
44) Acenaphthylene-d8	14.13	160	953591	38.19	ng/ul	0.00
49) 4-Nitrophenol-d4	14.62	143	114795	32.36	ng/ul	0.00
55) Fluorene-d10	15.42	176	686829	38.73	ng/ul	0.00
60) 4,6-Dinitro-2-methylphenol	15.54	200	55021	25.91	ng/ul	0.00
68) Anthracene-d10	17.28	188	1008827	38.61	ng/ul	0.00
76) Pyrene-d10	19.57	212	1054992	40.22	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.51	264	983473	39.07	ng/ul	0.00

Target Compounds

					Qvalue
2) Benzaldehyde	6.98	77	335	0.06 ng/ul#	1
6) Bis(2-Chloroethyl)ether	7.24	93	1441	0.19 ng/ul#	81
10) 2,2'-oxybis(1-Chloropropan	8.50	45	1582	0.16 ng/ul#	67
13) N-Nitroso-di-n-propylamine	8.50	70	5977	1.02 ng/ul#	1
18) Nitrobenzene	8.96	77	858	0.12 ng/ul#	28
19) Isophorone	9.69	82	7028	0.45 ng/ul#	62
26) Naphthalene	10.74	128	436	0.02 ng/ul#	1
42) Dimethylphthalate	13.84	163	556	0.03 ng/ul#	1
50) 4-Nitrophenol	14.75	109	1440	0.70 ng/ul#	1
52) 2,4-Dinitrotoluene	14.62	165	148	0.03 ng/ul#	1
56) Fluorene	15.42	166	1553	0.08 ng/ul#	8
58) 4-Nitroaniline	15.54	138	251	0.05 ng/ul#	1
61) 4,6-Dinitro-2-methylphenol	15.54	198	74	0.03 ng/ul#	1
62) N-Nitrosodiphenylamine	15.54	169	279	0.02 ng/ul#	1
67) Phenanthrene	17.18	178	224	0.01 ng/ul#	1
69) Anthracene	17.27	178	503	0.02 ng/ul#	1
70) 1,2,3,4-Tetrachlorobenzene	13.23	216	18362	2.62 ng/uL	97
71) Pentachlorobenzene	14.75	250	35833	5.22 ng/uL	98
73) Di-n-butylphthalate	18.14	149	840	0.03 ng/ul#	78
77) Pyrene	19.57	202	1504	0.04 ng/ul#	1
80) Benzo(a)anthracene	21.35	228	1589	0.05 ng/ul#	66
81) Bis(2-ethylhexyl)phthalate	21.27	149	166	0.01 ng/ul#	48
82) Chrysene	21.35	228	1589	0.05 ng/ul#	64
84) Di-n-octyl phthalate	22.17	149	190	0.01 ng/ul	100

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021915\
Data File : BG016019.D
Acq On : 20 Feb 2015 2:19
Operator : TP/IZ
Sample : MDL-04-S
Misc : MDL-SOIL-3PPM-SOM0
ALS Vial : 19 Sample Multiplier: 1


Quant Time: Feb 20 07:27:26 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-MA-BG021915.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 20 07:16:53 2015
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
88) Benzo(a)pyrene	23.50	252	3932	0.13	ng/ul#	82
89) Indeno(1,2,3-cd)pyrene	25.99	276	820	0.02	ng/ul#	63
90) Dibenzo(a,h)anthracene	26.02	278	435	0.01	ng/ul#	28
91) Benzo(g,h,i)perylene	26.73	276	275	0.01	ng/ul#	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG021915\
Data File : BG016019.D
Acq On : 20 Feb 2015 2:19
Operator : TP/IZ
Sample : MDL-04-S
Misc : MDL-SOIL-3PPM-SOM0
ALS Vial : 19 Sample Multiplier: 1

Quant Time: Feb 20 07:27:26 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-MA-BG021915.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 20 07:16:53 2015
Response via : Initial Calibration





#2

Benzaldehyde

Concen: 0.06 ng/ul

RT: 6.98 min Scan# 730

Delta R.T. 0.02 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19


Tgt Ion: 77 Resp: 335

Ion Ratio Lower Upper

77 100

105 0.0 76.6 114.8#




106 0.0 78.2 117.4#



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#6

Bis(2-Chloroethyl)ether

Concen: 0.19 ng/ul

RT: 7.24 min Scan# 774

Delta R.T. -0.00 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19


Tgt Ion: 93 Resp: 1441

Ion Ratio Lower Upper

93 100

63 53.2 57.5 86.3#



95 26.1 25.8 38.6




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 63.00 (62.70 to 63.70): BGC


Ion 95.00 (94.70 to 95.70): BGC








#10
 2,2'-oxybis(1-Chloropropane)
 Concen: 0.16 ng/ul
 RT: 8.50 min Scan# 990
 Delta R.T. 0.16 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion: 45 Resp: 1582
 Ion Ratio Lower Upper
 45 100
 77 0.0 11.4 17.2#
 79 22.9 9.0 13.6#




Abundance Ion 45.00 (44.70 to 45.70): BGC
 Ion 77.00 (76.70 to 77.70): BGC
 Ion 79.00 (78.70 to 79.70): BGC






#13
 N-Nitroso-di-n-propylamine
 Concen: 1.02 ng/ul
 RT: 8.50 min Scan# 990
 Delta R.T. -0.11 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion: 70 Resp: 5977
 Ion Ratio Lower Upper
 70 100
 42 331.6 37.8 56.6#
 101 0.0 9.3 13.9#
 130 0.0 22.2 33.2#




Abundance Ion 70.00 (69.70 to 70.70): BGC
 Ion 42.00 (41.70 to 42.70): BGC
 Ion 101.00 (100.70 to 101.70): BGC
 Ion 130.00 (129.70 to 130.70): BGC








#18
Nitrobenzene
Concen: 0.12 ng/ul
RT: 8.96 min Scan# 1068
Delta R.T. -0.04 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19

Tgt Ion: 77 Resp: 858
Ion Ratio Lower Upper
77 100
123 0.0 39.9 59.9#
65 42.5 11.0 16.6#




Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 123.00 (122.70 to 123.70): BGC
Ion 65.00 (64.70 to 65.70): BGC






#19
Isophorone
Concen: 0.45 ng/ul
RT: 9.69 min Scan# 1191
Delta R.T. 0.16 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19

Tgt Ion: 82 Resp: 7028
Ion Ratio Lower Upper
82 100
95 13.7 4.6 6.8#
138 0.0 15.4 23.2#







Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC








#26
Naphthalene
Concen: 0.02 ng/ul
RT: 10.74 min Scan# 1370
Delta R.T. 0.09 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19


Tgt Ion:128 Resp: 436
Ion Ratio Lower Upper
128 100
129 190.9 8.7 13.1#
127 0.0 10.4 15.6#



#42
Dimethylphthalate
Concen: 0.03 ng/ul
RT: 13.84 min Scan# 1899
Delta R.T. -0.05 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19


Tgt Ion:163 Resp: 556
Ion Ratio Lower Upper
163 100
194 0.0 5.3 7.9#
164 300.0 8.5 12.7#








#50
4-Nitrophenol
Concen: 0.70 ng/ul
RT: 14.75 min Scan# 2053
Delta R.T. 0.11 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19

Tgt Ion	Ratio	Lower	Upper
109	100		
139	0.0	146.1	219.1#
65	0.0	123.5	185.3#




Abundance Ion 109.00 (108.70 to 109.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 65.00 (64.70 to 65.70): BGC






#52
2,4-Dinitrotoluene
Concen: 0.03 ng/ul
RT: 14.62 min Scan# 2031
Delta R.T. -0.17 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19

Tgt Ion	Ratio	Lower	Upper
165	100		
63	297.8	28.0	42.0#
182	0.0	4.1	6.1#




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BGC
Ion 182.00 (181.70 to 182.70): E








#56
 Fluorene
 Concen: 0.08 ng/ul
 RT: 15.42 min Scan# 2168
 Delta R.T. -0.06 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion:166 Resp: 1553
 Ion Ratio Lower Upper
 166 100
 165 0.0 75.4 113.2#
 167 34.7 11.0 16.4#




Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E






#58
 4-Nitroaniline
 Concen: 0.05 ng/ul
 RT: 15.54 min Scan# 2188
 Delta R.T. 0.04 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion:138 Resp: 251
 Ion Ratio Lower Upper
 138 100
 92 141.7 38.1 57.1#
 108 2998.6 46.4 69.6#




Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 92.00 (91.70 to 92.70): BG
 Ion 108.00 (107.70 to 108.70): E








#61
 4,6-Dinitro-2-methylphenol
 Concen: 0.03 ng/ul
 RT: 15.54 min Scan# 2188
 Delta R.T. -0.01 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion	Ratio	Lower	Upper
198	100		
182	275.6	2.0	3.0#
77	912.9	15.8	23.8#




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 77.00 (76.70 to 77.70): B






#62
 N-Nitrosodiphenylamine
 Concen: 0.02 ng/ul
 RT: 15.54 min Scan# 2188
 Delta R.T. -0.15 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion	Ratio	Lower	Upper
169	100		
168	0.0	53.8	80.8#
167	163.8	29.2	43.8#



Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E





#67

Phenanthrene

Concen: 0.01 ng/ul

RT: 17.18 min Scan# 2466

Delta R.T. -0.04 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19


Tgt Ion:178 Resp: 224

Ion Ratio Lower Upper

178 100

179 107.4 12.8 19.2#



176 0.0 16.5 24.7#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



Time-->



#69

Anthracene

Concen: 0.02 ng/ul

RT: 17.27 min Scan# 2482

Delta R.T. -0.04 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19


Tgt Ion:178 Resp: 503

Ion Ratio Lower Upper

178 100

179 112.7 12.9 19.3#



176 0.0 15.7 23.5#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E




Time-->






#70
 1,2,3,4-Tetrachlorobenzene
 Concen: 2.62 ng/uL
 RT: 13.23 min Scan# 1795
 Delta R.T. -0.00 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion:	216	Resp:	18362
Ion Ratio	Lower	Upper	
216	100		
214	76.9	64.5	96.7
218	48.4	37.9	56.9




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 218.00 (217.70 to 218.70): E






#71
 Pentachlorobenzene
 Concen: 5.22 ng/uL
 RT: 14.75 min Scan# 2053
 Delta R.T. -0.00 min
 Lab File: BG016019.D
 Acq: 20 Feb 2015 2:19

Tgt Ion:	250	Resp:	35833
Ion Ratio	Lower	Upper	
250	100		
248	60.7	50.5	75.7
252	63.7	50.7	76.1



Abundance Ion 250.00 (249.70 to 250.70): E
 Ion 248.00 (247.70 to 248.70): E
 Ion 252.00 (251.70 to 252.70): E





#73

Di-n-butylphthalate

Concen: 0.03 ng/ul


RT: 18.14 min Scan# 2631

Delta R.T. -0.00 min




Lab File: BG016019.D

Acq: 20 Feb 2015 2:19

Tgt Ion:	149	Resp:	840
Ion	Ratio	Lower	Upper
149	100		
150	0.0	7.6	11.4#
104	0.0	3.7	5.5#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E



#77

Pyrene

Concen: 0.04 ng/ul


RT: 19.57 min Scan# 2874

Delta R.T. -0.02 min



Lab File: BG016019.D


Acq: 20 Feb 2015 2:19

Tgt Ion:	202	Resp:	1504
Ion	Ratio	Lower	Upper
202	100		
101	116.4	15.8	23.8#
100	733.2	12.7	19.1#




Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 100.00 (99.70 to 100.70): E








#80
Benzo(a)anthracene
Concen: 0.05 ng/ul
RT: 21.35 min Scan# 3177
Delta R.T. 0.02 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19

Tgt Ion: 228 Resp: 1589
Ion Ratio Lower Upper
228 100
229 23.9 16.4 24.6
226 0.0 22.8 34.2#




Abundance Ion 228.00 (227.70 to 228.70): E
Ion 229.00 (228.70 to 229.70): E
Ion 226.00 (225.70 to 226.70): E






#81
Bis(2-ethylhexyl)phthalate
Concen: 0.01 ng/ul
RT: 21.27 min Scan# 3163
Delta R.T. 0.00 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19

Tgt Ion: 149 Resp: 166
Ion Ratio Lower Upper
149 100
167 65.2 25.5 38.3#
279 0.0 6.0 9.0#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E





#82

Chrysene

Concen: 0.05 ng/ul

RT: 21.35 min Scan# 3177

Delta R.T. -0.03 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19


Tgt Ion: 228 Resp: 1589

Ion Ratio Lower Upper

228 100

226 0.0 24.7 37.1#




229 23.9 17.0 25.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Di-n-octyl phthalate

Concen: 0.01 ng/ul


RT: 22.17 min Scan# 3316

Delta R.T. 0.00 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19



Tgt Ion: 149 Resp: 190




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 149.00 (148.70 to 149.70): E





Ion 149.00 (148.70 to 149.70): E








#88
Benzo(a)pyrene
Concen: 0.13 ng/ul
RT: 23.50 min Scan# 3543
Delta R.T. -0.05 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19


Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.8	17.9	26.9
125	0.0	13.2	19.8#



#89
Indeno(1,2,3-cd)pyrene
Concen: 0.02 ng/ul
RT: 25.99 min Scan# 3967
Delta R.T. -0.02 min
Lab File: BG016019.D
Acq: 20 Feb 2015 2:19

Tgt Ion	Ratio	Lower	Upper
276	100		
138	0.0	26.2	39.4#
277	32.2	21.0	31.6#





#90

Dibenzo(a,h)anthracene

Concen: 0.01 ng/ul

RT: 26.02 min Scan# 3971


Delta R.T. -0.01 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19

Tgt Ion: 278 Resp: 435




Ion	Ratio	Lower	Upper
278	100		
139	76.4	17.0	25.4#
279	40.0	19.3	28.9#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 0.01 ng/ul

RT: 26.73 min Scan# 4093


Delta R.T. 0.00 min

Lab File: BG016019.D

Acq: 20 Feb 2015 2:19

Tgt Ion: 276 Resp: 275

Ion	Ratio	Lower	Upper
276	100		
138	0.0	23.0	34.4#
277	0.0	20.0	30.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

