


Data Path : Z:\HPCHEM1\BNA G\Data\BG021916\
 Data File : BG020968.D
 Acq On : 19 Feb 2016 12:21
 Operator : SJ/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG021116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Feb 11 15:26:53 2016



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.0	9930	PASS
68	69	0.00	2	0.8	99	PASS
69	198	0.00	100	33.8	11994	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.3	18526	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	35437	PASS
199	198	5	9	6.9	2456	PASS
275	198	10	60	24.1	8524	PASS
365	198	1	100	2.6	935	PASS
441	442	0.01	24	17.2	4101	PASS
442	198	50	100	67.3	23864	PASS
443	442	15	24	20.2	4829	PASS