


Data Path : Z:\HPCHEM1\BNA G\DATA\BG021916\
 Data File : BG020987.D
 Acq On : 20 Feb 2016 00:50
 Operator : SJ/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG021116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Feb 11 15:26:53 2016



AutoFind: Scans 2583, 2584, 2585; Background Corrected with Scan 2577

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.1	4388	PASS
68	69	0.00	2	1.0	56	PASS
69	198	0.00	100	33.0	5543	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.9	8551	PASS
197	198	0.00	2	0.3	51	PASS
198	198	100	100	100.0	16814	PASS
199	198	5	9	6.4	1080	PASS
275	198	10	60	24.5	4120	PASS
365	198	1	100	3.0	504	PASS
441	442	0.01	24	16.9	2121	PASS
442	198	50	100	74.6	12541	PASS
443	442	15	24	18.9	2368	PASS