


Data Path : Z:\HPCHEM1\BNA G\DATA\BG021916\
 Data File : BG021005.D
 Acq On : 20 Feb 2016 12:37
 Operator : SJ/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG021116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Feb 11 15:26:53 2016



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2577

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.3	5220	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	36.0	6405	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.7	9370	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	17789	PASS
199	198	5	9	7.1	1266	PASS
275	198	10	60	23.8	4238	PASS
365	198	1	100	2.7	486	PASS
441	442	0.01	24	16.8	1996	PASS
442	198	50	100	66.9	11902	PASS
443	442	15	24	20.1	2390	PASS