

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG022017\
 Data File : BG025881.D
 Acq On : 20 Feb 2017 20:06
 Operator : SJ/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02023

Quant Time: Feb 21 01:12:12 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.3-EPA-BG021617MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 21 01:07:13 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
3	Naphthalene	0.986	1.026	-4.1	122	0.00
4	2-Methylnaphthalene	0.777	0.804	-3.5	120	0.00
5	1-Methylnaphthalene	0.758	0.783	-3.3	120	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
7 S	Acenaphthylene-d8	1.656	1.779	-7.4	118	0.00
8	Acenaphthylene	1.697	1.816	-7.0	118	0.00
9 C	Acenaphthene	1.212	1.295	-6.8	120	0.00
10 S	Fluorene-d10	1.272	1.297	-2.0	114	0.00
11	Fluorene	1.423	1.476	-3.7	116	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
13	Phenanthrene	1.067	1.120	-5.0	113	0.00
14 S	Anthracene-d10	0.904	0.945	-4.5	112	0.00
15	Anthracene	1.084	1.155	-6.5	113	0.00
16 C	Fluoranthene	1.172	1.227	-4.7	110	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
18 S	Pyrene-d10	0.964	1.022	-6.0	109	0.00
19	Pyrene	1.210	1.295	-7.0	109	0.00
20	Benzo(a)anthracene	1.125	1.195	-6.2	110	0.00
21	Chrysene	1.075	1.136	-5.7	109	0.00
22 I	Perylene-d12	1.000	1.000	0.0	112	-0.01
23	Benzo(b)fluoranthene	1.156	1.185	-2.5	107	-0.01
24	Benzo(k)fluoranthene	1.116	1.159	-3.9	112	-0.01
25 S	Benzo(a)pyrene-d12	0.885	0.918	-3.7	109	-0.01
26 C	Benzo(a)pyrene	1.118	1.158	-3.6	109	-0.01
27	Indeno(1,2,3-cd)pyrene	1.310	1.352	-3.2	112	-0.02
28	Dibenzo(a,h)anthracene	1.085	1.114	-2.7	110	-0.02
29	Benzo(g,h,i)perylene	1.081	1.108	-2.5	109	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0