

Data Path : Z:\HPCHEM1\BNA G\Data\BG022117\
 Data File : BG025931.D
 Acq On : 22 Feb 2017 8:28
 Operator : SJ/MA
 Sample : I1901-02
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DABE8

Quant Time: Feb 22 10:43:33 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.3-EPA-BG021617MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 04:47:03 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.09	152	166812	20.00	ng/ul	0.00
2) Naphthalene-d8	10.89	136	753734	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.69	164	420908	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.42	188	844426	20.00	ng/ul	0.00
17) Chrysene-d12	21.67	240	839719	20.00	ng/ul	0.00
22) Perylene-d12	24.90	264	804619	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
7) Acenaphthylene-d8	14.38	160	531999	15.26	ng/ul	0.00
10) Fluorene-d10	15.68	176	384314	14.36	ng/ul	0.00
14) Anthracene-d10	17.52	188	538164	14.10	ng/ul	0.00
18) Pyrene-d10	19.79	212	542425	13.40	ng/ul	0.00
25) Benzo(a)pyrene-d12	24.68	264	490878	13.79	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Phenanthrene	17.46	178	49691	1.10	ng/ul	96
15) Anthracene	17.46	178	50097	1.09	ng/ul	95
16) Fluoranthene	19.47	202	101265	2.05	ng/ul	99
19) Pyrene	19.82	202	92308	1.82	ng/ul	98
20) Benzo(a)anthracene	21.66	228	54743	1.16	ng/ul#	85
21) Chrysene	21.72	228	67294	1.49	ng/ul#	93
23) Benzo(b)fluoranthene	23.88	252	95387	2.05	ng/ul#	80
24) Benzo(k)fluoranthene	23.88	252	95387	2.12	ng/ul#	80
26) Benzo(a)pyrene	24.75	252	58435	1.30	ng/ul#	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\Data\BG022117\
Data File : BG025931.D
Acq On : 22 Feb 2017 8:28
Operator : SJ/MA
Sample : I1901-02
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DABE8

Quant Time: Feb 22 10:43:33 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.3-EPA-BG021617MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 22 04:47:03 2017
Response via : Initial Calibration

