

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022117\
 Data File : BG025897.D
 Acq On : 21 Feb 2017 9:15
 Operator : SJ/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02025

Quant Time: Feb 21 10:42:32 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.3-EPA-BG021617MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 21 02:22:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	155#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	152#	0.00
3	Naphthalene	0.986	1.015	-2.9	152#	0.00
4	2-Methylnaphthalene	0.777	0.786	-1.2	148	0.00
5	1-Methylnaphthalene	0.758	0.768	-1.3	148	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	138	0.00
7 S	Acenaphthylene-d8	1.656	1.779	-7.4	142	0.00
8	Acenaphthylene	1.697	1.796	-5.8	140	0.00
9 C	Acenaphthene	1.212	1.267	-4.5	141	0.00
10 S	Fluorene-d10	1.272	1.277	-0.4	135	0.00
11	Fluorene	1.423	1.413	0.7	133	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
13	Phenanthrene	1.067	1.106	-3.7	123	0.00
14 S	Anthracene-d10	0.904	0.959	-6.1	125	0.00
15	Anthracene	1.084	1.142	-5.4	123	0.00
16 C	Fluoranthene	1.172	1.165	0.6	115	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
18 S	Pyrene-d10	0.964	0.992	-2.9	116	0.00
19	Pyrene	1.210	1.230	-1.7	113	0.00
20	Benzo(a)anthracene	1.125	1.191	-5.9	120	0.00
21	Chrysene	1.075	1.112	-3.4	117	0.00
22 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
23	Benzo(b)fluoranthene	1.156	1.200	-3.8	119	-0.01
24	Benzo(k)fluoranthene	1.116	1.127	-1.0	119	0.00
25 S	Benzo(a)pyrene-d12	0.885	0.929	-5.0	121	-0.01
26 C	Benzo(a)pyrene	1.118	1.150	-2.9	119	-0.01
27	Indeno(1,2,3-cd)pyrene	1.310	1.368	-4.4	124	-0.01
28	Dibenzo(a,h)anthracene	1.085	1.134	-4.5	122	-0.01
29	Benzo(g,h,i)perylene	1.081	1.117	-3.3	121	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0