

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022117\
 Data File : BG025907.D
 Acq On : 21 Feb 2017 16:34
 Operator : SJ/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02026

Quant Time: Feb 21 17:36:36 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.3-EPA-BG021617MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 21 12:19:58 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	150	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	148	0.00
3	Naphthalene	0.986	1.019	-3.3	150	0.00
4	2-Methylnaphthalene	0.777	0.799	-2.8	147	0.00
5	1-Methylnaphthalene	0.758	0.770	-1.6	146	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	138	0.00
7 S	Acenaphthylene-d8	1.656	1.771	-6.9	141	0.00
8	Acenaphthylene	1.697	1.796	-5.8	140	0.00
9 C	Acenaphthene	1.212	1.249	-3.1	139	0.00
10 S	Fluorene-d10	1.272	1.283	-0.9	136	0.00
11	Fluorene	1.423	1.427	-0.3	134	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
13	Phenanthrene	1.067	1.109	-3.9	126	0.00
14 S	Anthracene-d10	0.904	0.946	-4.6	126	0.00
15	Anthracene	1.084	1.135	-4.7	124	0.00
16 C	Fluoranthene	1.172	1.192	-1.7	120	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
18 S	Pyrene-d10	0.964	0.972	-0.8	119	0.00
19	Pyrene	1.210	1.221	-0.9	117	0.00
20	Benzo(a)anthracene	1.125	1.192	-6.0	125	0.00
21	Chrysene	1.075	1.124	-4.6	124	0.00
22 I	Perylene-d12	1.000	1.000	0.0	127	0.00
23	Benzo(b)fluoranthene	1.156	1.185	-2.5	121	-0.01
24	Benzo(k)fluoranthene	1.116	1.179	-5.6	129	0.00
25 S	Benzo(a)pyrene-d12	0.885	0.932	-5.3	126	-0.01
26 C	Benzo(a)pyrene	1.118	1.167	-4.4	125	0.00
27	Indeno(1,2,3-cd)pyrene	1.310	1.379	-5.3	129	-0.01
28	Dibenzo(a,h)anthracene	1.085	1.146	-5.6	128	-0.01
29	Benzo(g,h,i)perylene	1.081	1.132	-4.7	127	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0