

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG022118\
 Data File : BG032755.D
 Acq On : 21 Feb 2018 10:55
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

Sohil
 2/22/2018 3:13:41 PM

Quant Time: Feb 21 15:33:44 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 21 13:32:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.93	152	52431	20.00	ng	0.00
21) Naphthalene-d8	11.82	136	242260	20.00	ng	0.00
38) Acenaphthene-d10	15.55	164	143424	20.00	ng	0.00
63) Phenanthrene-d10	18.27	188	336238	20.00	ng	0.00
75) Chrysene-d12	22.63	240	318957	20.00	ng	0.00
86) Perylene-d12	26.45	264	319991	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	6.34	112	14868	5.69	ng	0.00
7) Phenol-d6	8.02	99	21046	5.89	ng	0.00
23) Nitrobenzene-d5	10.13	82	8304	2.18	ng	0.00
41) 2,4,6-Tribromophenol	0.00	330	0d	0.00	ng	
44) 2-Fluorobiphenyl	14.18	172	50900	4.44	ng	0.00
78) Terphenyl-d14	20.79	244	75743	5.41	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.98	88	3752	3.432	ng	# 79
3) Pyridine	4.45	79	8013	2.858	ng	# 89
4) n-Nitrosodimethylamine	4.34	42	3663	2.780	ng	# 59
6) Aniline	8.22	93	14881	3.524	ng	# 94
8) 2-Chlorophenol	8.48	128	8063	2.683	ng	# 88
10) Phenol	8.05	94	10879	3.130	ng	# 90
11) bis(2-Chloroethyl)ether	8.31	93	9098	3.265	ng	# 95
12) 1,3-Dichlorobenzene	8.82	146	10309	2.432	ng	# 96
13) 1,4-Dichlorobenzene	8.97	146	10394	2.482	ng	# 95
14) 1,2-Dichlorobenzene	9.30	146	10245	2.426	ng	# 96
15) Benzyl Alcohol	9.16	79	7549	2.819	ng	# 96
16) 2,2'-oxybis(1-Chloropropan	9.46	45	16305	5.438	ng	# 99
17) 2-Methylphenol	9.36	107	8017	3.257	ng	# 88
18) Hexachloroethane	10.06	117	3374	2.207	ng	# 76
19) n-Nitroso-di-n-propylamine	9.74	70	7667	3.192	ng	# 98
20) 3+4-Methylphenols	9.69	107	11169	3.003	ng	# 97
22) Acetophenone	9.77	105	14662	2.846	ng	# 96
25) Isophorone	10.70	82	20829	2.866	ng	# 97
27) 2,4-Dimethylphenol	10.92	122	10036	3.579	ng	# 87
28) bis(2-Chloroethoxy)methane	11.18	93	12325	3.317	ng	# 94
31) Naphthalene	11.87	128	31929	2.714	ng	# 98
33) 4-Chloroaniline	11.96	127	13659	2.926	ng	# 91
35) Caprolactam	12.66	113	2733	2.345	ng	# 90
36) 4-Chloro-3-methylphenol	13.00	107	8461	2.292	ng	# 93
37) 2-Methylnaphthalene	13.42	142	22891	2.399	ng	# 93
45) 1,1'-Biphenyl	14.39	154	31632	2.680	ng	# 98
46) 2-Chloronaphthalene	14.44	162	23401	2.506	ng	# 98
48) Acenaphthylene	15.27	152	37243	2.711	ng	# 98
49) Dimethylphthalate	14.97	163	30613	2.469	ng	# 98
51) Acenaphthene	15.61	154	23424	2.497	ng	# 94
54) Dibenzofuran	15.94	168	34889	2.499	ng	# 96
57) Fluorene	16.58	166	29248	2.548	ng	# 99
59) Diethylphthalate	16.31	149	32165	2.666	ng	# 98
62) Azobenzene	16.85	77	29334	3.785	ng	# 95

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG022118\
 Data File : BG032755.D
 Acq On : 21 Feb 2018 10:55
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

Sohil
 2/22/2018 3:13:41 PM

Quant Time: Feb 21 15:33:44 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 21 13:32:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
65) n-Nitrosodiphenylamine	16.76	169	25602	2.596	nq	96
68) Atrazine	17.68	200	8505	2.011	nq	93
70) Phenanthrene	18.31	178	47051	2.605	nq	99
71) Anthracene	18.40	178	47332	2.561	nq	99
72) Carbazole	18.66	167	47404	3.008	nq	99
73) Di-n-butylphthalate	19.17	149	53992	3.043	nq	98
74) Fluoranthene	20.27	202	53687	2.226	nq	94
76) Benzidine	20.42	184	31233	4.159	nq	100
77) Pyrene	20.62	202	58050	3.059	nq	98
79) Butylbenzylphthalate	21.50	149	19318	3.090	nq	95
80) Benzo(a)anthracene	22.61	228	51754	2.592	nq	96
81) 3,3'-Dichlorobenzidine	22.49	252	17043	2.285	nq #	95
82) Chrysene	22.69	228	48490	2.462	nq	97
83) Bis(2-ethylhexyl)phthalate	22.47	149	30520	3.386	nq	97
84) Di-n-octyl phthalate	23.90	149	40567	2.868	nq #	95
85) Indeno(1,2,3-cd)pyrene	30.89	276	49048m	2.103	nq	
87) Benzo(b)fluoranthene	25.22	252	45751	2.399	nq #	92
88) Benzo(k)fluoranthene	25.29	252	46561	2.359	nq #	95
89) Benzo(a)pyrene	26.27	252	42978	2.313	nq	97
90) Dibenzo(a,h)anthracene	30.97	278	41714	2.255	nq #	93
91) Benzo(g,h,i)perylene	32.26	276	41218	2.305	nq #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA G\DATA\BG022118\
Data File : BG032755.D
Acq On : 21 Feb 2018 10:55
Operator : SJ/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM

Quant Time: Feb 21 15:33:44 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 21 13:32:56 2018
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.93 min Scan# 961
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
152	100		
150	159.8	126.6	189.8
115	63.2	53.0	79.4

Manual Integrations
APPROVED




Sohil
2/22/2018 3:13:41 PM




#2

1,4-Dioxane
Concen: 3.432 ng
RT: 3.98 min Scan# 117
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
88	100		
58	75.8	79.6	119.4#
43	42.0	27.4	41.0#





#3

Pvridine

Concen: 2.858 ng

RT: 4.45 min Scan# 198

Delta R.T. 0.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	79	Resp:	8013
Ion	Ratio	Lower	Upper
79	100		
52	91.6	69.9	104.9
51	58.3	34.0	51.0#

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#4

n-Nitrosodimethylamine

Concen: 2.780 ng




RT: 4.34 min Scan# 179


Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion:	42	Resp:	3663
Ion	Ratio	Lower	Upper
42	100		
74	94.7	102.5	153.7#
44	75.3	18.9	28.3#





#5

2-Fluorophenol

Concen: 5.689 ng

RT: 6.34 min Scan# 519

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampled :

SSTDICC2.5

Tot Ion:	112	Resp:	14868
Ion Ratio	Lower	Upper	
112	100		
64	64.6	56.3	84.5
63	33.7	30.1	45.1

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#6

Aniline

Concen: 3.524 ng




RT: 8.22 min Scan# 839


Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion:	93	Resp:	14881
Ion Ratio	Lower	Upper	
93	100		
66	39.1	33.7	50.5
65	16.0	16.1	24.1#





#7

Phenol-d6

Concen: 5.890 ng

RT: 8.02 min Scan# 805

Delta R.T. -0.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G


ClientSampled :

SSTDICC2.5




Tot Ion:	99	Resp:	21046
Ion	Ratio	Lower	Upper
99	100		
42	19.7	14.1	21.1
71	34.4	26.1	39.1

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 2.683 ng


RT: 8.48 min Scan# 883

Delta R.T. -0.00 min



Lab File: BG032755.D


Acq: 21 Feb 2018 10:55

Tgt Ion:	128	Resp:	8063
Ion	Ratio	Lower	Upper
128	100		
130	25.9	14.0	54.0
64	49.9	37.7	77.7



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 130.00 (129.70 to 130.70): BGC
Ion 64.00 (63.70 to 64.70): BGC





#10

Phenol

Concen: 3.130 ng

RT: 8.05 min Scan# 810

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	94	Resp:	10879
Ion	Ratio	Lower	Upper
94	100		
65	25.3	11.9	51.9
66	37.1	22.2	62.2

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#11

bis(2-Chloroethyl)ether

Concen: 3.265 ng




RT: 8.31 min Scan# 855


Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion:	93	Resp:	9098
Ion	Ratio	Lower	Upper
93	100		
63	83.3	67.1	107.1
95	35.4	11.5	51.5









#12
 1,3-Dichlorobenzene
 Concen: 2.432 ng
 RT: 8.82 min Scan# 941
 Delta R.T. -0.00 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
146	100		
148	67.5	51.4	77.2
75	31.7	26.6	39.8




Manual Integrations
 APPROVED


Sohil
 2/22/2018 3:13:41 PM



#13
 1,4-Dichlorobenzene
 Concen: 2.482 ng
 RT: 8.97 min Scan# 967
 Delta R.T. -0.00 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
146	100		
148	61.6	53.7	80.5
111	43.3	35.2	52.8






#14
 1,2-Dichlorobenzene
 Concen: 2.426 ng
 RT: 9.30 min Scan# 1024
 Delta R.T. -0.00 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5




Tot Ion	Ratio	Lower	Upper
146	100		
148	64.6	49.3	73.9
111	40.9	34.3	51.5

Manual Integrations
 APPROVED

Sohil
 2/22/2018 3:13:41 PM




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#15
 Benzyl Alcohol
 Concen: 2.819 ng
 RT: 9.16 min Scan# 1000
 Delta R.T. -0.00 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
79	100		
108	71.6	57.2	85.8
77	63.5	46.1	69.1



Abundance Ion 79.00 (78.70 to 79.70): BGC
 Ion 108.00 (107.70 to 108.70): E
 Ion 77.00 (76.70 to 77.70): BGC









#16
2,2'-oxvbis(1-Chloropropane)
Concen: 5.438 ng
RT: 9.46 min Scan# 1051
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
45	100		
77	10.8	0.0	30.2
79	7.5	0.0	27.9




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#17
2-Methylphenol
Concen: 3.257 ng
RT: 9.36 min Scan# 1033
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
107	100		
108	123.8	83.9	125.9
77	46.1	37.2	55.8
79	38.6	36.2	54.2









#18
Hexachloroethane
Concen: 2.207 ng
RT: 10.06 min Scan# 1153
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
117	100		
119	87.6	76.7	115.1
201	79.4	95.7	143.5#




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#19
n-Nitroso-di-n-propylamine
Concen: 3.192 ng
RT: 9.74 min Scan# 1098
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
70	100		
42	62.6	49.1	73.7
101	8.8	8.5	12.7
130	18.3	13.4	20.0





#20

3+4-Methylphenols

Concen: 3.003 ng

RT: 9.69 min Scan# 1089

Delta R.T. -0.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5

Tot Ion:	107	Resp:	11169
Ion	Ratio	Lower	Upper
107	100		
108	79.5	61.6	101.6
77	29.3	13.9	53.9
79	28.4	8.8	48.8

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 11.82 min Scan# 1452

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion:	136	Resp:	242260
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.2	13.8
54	9.5	10.3	15.5#
68	6.5	4.5	6.7




Abundance


Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC





#22

Acetophenone

Concen: 2.846 ng

RT: 9.77 min Scan# 1104

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	105	Resp:	14662
Ion Ratio	Lower	Upper	
105	100		
71	0.0	3.0	4.4#
51	32.1	26.1	39.1
120	25.9	17.4	26.2

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#23

Nitrobenzene-d5

Concen: 2.178 ng




RT: 10.13 min Scan# 1164


Delta R.T. -0.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion:	82	Resp:	8304
Ion Ratio	Lower	Upper	
82	100		
128	42.2	29.7	44.5
54	55.5	43.1	64.7









#25
Isophorone
Concen: 2.866 ng
RT: 10.70 min Scan# 1262
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
82	100	20829		
95	6.6		6.2	9.2
138	16.3		12.1	18.1




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#27
2,4-Dimethylphenol
Concen: 3.579 ng
RT: 10.92 min Scan# 1300
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Resp	Lower	Upper
122	100	10036		
107	106.7		100.2	150.2
121	58.6		44.4	66.6









#28
bis(2-Chloroethoxy)methane
Concen: 3.317 ng
RT: 11.18 min Scan# 1343
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
93	100		
95	28.2	24.3	36.5
123	15.1	8.4	12.6




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#31
Naphthalene
Concen: 2.714 ng
RT: 11.87 min Scan# 1461
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.5	8.6	12.8
127	13.8	11.6	17.4









#33
4-Chloroaniline
Concen: 2.926 ng
RT: 11.96 min Scan# 1476
Delta R.T. 0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ion	Ratio	Resp	Lower	Upper
127	100				
129	26.7	24.0	36.0		
65	27.5	27.0	40.4		
92	15.3	16.6	25.0		




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#35
Caprolactam
Concen: 2.345 ng
RT: 12.66 min Scan# 1595
Delta R.T. -0.04 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
113	100				
55	193.7	186.9	226.9		
56	135.9	130.9	170.9		






#36
 4-Chloro-3-methylphenol
 Concen: 2.292 ng
 RT: 13.00 min Scan# 1654
 Delta R.T. -0.01 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
107	100		
144	29.4	18.2	27.4#
142	77.7	59.0	88.4

Manual Integrations
 APPROVED

Sohil
 2/22/2018 3:13:41 PM






Abundance

Ion 107.00 (106.70 to 107.70): E


Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E



#37
 2-Methylnaphthalene
 Concen: 2.399 ng
 RT: 13.42 min Scan# 1724
 Delta R.T. -0.00 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.7	67.1	100.7
115	32.6	30.7	46.1






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E





#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 15.55 min Scan# 2087

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion	Ion	Ratio	Resp	Lower	Upper
164	100				
162	108.6	78.8	118.2		
160	46.9	35.4	53.0		

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#41

2,4,6-Tribromophenol

Concen: 0.000 ng

Expected RT: 17.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55


Tgt Ion: 330


Sig Exp Ratio

330 100

332 96.3

141 31.5





#44

2-Fluorobiphenyl

Concen: 4.438 ng

RT: 14.18 min Scan# 1854

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	172	Resp:	50900
Ion Ratio	Lower	Upper	
172	100		
171	37.2	30.0	45.0
170	23.6	19.5	29.3

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#45

1,1'-Biphenyl

Concen: 2.680 ng




RT: 14.39 min Scan# 1890


Delta R.T. 0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion:	154	Resp:	31632
Ion Ratio	Lower	Upper	
154	100		
153	40.9	19.8	59.8
76	12.6	0.0	31.9






#46
2-Chloronaphthalene
Concen: 2.506 ng
RT: 14.44 min Scan# 1899
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
162	100		
127	39.2	30.7	46.1
164	30.9	26.0	39.0

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM






Abundance

Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E



#48
Acenaphthylene
Concen: 2.711 ng
RT: 15.27 min Scan# 2040
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55




Tgt Ion	Ratio	Lower	Upper
152	100		
151	22.3	17.2	25.8
153	12.1	10.2	15.4


Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 2.469 ng
RT: 14.97 min Scan# 1988
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampled :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
163	100	30613		
194	5.0		3.4	5.2
164	10.9		8.2	12.4




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#51
Acenaphthene
Concen: 2.497 ng
RT: 15.61 min Scan# 2097
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	23424		
153	121.5		90.4	135.6
152	53.5		41.6	62.4





#54

Dibenzofuran

Concen: 2.499 ng

RT: 15.94 min Scan# 2153

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:168	Resp:	34889
Ion Ratio	Lower	Upper
168	100	
139	38.7	28.6 43.0
169	14.1	11.2 16.8

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#57

Fluorene

Concen: 2.548 ng




RT: 16.58 min Scan# 2262


Delta R.T. -0.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion:166	Resp:	29248
Ion Ratio	Lower	Upper
166	100	
165	99.8	80.6 121.0
167	14.3	10.4 15.6









#59
Diethylphthalate
Concen: 2.666 ng
RT: 16.31 min Scan# 2216
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
149	100		
177	21.9	18.5	27.7
150	12.9	10.2	15.2




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#62
Azobenzene
Concen: 3.785 ng
RT: 16.85 min Scan# 2308
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
77	100		
182	21.9	7.4	47.4
105	19.5	0.3	40.3
51	30.0	11.6	51.6





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 18.27 min Scan# 2551

Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5




Tot Ion:	188	Resp:	336238
Ion Ratio	Lower	Upper	
188	100		
94	10.2	6.9	10.3
80	11.5	7.7	11.5

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



Abundance Ion 188.00 (187.70 to 188.70): E



#65

n-Nitrosodiphenylamine

Concen: 2.596 ng


RT: 16.76 min Scan# 2293

Delta R.T. -0.01 min



Lab File: BG032755.D


Acq: 21 Feb 2018 10:55

Tgt Ion:	169	Resp:	25602
Ion Ratio	Lower	Upper	
169	100		
168	72.0	55.7	83.5
167	40.7	30.1	45.1



Abundance Ion 169.00 (168.70 to 169.70): E









#68
Atrazine
Concen: 2.011 ng
RT: 17.68 min Scan# 2450
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
200	100		
173	27.6	5.2	45.2
215	44.7	30.4	70.4




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#70
Phenanthrene
Concen: 2.605 ng
RT: 18.31 min Scan# 2558
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.3	15.7	23.5
179	15.1	12.5	18.7









#71
 Anthracene
 Concen: 2.561 ng
 RT: 18.40 min Scan# 2573
 Delta R.T. -0.01 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
178	100		
176	20.4	16.0	24.0
179	15.1	12.5	18.7




Manual Integrations
 APPROVED


Sohil
 2/22/2018 3:13:41 PM



#72
 Carbazole
 Concen: 3.008 ng
 RT: 18.66 min Scan# 2616
 Delta R.T. -0.00 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
167	100		
166	22.8	18.2	27.4
139	12.8	9.4	14.2









#73
Di-n-butylphthalate
Concen: 3.043 ng
RT: 19.17 min Scan# 2703
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampled :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.8	11.6
104	5.3	3.9	5.9




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#74
Fluoranthene
Concen: 2.226 ng
RT: 20.27 min Scan# 2890
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
202	100		
101	14.4	0.0	28.9
203	17.0	0.0	37.5





#75

Chrysene-d12

Concen: 20.000 ng

RT: 22.63 min Scan# 3293

Delta R.T. -0.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
240	100		
120	13.0	6.8	10.2#
236	26.5	20.9	31.3

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



#76

Benzidine

Concen: 4.159 ng




RT: 20.42 min Scan# 2916


Delta R.T. -0.00 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.1	16.7
183	13.4	10.6	16.0









#77
Pvrene
Concen: 3.059 ng
RT: 20.62 min Scan# 2950
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampled :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
202	100		
200	19.7	16.9	25.3
203	18.3	13.9	20.9




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#78
Terphenyl-d14
Concen: 5.412 ng
RT: 20.79 min Scan# 2979
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.8	8.8
122	13.5	7.0	10.4









#79
Butylbenzylphthalate
Concen: 3.090 ng
RT: 21.50 min Scan# 3100
Delta R.T. -0.00 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
149	100		
91	74.1	62.2	93.2
206	20.0	18.6	27.8




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#80
Benzo(a)anthracene
Concen: 2.592 ng
RT: 22.61 min Scan# 3289
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.3	22.7	34.1
229	18.5	16.2	24.2









#81
 3,3'-Dichlorobenzidine
 Concen: 2.285 ng
 RT: 22.49 min Scan# 3269
 Delta R.T. -0.01 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Instrument :
 BNA_G
 ClientSampled :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
252	100		
254	67.0	50.8	76.2
126	13.1	7.4	11.2




Manual Integrations
 APPROVED


Sohil
 2/22/2018 3:13:41 PM



#82
 Chrysene
 Concen: 2.462 ng
 RT: 22.69 min Scan# 3302
 Delta R.T. -0.01 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.6	24.9	37.3
229	20.8	15.0	22.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 3.386 ng
 RT: 22.47 min Scan# 3265
 Delta R.T. -0.00 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Instrument :
 BNA_G
 ClientSampled :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
149	100		
167	28.3	24.3	36.5
279	4.7	4.0	6.0




Manual Integrations
 APPROVED


Sohil
 2/22/2018 3:13:41 PM



#84
 Di-n-octyl phthalate
 Concen: 2.868 ng
 RT: 23.90 min Scan# 3509
 Delta R.T. -0.01 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
149	100		
167	1.2	1.4	2.0#
43	13.1	8.9	13.3





#85

Indeno(1,2,3-cd)pyrene

Concen: 2.103 ng

RT: 30.89 min Scan# 4698

Delta R.T. -0.04 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G


ClientSampled :

SSTDICC2.5




Tot Ion:	276	Resp:	49048
Ion Ratio	Lower	Upper	
276	100		
138	20.5	16.0	24.0
277	22.2	20.0	30.0

Manual Integrations
APPROVED

Sohil
2/22/2018 3:13:41 PM



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.000 ng


RT: 26.45 min Scan# 3943

Delta R.T. -0.01 min



Lab File: BG032755.D


Acq: 21 Feb 2018 10:55

Tgt Ion:	264	Resp:	319991
Ion Ratio	Lower	Upper	
264	100		
260	22.5	17.4	26.0
265	22.7	17.7	26.5



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 2.399 ng

RT: 25.22 min Scan# 3734

Delta R.T. -0.01 min

Lab File: BG032755.D

Acq: 21 Feb 2018 10:55

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5


Tot Ion:	252	Resp:	45751
Ion Ratio	Lower	Upper	
252	100		
253	18.8	18.2	27.4
125	12.0	7.2	10.8#

Manual Integrations
APPROVED



Sohil
2/22/2018 3:13:41 PM



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



Time--> 25.10 25.15 25.20 25.25



#88

Benzo(k)fluoranthene

Concen: 2.359 ng

RT: 25.29 min Scan# 3746


Delta R.T. -0.03 min

Lab File: BG032755.D



Acq: 21 Feb 2018 10:55


Tgt Ion:	252	Resp:	46561
Ion Ratio	Lower	Upper	
252	100		
253	22.8	17.4	26.2
125	12.2	6.6	9.8#

Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



Time--> 25.25 25.30 25.35









#89
Benzo(a)pyrene
Concen: 2.313 ng
RT: 26.27 min Scan# 3912
Delta R.T. -0.01 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.9	18.3	27.5
125	10.4	7.3	10.9




Manual Integrations
APPROVED


Sohil
2/22/2018 3:13:41 PM



#90
Dibenzo(a,h)anthracene
Concen: 2.255 ng
RT: 30.97 min Scan# 4712
Delta R.T. -0.06 min
Lab File: BG032755.D
Acq: 21 Feb 2018 10:55

Tgt Ion	Ratio	Lower	Upper
278	100		
139	19.0	9.8	14.6#
279	23.9	18.4	27.6





#91
 Benzo(a,h,i)perylene
 Concen: 2.305 ng
 RT: 32.26 min Scan# 4932
 Delta R.T. -0.06 min
 Lab File: BG032755.D
 Acq: 21 Feb 2018 10:55

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
276	100		
277	21.0	18.3	27.5
138	22.0	13.9	20.9

Manual Integrations
 APPROVED

Sohil
 2/22/2018 3:13:41 PM

