

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022122\
 Data File : BG052430.D
 Acq On : 21 Feb 2022 10:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 21 13:56:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	41.152	-2.9	107	0.00
3	Pyridine	40.000	40.520	-1.3	98	0.00
4	n-Nitrosodimethylamine	40.000	36.904	7.7	92	0.00
5 S	2-Fluorophenol	80.000	80.277	-0.3	99	0.00
6	Aniline	40.000	38.395	4.0	92	0.00
7 S	Phenol-d6	80.000	76.659	4.2	93	0.00
8	2-Chlorophenol	40.000	37.629	5.9	95	0.00
9	Benzaldehyde	40.000	36.224	9.4	90	0.00
10 C	Phenol	40.000	38.994	2.5	95	0.00
11	bis(2-Chloroethyl)ether	40.000	37.364	6.6	95	0.00
12	1,3-Dichlorobenzene	40.000	38.829	2.9	97	0.00
13 C	1,4-Dichlorobenzene	40.000	39.098	2.3	98	0.00
14	1,2-Dichlorobenzene	40.000	38.328	4.2	96	0.00
15	Benzyl Alcohol	40.000	37.550	6.1	89	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.673	8.3	94	0.00
17	2-Methylphenol	40.000	38.672	3.3	95	0.00
18	Hexachloroethane	40.000	38.894	2.8	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.347	11.6	88	0.00
20	3+4-Methylphenols	40.000	38.314	4.2	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	36.960	7.6	92	0.00
23 S	Nitrobenzene-d5	80.000	74.600	6.8	92	0.00
24	Nitrobenzene	40.000	37.083	7.3	90	0.00
25	Isophorone	40.000	36.895	7.8	90	0.00
26 C	2-Nitrophenol	40.000	37.549	6.1	89	0.00
27	2,4-Dimethylphenol	40.000	38.409	4.0	93	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.052	7.4	91	0.00
29 C	2,4-Dichlorophenol	40.000	38.287	4.3	94	0.00
30	1,2,4-Trichlorobenzene	40.000	37.706	5.7	94	0.00
31	Naphthalene	40.000	38.076	4.8	94	0.00
32	Benzoic acid	40.000	32.663	18.3	80	0.00
33	4-Chloroaniline	40.000	38.555	3.6	92	0.00
34 C	Hexachlorobutadiene	40.000	37.192	7.0	94	0.00
35	Caprolactam	40.000	38.778	3.1	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.912	2.7	93	0.00
37	2-Methylnaphthalene	40.000	38.276	4.3	95	0.00
38	1-Methylnaphthalene	40.000	38.090	4.8	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.122	7.2	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.784	13.0	87	0.00
42 S	2,4,6-Tribromophenol	80.000	76.872	3.9	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.345	6.6	90	0.00
44	2,4,5-Trichlorophenol	40.000	36.853	7.9	90	0.00
45 S	2-Fluorobiphenyl	80.000	75.650	5.4	94	0.00
46	1,1'-Biphenyl	40.000	38.193	4.5	94	0.00
47	2-Chloronaphthalene	40.000	37.837	5.4	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022122\
 Data File : BG052430.D
 Acq On : 21 Feb 2022 10:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 21 13:56:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.340	4.1	93	0.00
49	Acenaphthylene	40.000	38.175	4.6	95	0.00
50	Dimethylphthalate	40.000	37.613	6.0	94	0.00
51	2,6-Dinitrotoluene	40.000	37.748	5.6	92	0.00
52 C	Acenaphthene	40.000	37.912	5.2	93	0.00
53	3-Nitroaniline	40.000	39.754	0.6	96	0.00
54 P	2,4-Dinitrophenol	40.000	33.551	16.1	82	0.00
55	Dibenzofuran	40.000	37.766	5.6	95	0.00
56 P	4-Nitrophenol	40.000	38.595	3.5	91	0.00
57	2,4-Dinitrotoluene	40.000	38.870	2.8	94	0.00
58	Fluorene	40.000	39.019	2.5	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.486	3.8	95	0.00
60	Diethylphthalate	40.000	38.126	4.7	95	0.00
61	4-Chlorophenyl-phenylether	40.000	37.652	5.9	95	0.00
62	4-Nitroaniline	40.000	40.643	-1.6	98	0.00
63	Azobenzene	40.000	38.062	4.8	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.522	3.7	91	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.183	4.5	95	0.00
67	4-Bromophenyl-phenylether	40.000	37.901	5.2	95	0.00
68	Hexachlorobenzene	40.000	37.261	6.8	94	0.00
69	Atrazine	40.000	38.646	3.4	96	0.00
70 C	Pentachlorophenol	40.000	34.591	13.5	84	0.00
71	Phenanthrene	40.000	38.153	4.6	97	0.00
72	Anthracene	40.000	38.175	4.6	97	0.00
73	Carbazole	40.000	38.941	2.6	99	0.00
74	Di-n-butylphthalate	40.000	39.348	1.6	99	0.00
75 C	Fluoranthene	40.000	38.362	4.1	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	37.342	6.6	86	0.00
78	Pyrene	40.000	38.298	4.3	99	0.00
79 S	Terphenyl-d14	80.000	74.444	6.9	99	0.00
80	Butylbenzylphthalate	40.000	40.215	-0.5	103	0.00
81	Benzo(a)anthracene	40.000	38.026	4.9	100	0.00
82	3,3'-Dichlorobenzidine	40.000	37.252	6.9	95	0.00
83	Chrysene	40.000	37.346	6.6	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.378	1.6	98	-0.01
85 c	Di-n-octyl phthalate	40.000	40.140	-0.4	102	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.773	5.6	101	0.00
88	Benzo(b)fluoranthene	40.000	37.875	5.3	102	0.00
89	Benzo(k)fluoranthene	40.000	37.302	6.7	99	0.00
90 C	Benzo(a)pyrene	40.000	37.603	6.0	100	0.00
91	Dibenzo(a,h)anthracene	40.000	37.578	6.1	101	-0.02
92	Benzo(g,h,i)perylene	40.000	37.802	5.5	101	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022122\
Data File : BG052430.D
Acq On : 21 Feb 2022 10:06
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Feb 21 13:56:17 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 08 01:21:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0