


Data Path : Z:\HPCHEM1\BNA G\DATA\BG022216\  
 Data File : BG021016.D  
 Acq On : 22 Feb 2016 13:06  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG022216.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Feb 23 14:08:44 2016



AutoFind: Scans 2581, 2582, 2583; Background Corrected with Scan 2575

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 33.3      | 3931    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 38.4      | 4536    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 55.0      | 6488    | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 51      | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 11806   | PASS             |
| 199         | 198          | 5            | 9            | 7.3       | 860     | PASS             |
| 275         | 198          | 10           | 60           | 26.3      | 3100    | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 405     | PASS             |
| 441         | 442          | 0.01         | 24           | 17.1      | 1782    | PASS             |
| 442         | 198          | 50           | 100          | 88.4      | 10437   | PASS             |
| 443         | 442          | 15           | 24           | 18.7      | 1950    | PASS             |