

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG022221\
 Data File : BG048273.D
 Acq On : 22 Feb 2021 18:03
 Operator : CG/JU
 Sample : PB134720BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK20

Quant Time: Feb 23 00:01:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG021321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 22 11:08:29 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.12	152	39352	20.00	ng/ul	0.00
18) Naphthalene-d8	10.94	136	153756	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.75	164	108198	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.49	188	259349	20.00	ng/ul	0.00
78) Chrysene-d12	21.77	240	262735	20.00	ng/ul	0.00
86) Perylene-d12	25.08	264	244049	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.52	96	4812	5.09	ng/uL	0.00
5) Phenol-d5	7.32	99	115539	34.84	ng/ul	0.01
7) Bis-(2-Chloroethyl)ether-d	7.44	67	69537	35.21	ng/ul	0.00
9) 2-Chlorophenol-d4	7.67	132	83692	35.96	ng/ul	0.00
13) 4-Methylphenol-d8	8.86	113	93032	35.77	ng/ul	0.00
19) Nitrobenzene-d5	9.30	128	42512	37.02	ng/ul	0.01
22) 2-Nitrophenol-d4	10.02	143	53768	38.70	ng/ul	0.01
26) 2,4-Dichlorophenol-d3	10.58	165	93487	34.36	ng/ul	0.00
29) 4-Chloroaniline-d4	11.08	131	116565	43.36	ng/ul	0.00
44) Dimethylphthalate-d6	14.15	166	294525	36.75	ng/ul	0.00
47) Acenaphthylene-d8	14.45	160	358018	36.55	ng/ul	0.00
52) 4-Nitrophenol-d4	15.02	143	41984	30.83	ng/ul	0.01
58) Fluorene-d10	15.74	176	259048	36.24	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.88	200	50227	31.59	ng/ul	0.00
71) Anthracene-d10	17.59	188	414417	37.27	ng/ul	0.00
79) Pyrene-d10	19.87	212	478654	36.48	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.85	264	490728	39.42	ng/ul	0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG022221\
Data File : BG048273.D
Acq On : 22 Feb 2021 18:03
Operator : CG/JU
Sample : PB134720BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK20

Quant Time: Feb 23 00:01:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG021321MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 22 11:08:29 2021
Response via : Initial Calibration

