




Data Path : Z:\HPCHEM1\BNA G\DATA\BG022318\
Data File : BG032808.D
Acq On : 24 Feb 2018 3:05
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 26 00:16:48 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 17:04:54 2018
Response via : Initial Calibration



TIC: BG032808.D

(69) Pentachlorophenol (C)




17.904min (+0.004) 25124.35ng

response 72695

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	61.38
264.00	62.00	58.64
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022318\
Data File : BG032808.D
Acq On : 24 Feb 2018 3:05
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 26 00:16:48 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 23 17:04:54 2018
Response via : Initial Calibration



TIC: BG032808.D

(76) Benzidine

20.418min (+0.004) 129211.12ng

response 1277271

Ion	Exp%	Act%
184.00	100	100
185.00	13.90	16.14
183.00	13.30	14.01
0.00	0.00	0.00