



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022322\
 Data File : BG052456.D
 Acq On : 23 Feb 2022 8:55
 Operator : CG/JU
 Sample : DFTPP466
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP466

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Title : SVOA CALIBRATION
 Last Update : Tue Feb 22 17:32:12 2022



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	22642	PASS
68	69	0.00	2	0.9	215	PASS
69	198	0.00	100	36.2	25007	PASS
70	69	0.00	2	0.2	54	PASS
127	198	10	80	43.3	29949	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	69131	PASS
199	198	5	9	6.6	4580	PASS
275	198	10	60	31.1	21493	PASS
365	198	1	100	4.0	2755	PASS
441	198	0.01	100	14.1	9767	PASS
442	198	50	100	89.8	62067	PASS
443	442	15	24	19.4	12045	PASS

Scan 2471 (18.003 min): BG052456.D\data.ms
DFTPP466

m/z	Abundance
40.05	277.0
44.05	436.0
73.15	155.0
206.95	340.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

m/z	Abundance
38.05	344.0
39.15	2063.0
40.05	385.0
44.05	484.0
50.15	7489.0
51.15	28752.0
52.15	1548.0
56.15	915.0
57.05	2120.0
60.95	487.0
62.15	332.0
63.15	1238.0
64.15	203.0
65.15	657.0
69.05	31544.0
73.25	369.0
74.15	3242.0
75.05	5566.0
76.15	2190.0
77.15	35304.0
78.15	2669.0
79.15	2462.0
80.15	1629.0
81.05	2742.0
82.05	836.0
83.15	485.0
85.05	584.0
86.15	634.0
87.05	383.0
88.05	214.0
91.05	622.0
92.05	700.0
93.15	4040.0
94.05	391.0
96.15	289.0
96.95	153.0
98.15	3155.0
99.05	2393.0
100.05	193.0
101.05	1252.0
102.25	153.0
103.05	364.0
104.05	861.0
105.05	1040.0
106.15	473.0
107.05	11248.0
108.15	1529.0
108.95	256.0
110.05	18752.0
111.15	2770.0
112.15	218.0
116.15	953.0
117.05	8752.0
118.15	507.0
122.05	595.0
123.05	993.0
124.15	349.0
125.05	455.0
127.05	34432.0
128.05	3240.0
129.05	13913.0
130.05	1375.0
130.95	249.0
134.15	606.0
135.15	1405.0
136.05	507.0
136.95	570.0
140.05	250.0
141.05	1734.0
142.05	565.0
142.95	354.0
143.25	377.0
145.05	251.0
146.05	388.0
147.15	1027.0
148.05	2131.0
149.05	429.0
150.95	225.0
153.05	653.0
154.05	382.0
155.05	1155.0
156.05	1226.0
157.25	220.0
158.15	282.0
158.95	443.0
160.15	593.0
161.15	923.0
162.15	368.0
163.95	208.0
165.05	722.0
166.05	602.0
167.05	4033.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

168.15	1773.0
169.05	344.0
172.05	506.0
173.05	669.0
174.05	748.0
175.05	1591.0
176.25	432.0
177.05	662.0
177.95	247.0
179.05	3550.0
180.05	1771.0
181.05	968.0
181.95	154.0
184.15	284.0
185.05	1289.0
186.05	10847.0
187.15	3027.0
188.15	558.0
189.05	731.0
191.05	442.0
191.95	927.0
193.15	1313.0
196.05	2024.0
198.05	76400.0
199.05	5141.0
200.05	332.0
201.55	384.0
202.95	747.0
204.05	3018.0
205.05	4811.0
206.05	22112.0
207.05	3038.0
208.05	800.0
209.05	230.0
210.15	273.0
211.15	707.0
214.95	333.0
216.05	604.0
217.05	6401.0
218.05	896.0
221.05	3954.0
221.95	762.0
222.15	757.0
222.95	1155.0
224.05	11532.0
225.05	3649.0
225.95	319.0
227.05	5712.0
227.95	556.0
229.05	728.0
230.05	158.0
231.05	525.0
233.05	268.0
234.05	455.0
234.95	586.0
236.05	353.0
237.15	447.0
239.15	272.0
240.15	259.0
240.95	300.0
242.05	644.0
243.15	811.0
244.05	9058.0
245.05	1685.0
246.05	1752.0
247.05	559.0
248.05	164.0
249.05	507.0
252.95	245.0
255.05	50896.0
256.05	7689.0
257.05	498.0
258.05	3117.0
259.05	341.0
265.05	1057.0
265.85	180.0
273.05	1568.0
274.05	3351.0
275.05	22936.0
276.05	2854.0
277.05	1872.0
277.95	281.0
282.95	229.0
285.05	261.0
285.95	152.0
293.05	650.0
294.05	168.0
294.95	221.0
296.05	6910.0
297.05	891.0
303.05	777.0
304.05	297.0
314.15	339.0
314.95	576.0
315.95	355.0
321.85	150.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

323.05	1942.0
324.05	369.0
327.05	321.0
328.05	287.0
332.05	229.0
334.05	1326.0
335.05	268.0
341.15	329.0
346.15	328.0
352.05	518.0
353.05	300.0
353.95	517.0
364.95	2813.0
366.05	350.0
371.95	872.0
372.95	293.0
382.95	260.0
402.05	320.0
402.95	699.0
404.05	216.0
421.05	482.0
421.95	390.0
423.05	3011.0
424.05	790.0
441.05	8297.0
442.05	55120.0
443.05	10905.0
444.05	1444.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

m/z	Abundance
38.15	319.0
39.15	2342.0
39.95	262.0
44.15	540.0
49.15	280.0
50.15	7654.0
51.15	25480.0
52.15	1349.0
53.15	195.0
55.05	210.0
56.05	1000.0
57.05	1704.0
61.05	314.0
62.05	418.0
63.05	1125.0
65.15	671.0
68.15	646.0
69.05	28848.0
69.85	162.0
73.05	470.0
74.15	2616.0
75.05	4561.0
76.05	1509.0
77.15	32400.0
78.15	2360.0
79.05	2378.0
80.15	1899.0
81.05	2528.0
82.05	565.0
83.05	415.0
85.05	465.0
86.15	703.0
87.15	374.0
88.25	210.0
91.05	655.0
92.05	562.0
93.05	3490.0
94.05	162.0
98.15	3228.0
99.15	2503.0
100.15	220.0
101.05	1506.0
103.05	416.0
104.05	1044.0
105.05	891.0
106.05	386.0
107.15	10110.0
108.05	1881.0
110.15	15873.0
111.05	2530.0
111.95	428.0
112.85	165.0
113.05	160.0
116.05	859.0
117.05	8141.0
118.05	582.0
120.15	241.0
121.95	734.0
123.15	904.0
124.15	547.0
125.05	582.0
127.05	34872.0
128.05	3001.0
129.05	14666.0
130.15	1498.0
130.95	348.0
132.15	155.0
134.05	485.0
135.05	1478.0
136.05	387.0
137.15	647.0
137.95	170.0
141.05	2330.0
142.05	618.0
143.15	360.0
145.15	160.0
146.15	301.0
147.15	1053.0
148.15	2069.0
149.05	661.0
151.25	288.0
153.05	779.0
154.15	435.0
155.15	1137.0
156.05	1647.0
157.15	305.0
157.95	556.0
158.95	399.0
160.05	974.0
161.15	943.0
161.85	205.0
162.05	209.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

165.05	1063.0
166.05	662.0
167.05	3877.0
168.05	1616.0
169.05	267.0
170.25	262.0
172.05	434.0
173.05	533.0
174.05	684.0
175.15	1461.0
176.15	542.0
177.05	837.0
178.05	207.0
179.05	3037.0
180.15	2228.0
181.05	967.0
184.05	280.0
185.15	1587.0
186.05	12034.0
187.05	3222.0
188.05	392.0
189.05	736.0
190.05	156.0
191.05	209.0
192.05	927.0
193.15	1060.0
193.95	242.0
195.05	203.0
196.15	2116.0
198.05	82312.0
199.05	5478.0
200.05	412.0
201.65	342.0
202.95	696.0
204.05	2922.0
205.05	5465.0
206.15	21360.0
207.15	3074.0
208.15	911.0
209.05	267.0
210.05	324.0
211.15	854.0
215.05	314.0
216.15	668.0
217.05	6370.0
218.15	741.0
219.05	175.0
221.15	3026.0
222.05	849.0
222.95	1237.0
224.05	12756.0
225.05	3298.0
226.05	456.0
227.05	5468.0
228.05	654.0
229.05	1171.0
230.05	339.0
231.05	584.0
234.05	398.0
235.05	437.0
236.15	342.0
237.05	481.0
241.05	369.0
242.05	942.0
243.15	912.0
244.05	10369.0
245.05	1469.0
246.05	1981.0
247.15	422.0
249.05	430.0
253.05	255.0
255.05	57176.0
256.05	7324.0
257.15	541.0
258.15	2542.0
259.05	591.0
259.95	200.0
264.05	285.0
265.05	1014.0
266.05	210.0
271.05	157.0
272.05	230.0
273.05	1887.0
274.05	4371.0
275.05	25288.0
276.05	3263.0
277.05	1876.0
278.15	385.0
278.95	181.0
281.15	163.0
282.95	226.0
284.25	184.0
285.05	474.0
289.15	201.0
293.05	562.0
294.85	159.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

296.05	7262.0
297.05	932.0
302.15	172.0
303.05	1189.0
304.05	235.0
314.05	407.0
314.95	724.0
315.95	524.0
321.05	227.0
322.05	167.0
323.05	2853.0
324.05	382.0
327.05	501.0
328.15	169.0
333.05	418.0
333.95	1458.0
334.95	309.0
341.05	234.0
342.25	151.0
346.05	297.0
352.05	941.0
353.05	495.0
354.05	873.0
364.95	3054.0
366.05	390.0
370.95	232.0
372.05	1289.0
372.85	233.0
382.95	416.0
390.05	303.0
402.05	405.0
403.05	521.0
404.05	272.0
421.05	736.0
422.15	676.0
423.05	4386.0
424.15	999.0
441.05	12115.0
442.05	75360.0
443.05	14484.0
444.05	1214.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

m/z	Abundance
39.15	1529.0
39.95	160.0
44.15	731.0
50.15	3121.0
51.15	13693.0
52.15	681.0
56.15	371.0
57.05	774.0
60.95	156.0
62.15	228.0
63.15	551.0
65.05	319.0
69.05	14629.0
74.15	1810.0
75.15	2222.0
76.15	757.0
77.15	18304.0
78.15	1172.0
79.15	1133.0
80.15	1128.0
81.15	1116.0
82.05	443.0
83.25	378.0
85.05	271.0
86.05	291.0
87.05	246.0
90.95	290.0
92.05	520.0
93.15	2004.0
98.05	1218.0
99.05	1887.0
101.15	650.0
103.15	253.0
104.15	564.0
105.15	644.0
107.05	6294.0
108.05	893.0
110.05	10694.0
111.05	1665.0
116.05	448.0
117.15	4487.0
118.05	491.0
121.95	371.0
123.05	535.0
123.95	353.0
125.15	441.0
127.15	20544.0
128.15	1427.0
129.05	7629.0
130.15	760.0
131.15	173.0
134.05	407.0
135.25	821.0
135.95	364.0
137.05	309.0
141.05	1287.0
142.15	316.0
146.05	218.0
147.05	541.0
148.05	1278.0
148.95	633.0
151.15	341.0
153.05	593.0
154.15	247.0
155.05	589.0
156.15	1052.0
158.15	197.0
159.05	250.0
160.05	657.0
161.05	548.0
162.05	198.0
164.05	156.0
165.05	585.0
166.05	523.0
167.15	2202.0
168.05	1000.0
171.15	170.0
172.05	217.0
172.85	333.0
173.15	341.0
174.15	618.0
175.15	1097.0
175.95	151.0
176.95	360.0
177.75	156.0
179.05	2184.0
180.05	1144.0
181.15	754.0
183.95	258.0
185.05	758.0
186.05	6997.0
187.05	1644.0

Instrument :

BNA_G

ClientSampleId :

DFTPP466

188.25	226.0
189.15	525.0
191.15	318.0
192.15	515.0
193.05	545.0
193.85	298.0
196.05	1927.0
198.05	48680.0
198.95	3122.0
200.05	169.0
201.55	206.0
203.05	526.0
204.05	1826.0
205.15	3163.0
206.05	13591.0
207.15	2248.0
208.15	516.0
210.25	335.0
210.95	501.0
215.05	227.0
216.15	500.0
217.05	4626.0
218.05	485.0
221.05	2365.0
221.75	689.0
223.05	851.0
224.05	7962.0
225.15	1541.0
227.05	3796.0
228.05	520.0
229.05	629.0
229.95	172.0
231.05	342.0
234.05	153.0
235.05	268.0
236.15	262.0
237.05	361.0
238.95	152.0
240.15	241.0
242.05	604.0
243.15	775.0
244.15	6644.0
245.05	748.0
246.05	1383.0
246.95	233.0
249.15	246.0
252.05	187.0
253.05	230.0
255.05	35992.0
256.05	5239.0
257.05	451.0
258.05	1863.0
258.95	611.0
265.05	985.0
273.05	967.0
274.05	2643.0
275.05	16256.0
276.05	2499.0
277.05	1654.0
278.05	244.0
283.05	276.0
283.95	223.0
293.05	471.0
294.05	187.0
296.05	4503.0
297.05	632.0
303.05	545.0
304.05	178.0
314.05	219.0
315.05	604.0
315.85	195.0
321.05	206.0
323.05	1736.0
324.05	232.0
326.95	387.0
328.15	261.0
332.95	225.0
334.15	950.0
345.95	479.0
352.15	616.0
353.05	431.0
354.05	566.0
364.95	2397.0
366.15	306.0
372.05	1004.0
373.05	187.0
383.05	282.0
401.95	333.0
403.05	524.0
404.05	165.0
420.95	418.0
422.05	364.0
423.05	3667.0
424.15	638.0
441.05	8888.0
442.05	55720.0

Instrument :
BNA_G
ClientSampleId :
DFTPP466

443.0510747.0

444.05879.0

Instrument :

BNA_G

ClientSampleId :

DFTPP466