

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022416\
 Data File : BG021064.D
 Acq On : 24 Feb 2016 23:11
 Operator : UM/SJ
 Sample : H1521-09
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP-3(0-2)

Manual Integrations
 APPROVED

Sohil
 2/25/2016 2:40:26 PM

Quant Time: Feb 25 03:41:55 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 24 00:35:06 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	5855	20.00	ng	0.00
21) Naphthalene-d8	11.03	136	27658	20.00	ng	0.00
38) Acenaphthene-d10	14.82	164	12947	20.00	ng	0.00
63) Phenanthrene-d10	17.56	188	22441	20.00	ng	0.00
75) Chrysene-d12	21.83	240	16560	20.00	ng	0.00
86) Perylene-d12	25.16	264	13816	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.78	112	45330	135.73	ng	0.00
7) Phenol-d6	7.41	99	69552	135.23	ng	0.00
23) Nitrobenzene-d5	9.39	82	46134	97.93	ng	0.00
41) 2,4,6-Tribromophenol	16.31	330	19237	124.03	ng	0.00
44) 2-Fluorobiphenyl	13.45	172	75081	80.73	ng	0.00
78) Terphenyl-d14	20.14	244	53613	42.01	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	14.27	163	15355	14.54	ng	98
70) Phenanthrene	17.60	178	8887	7.00	ng	96
74) Fluoranthene	19.59	202	10650	8.34	ng	97
77) Pyrene	19.96	202	8835	4.67	ng	# 96
79) Butylbenzylphthalate	20.83	149	278093	416.78	ng	98
80) Benzo(a)anthracene	21.81	228	3749	3.78	ng	# 93
82) Chrysene	21.87	228	3824	4.13	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022416\
Data File : BG021064.D
Acq On : 24 Feb 2016 23:11
Operator : UM/SJ
Sample : H1521-09
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SP-3(0-2)

Manual Integrations
APPROVED

Sohil
2/25/2016 2:40:26 PM

Quant Time: Feb 25 03:41:55 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 24 00:35:06 2016
Response via : Initial Calibration

