




Data File : BG025950.D
Acq On : 25 Feb 2017 2:59
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 25 04:13:07 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG021717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 15:48:21 2017
Response via : Initial Calibration



TIC: BG025950.D

(69) Pentachlorophenol (C)




17.066min (-0.000) 0.00ng

response 383708

Ion	Exp%	Act%
265.70	100	100
268.00	60.70	63.88
264.00	62.60	62.92
0.00	0.00	0.00

Data File : BG025950.D
Acq On : 25 Feb 2017 2:59
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 25 04:13:07 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG021717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 15:48:21 2017
Response via : Initial Calibration



TIC: BG025950.D

(76) Benzidine

19.627min (-0.006) 48016.56ng

response 1779497

Ion	Exp%	Act%
184.00	100	100
185.00	16.30	15.56
183.00	13.80	14.07
0.00	0.00	0.00