



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG022522\
 Data File : BG052484.D
 Acq On : 25 Feb 2022 9:18
 Operator : CG/JU
 Sample : DFTPP468
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP468

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022222.M
 Title : SVOA CALIBRATION
 Last Update : Tue Feb 22 17:32:12 2022



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	36029	PASS
68	69	0.00	2	1.7	647	PASS
69	198	0.00	100	37.5	38776	PASS
70	69	0.00	2	0.5	185	PASS
127	198	10	80	45.0	46611	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	103467	PASS
199	198	5	9	6.8	7086	PASS
275	198	10	60	30.9	31987	PASS
365	198	1	100	3.3	3435	PASS
441	198	0.01	100	11.8	12166	PASS
442	198	50	100	76.6	79267	PASS
443	442	15	24	19.2	15206	PASS

Scan 2471 (18.004 min): BG052484.D\data.ms
DFTPP468

m/z	Abundance
40.05	230.0
42.95	152.0
44.05	682.0
207.15	335.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

m/z	Abundance
38.15	549.0
39.15	2594.0
40.05	385.0
44.05	709.0
49.15	252.0
50.15	8221.0
51.15	30504.0
52.15	1674.0
56.05	1128.0
57.15	2446.0
61.15	784.0
62.15	482.0
63.15	1357.0
65.15	678.0
69.05	31736.0
70.95	161.0
73.15	159.0
74.15	3633.0
75.15	5065.0
76.25	1696.0
77.15	36696.0
78.15	2070.0
79.15	3032.0
80.05	2099.0
81.05	3084.0
82.15	746.0
83.15	640.0
85.05	703.0
86.15	745.0
87.05	358.0
88.05	166.0
90.95	422.0
92.05	540.0
93.15	4243.0
94.15	353.0
96.05	296.0
98.05	3602.0
99.05	2558.0
99.95	351.0
101.05	1596.0
103.05	358.0
104.05	768.0
105.05	624.0
107.15	10819.0
108.05	1820.0
110.05	18320.0
111.05	3441.0
112.05	289.0
113.15	173.0
115.05	153.0
116.15	536.0
117.05	8261.0
118.05	577.0
119.05	179.0
122.05	469.0
123.05	1050.0
124.05	428.0
125.05	507.0
127.05	35968.0
128.05	2373.0
129.05	14530.0
130.15	1080.0
131.05	317.0
134.05	222.0
135.05	1266.0
136.15	396.0
137.15	588.0
140.15	236.0
141.05	1959.0
142.05	829.0
143.05	408.0
145.95	274.0
147.05	1012.0
148.15	2194.0
149.05	450.0
151.15	360.0
153.05	982.0
154.15	323.0
155.15	1406.0
156.05	1447.0
157.25	208.0
158.05	268.0
159.05	354.0
159.95	766.0
161.05	695.0
161.95	248.0
165.05	639.0
166.15	646.0
167.15	3574.0
168.05	1705.0
169.15	275.0
170.05	166.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

170.85	152.0
172.05	341.0
173.15	287.0
174.05	889.0
175.05	1019.0
176.05	414.0
177.15	728.0
179.05	3226.0
180.05	1991.0
181.15	1138.0
184.05	258.0
185.15	1797.0
186.05	10438.0
187.15	3070.0
188.15	284.0
189.15	820.0
190.15	177.0
191.15	478.0
192.15	992.0
192.95	815.0
194.05	163.0
196.15	1976.0
198.05	75656.0
199.05	5229.0
199.95	320.0
201.35	371.0
203.05	686.0
204.15	2932.0
205.05	4861.0
206.05	20576.0
207.05	2966.0
208.05	961.0
209.15	328.0
210.05	392.0
211.05	856.0
216.15	321.0
217.05	5514.0
217.95	399.0
221.05	2575.0
222.05	784.0
223.05	1436.0
224.05	10904.0
225.05	3062.0
226.15	344.0
227.05	4037.0
227.95	635.0
229.15	945.0
230.95	450.0
234.05	421.0
234.95	343.0
236.05	247.0
237.05	413.0
238.85	227.0
240.15	293.0
240.95	334.0
242.15	623.0
243.15	566.0
244.05	8314.0
245.05	1204.0
246.05	2158.0
246.95	331.0
248.95	424.0
253.05	216.0
253.85	324.0
255.05	44864.0
256.05	7638.0
256.95	507.0
258.05	2484.0
259.05	254.0
260.95	185.0
265.05	1001.0
271.05	191.0
272.05	220.0
273.05	1204.0
274.05	3357.0
275.05	20968.0
276.05	2610.0
277.05	1616.0
283.05	259.0
283.95	201.0
284.95	397.0
293.05	267.0
294.15	222.0
295.15	193.0
296.05	5350.0
296.95	751.0
298.05	184.0
302.05	167.0
303.15	741.0
303.95	169.0
314.15	431.0
315.05	804.0
315.95	315.0
320.95	181.0
323.15	1393.0
323.95	324.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

326.85	394.0
328.15	217.0
332.95	161.0
334.05	1514.0
335.05	169.0
340.85	259.0
346.05	312.0
352.05	465.0
353.15	371.0
354.05	561.0
364.95	1886.0
365.95	260.0
371.25	202.0
372.05	1109.0
373.05	297.0
382.85	150.0
401.85	218.0
402.95	575.0
403.85	164.0
420.95	584.0
422.05	360.0
423.05	2567.0
424.05	571.0
441.05	6512.0
442.05	44192.0
443.05	8802.0
444.15	739.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

m/z	Abundance
37.05	337.0
38.15	556.0
39.15	2655.0
40.05	461.0
44.05	470.0
49.15	337.0
50.15	10574.0
51.15	43080.0
52.15	2175.0
55.15	228.0
56.15	1066.0
57.05	3393.0
61.15	608.0
62.15	653.0
63.15	1979.0
64.15	363.0
65.15	1097.0
68.15	1102.0
69.05	45712.0
70.05	345.0
73.05	480.0
74.15	4375.0
75.15	7632.0
76.15	2943.0
77.15	52056.0
78.15	3238.0
79.05	3689.0
80.05	2991.0
81.15	4006.0
82.05	729.0
83.05	902.0
84.15	190.0
85.15	608.0
86.15	958.0
87.15	312.0
88.15	223.0
91.05	1302.0
92.05	1099.0
93.15	5846.0
94.15	407.0
98.05	5031.0
99.05	3727.0
100.15	430.0
101.05	2018.0
102.05	237.0
103.05	686.0
104.05	1225.0
105.05	1490.0
106.15	441.0
107.15	16528.0
108.15	2822.0
108.95	412.0
110.05	28296.0
111.05	4462.0
111.95	818.0
112.95	195.0
116.05	952.0
117.05	13932.0
118.15	943.0
120.15	216.0
121.15	187.0
122.15	1380.0
123.15	1864.0
124.15	736.0
125.05	750.0
127.15	55320.0
128.05	3514.0
129.05	23168.0
130.05	1990.0
131.15	258.0
132.15	234.0
134.05	532.0
135.05	1745.0
136.05	920.0
137.15	922.0
138.05	379.0
139.85	285.0
141.05	3343.0
142.15	1154.0
143.15	676.0
145.15	156.0
146.05	770.0
147.15	1666.0
148.15	3742.0
149.05	804.0
150.05	222.0
150.95	369.0
151.45	260.0
152.25	225.0
153.05	922.0
154.05	482.0
155.05	1806.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

156.05	2416.0
157.05	518.0
158.05	589.0
159.05	478.0
160.05	1089.0
161.05	1736.0
162.05	415.0
165.05	1312.0
166.25	840.0
167.05	4997.0
168.05	2508.0
169.05	528.0
170.15	265.0
171.05	488.0
172.05	759.0
173.05	665.0
174.05	1196.0
175.15	2498.0
176.15	605.0
177.05	1138.0
178.05	577.0
179.05	4329.0
180.05	3237.0
181.05	1449.0
182.15	190.0
184.15	355.0
185.15	2195.0
186.15	18544.0
187.15	4809.0
188.05	532.0
188.95	897.0
190.05	202.0
191.15	419.0
192.05	1857.0
193.05	1484.0
194.05	352.0
196.05	3152.0
198.05	122320.0
199.05	8399.0
199.95	728.0
201.45	892.0
202.25	165.0
203.15	1046.0
204.15	4846.0
205.05	7713.0
206.15	33216.0
207.15	4667.0
208.05	917.0
209.05	453.0
210.05	483.0
211.15	1221.0
212.15	174.0
213.05	187.0
215.05	233.0
216.05	1100.0
217.05	9349.0
218.05	944.0
219.05	192.0
221.05	4966.0
221.85	1224.0
223.05	2105.0
224.05	17936.0
225.15	5099.0
225.95	717.0
227.05	8209.0
228.05	1115.0
229.05	1719.0
229.85	227.0
231.05	846.0
234.05	427.0
235.15	546.0
236.05	430.0
236.95	579.0
239.15	463.0
240.05	208.0
240.95	275.0
242.05	1425.0
243.25	864.0
244.05	14683.0
245.05	2043.0
246.05	3026.0
246.95	720.0
248.25	194.0
248.95	679.0
251.15	155.0
252.05	180.0
253.25	296.0
255.05	77968.0
256.05	11267.0
257.05	1200.0
258.05	4626.0
259.05	869.0
265.05	1806.0
266.15	440.0
271.05	191.0
273.05	2416.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

274.05	6180.0
275.05	37152.0
276.05	4806.0
277.05	2923.0
278.15	460.0
281.05	170.0
282.15	154.0
282.95	491.0
283.85	198.0
284.95	330.0
289.05	153.0
290.05	216.0
293.05	564.0
293.85	292.0
294.85	280.0
296.05	10777.0
297.15	1064.0
297.95	157.0
301.95	256.0
303.05	1417.0
304.05	423.0
308.05	250.0
313.25	224.0
313.95	334.0
314.95	1179.0
316.05	948.0
321.05	348.0
322.25	179.0
323.05	3219.0
324.05	553.0
327.05	610.0
327.95	246.0
331.95	187.0
332.95	226.0
334.05	1855.0
335.05	516.0
341.15	364.0
341.95	207.0
345.95	511.0
352.05	1115.0
352.95	499.0
354.05	965.0
355.15	198.0
365.05	4037.0
366.05	624.0
370.05	170.0
370.85	159.0
372.05	1594.0
373.05	442.0
382.95	363.0
383.95	175.0
389.95	208.0
391.05	189.0
391.95	158.0
402.05	806.0
403.05	1096.0
404.05	384.0
421.05	735.0
422.05	679.0
423.05	5589.0
424.05	1314.0
425.05	209.0
441.05	13979.0
442.05	88864.0
443.05	16912.0
444.05	1695.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

m/z	Abundance
38.15	459.0
39.15	2920.0
40.05	396.0
44.05	754.0
45.15	167.0
50.15	9570.0
51.15	34504.0
52.15	2034.0
55.05	217.0
56.05	1266.0
57.15	2431.0
57.95	218.0
61.15	654.0
62.15	596.0
63.15	1461.0
64.05	230.0
65.15	921.0
68.15	840.0
69.05	38880.0
70.15	210.0
74.15	3933.0
75.15	6141.0
76.25	2097.0
77.15	44832.0
78.15	2659.0
79.05	2779.0
80.15	2588.0
81.05	3660.0
82.15	732.0
83.05	732.0
85.05	471.0
86.05	820.0
86.95	379.0
87.85	167.0
91.15	528.0
92.15	1028.0
93.15	5156.0
94.15	336.0
95.15	198.0
96.05	249.0
98.15	4496.0
99.05	3379.0
100.05	324.0
101.05	1814.0
103.05	647.0
104.15	1539.0
105.05	1192.0
107.05	14535.0
108.15	1834.0
110.05	24792.0
111.05	4049.0
112.05	458.0
113.05	248.0
116.05	819.0
117.15	12134.0
118.05	842.0
119.05	230.0
120.25	230.0
122.15	947.0
123.05	1656.0
124.05	883.0
125.15	698.0
127.05	48544.0
128.15	4311.0
129.05	20088.0
130.05	1998.0
131.05	393.0
131.85	325.0
133.05	156.0
134.05	566.0
135.15	1820.0
136.05	791.0
137.15	796.0
138.15	192.0
138.95	155.0
140.05	369.0
141.05	2787.0
142.05	841.0
143.05	498.0
143.95	162.0
146.05	651.0
147.05	1638.0
148.05	2834.0
149.05	723.0
151.05	470.0
153.15	745.0
154.05	575.0
155.05	1568.0
156.15	2364.0
157.05	531.0
158.05	671.0
159.15	468.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

160.05	1112.0
161.05	1156.0
162.05	326.0
165.15	1005.0
166.05	838.0
167.05	5356.0
168.15	2385.0
169.05	430.0
169.95	251.0
171.05	307.0
172.05	414.0
173.05	740.0
174.05	1158.0
175.05	2451.0
176.05	683.0
177.05	870.0
177.95	225.0
179.05	3702.0
180.05	2606.0
181.05	1547.0
182.15	329.0
183.05	220.0
184.05	501.0
185.05	2151.0
186.05	16366.0
187.05	4647.0
187.95	470.0
189.05	1025.0
191.05	590.0
192.05	1383.0
193.05	1798.0
193.95	329.0
195.25	227.0
196.15	3418.0
198.05	112424.0
199.05	7630.0
200.05	607.0
201.55	592.0
203.05	828.0
204.15	3934.0
205.15	7428.0
206.15	32512.0
207.05	3746.0
208.15	1409.0
209.15	466.0
210.05	329.0
211.05	1272.0
211.65	476.0
215.05	350.0
215.95	623.0
217.05	9070.0
218.05	1068.0
221.05	3952.0
223.05	1953.0
224.15	19360.0
225.05	5495.0
226.05	697.0
227.05	7629.0
228.15	973.0
229.05	1721.0
230.05	274.0
231.05	770.0
232.25	195.0
234.05	414.0
234.95	758.0
236.05	353.0
237.05	675.0
239.15	359.0
240.05	208.0
240.95	574.0
242.05	1093.0
243.05	1012.0
244.05	13840.0
245.05	1541.0
246.15	2801.0
246.95	618.0
249.05	853.0
249.95	221.0
250.95	237.0
253.15	306.0
255.05	79888.0
256.05	11606.0
256.95	702.0
258.05	4408.0
259.05	620.0
260.95	172.0
265.05	1628.0
266.15	254.0
270.95	330.0
272.05	304.0
273.15	2218.0
274.05	6375.0
275.05	37840.0
276.05	4502.0
277.05	3138.0
278.05	599.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468

282.95	453.0
284.05	224.0
285.15	569.0
285.85	152.0
289.15	206.0
289.95	170.0
292.05	206.0
293.05	636.0
294.85	299.0
296.05	9723.0
296.95	1659.0
301.95	224.0
303.15	1421.0
304.05	492.0
308.05	180.0
313.95	465.0
315.05	1099.0
315.95	629.0
317.05	182.0
321.15	273.0
321.95	151.0
323.05	3252.0
324.05	474.0
326.95	666.0
328.05	291.0
332.05	289.0
333.05	483.0
334.05	2110.0
335.15	436.0
341.05	448.0
342.05	258.0
346.05	783.0
352.05	1029.0
353.05	874.0
354.05	1156.0
355.05	324.0
365.05	4382.0
366.05	470.0
371.15	168.0
372.05	1975.0
373.05	414.0
382.95	466.0
391.15	236.0
401.95	817.0
403.05	1082.0
404.15	213.0
421.05	943.0
422.05	745.0
423.05	6567.0
424.05	1256.0
441.05	16007.0
442.05	104744.0
443.05	19904.0
444.05	1972.0

Instrument :
BNA_G
ClientSampleId :
DFTPP468