

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG022615\
 Data File : BG016047.D
 Acq On : 26 Feb 2015 15:34
 Operator : TP/IZ
 Sample : PB81907BL
 Misc : SBLK27
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK27

Quant Time: Feb 27 00:16:22 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-MA-BG021915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 27 00:07:23 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	69025	20.00	ng/ul	0.00
16) Naphthalene-d8	10.59	136	308257	20.00	ng/ul	0.00
33) Acenaphthene-d10	14.43	164	187640	20.00	ng/ul	0.00
59) Phenanthrene-d10	17.17	188	383270	20.00	ng/ul	0.00
75) Chrysene-d12	21.34	240	366788	20.00	ng/ul	0.00
83) Perylene-d12	23.65	264	367247	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) Phenol-d5	6.96	99	194974	29.69	ng/ul	0.00
5) Bis-(2-Chloroethyl)ether-d	7.13	67	112954	29.12	ng/ul	0.00
7) 2-Chlorophenol-d4	7.33	132	147817	30.43	ng/ul	0.00
11) 4-Methylphenol-d8	8.50	113	150699	31.01	ng/ul	0.00
17) Nitrobenzene-d5	8.96	128	78797	36.17	ng/ul	0.00
20) 2-Nitrophenol-d4	9.67	143	71409	35.53	ng/ul	0.00
24) 2,4-Dichlorophenol-d3	10.21	165	131250	30.11	ng/ul	0.00
27) 4-Chloroaniline-d4	10.72	131	236370	49.22	ng/ul	0.00
41) Dimethylphthalate-d6	13.84	166	450722	35.93	ng/ul	0.00
44) Acenaphthylene-d8	14.12	160	569618	32.08	ng/ul	0.00
49) 4-Nitrophenol-d4	14.62	143	65041	25.78	ng/ul	0.00
55) Fluorene-d10	15.42	176	408884	32.43	ng/ul	0.00
60) 4,6-Dinitro-2-methylphenol	15.53	200	27321	18.49	ng/ul	0.00
68) Anthracene-d10	17.27	188	628893	34.60	ng/ul	0.00
76) Pyrene-d10	19.56	212	633178	36.75	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.50	264	598446	35.94	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG022615\
Data File : BG016047.D
Acq On : 26 Feb 2015 15:34
Operator : TP/IZ
Sample : PB81907BL
Misc : SBLK27
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK27

Quant Time: Feb 27 00:16:22 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-MA-BG021915.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 27 00:07:23 2015
Response via : Initial Calibration

