

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022616\
 Data File : BG021091.D
 Acq On : 26 Feb 2016 23:26
 Operator : UM/SJ
 Sample : H1558-04
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 S6

Quant Time: Feb 27 01:19:02 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 26 16:39:44 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	5577	20.00	ng	0.00
21) Naphthalene-d8	10.98	136	23885	20.00	ng	-0.01
38) Acenaphthene-d10	14.78	164	15457	20.00	ng	0.00
63) Phenanthrene-d10	17.51	188	41119	20.00	ng	0.00
75) Chrysene-d12	21.77	240	50443	20.00	ng	-0.02
86) Perylene-d12	25.06	264	45550	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	46503	145.98	ng	0.00
7) Phenol-d6	7.31	99	68599	144.72	ng	-0.01
23) Nitrobenzene-d5	9.34	82	46100	91.96	ng	-0.01
41) 2,4,6-Tribromophenol	16.26	330	31879	157.31	ng	0.00
44) 2-Fluorobiphenyl	13.41	172	107536	89.03	ng	0.00
78) Terphenyl-d14	20.11	244	223122	103.37	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.22	163	19017	14.28	ng	Qvalue # 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022616\
Data File : BG021091.D
Acq On : 26 Feb 2016 23:26
Operator : UM/SJ
Sample : H1558-04
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
S6

Quant Time: Feb 27 01:19:02 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 26 16:39:44 2016
Response via : Initial Calibration

