


Data Path : Z:\HPCHEM1\BNA G\Data\BG022618\  
 Data File : BG032814.D  
 Acq On : 26 Feb 2018 9:02  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG022118.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Feb 26 17:50:45 2018



AutoFind: Scans 2613, 2614, 2615; Background Corrected with Scan 2606

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 40.9      | 28029   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 429     | PASS             |
| 69          | 198          | 0.00         | 100          | 35.7      | 24464   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 144     | PASS             |
| 127         | 198          | 10           | 80           | 51.1      | 35000   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 68469   | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 4769    | PASS             |
| 275         | 198          | 10           | 60           | 26.7      | 18265   | PASS             |
| 365         | 198          | 1            | 100          | 3.4       | 2352    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.7      | 8110    | PASS             |
| 442         | 198          | 50           | 100          | 75.4      | 51658   | PASS             |
| 443         | 442          | 15           | 24           | 18.9      | 9775    | PASS             |