


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG022619\
 Data File : BG039596.D
 Acq On : 26 Feb 2019 12:05
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG012919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 29 15:41:51 2019



AutoFind: Scans 2525, 2526, 2527; Background Corrected with Scan 2519

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.2	25967	PASS
68	69	0.00	2	1.5	326	PASS
69	198	0.00	100	29.1	21462	PASS
70	69	0.00	2	0.7	154	PASS
127	198	10	80	41.6	30677	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	73696	PASS
199	198	5	9	6.3	4677	PASS
275	198	10	60	26.3	19365	PASS
365	198	1	100	3.9	2854	PASS
441	198	0.01	100	10.6	7816	PASS
442	198	50	100	63.3	46672	PASS
443	442	15	24	19.1	8929	PASS