

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022716\
 Data File : BG021101.D
 Acq On : 27 Feb 2016 5:58
 Operator : UM/SJ
 Sample : H1558-09
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 S11

Quant Time: Feb 29 00:24:32 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 26 16:39:44 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	5697	20.00	ng	0.00
21) Naphthalene-d8	10.98	136	24064	20.00	ng	-0.01
38) Acenaphthene-d10	14.77	164	15170	20.00	ng	0.00
63) Phenanthrene-d10	17.51	188	40032	20.00	ng	0.00
75) Chrysene-d12	21.78	240	50033	20.00	ng	-0.01
86) Perylene-d12	25.06	264	45332	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.73	112	41353	127.08	ng	0.00
7) Phenol-d6	7.31	99	60170	124.26	ng	-0.01
23) Nitrobenzene-d5	9.33	82	41079	81.34	ng	-0.01
41) 2,4,6-Tribromophenol	16.25	330	24694	124.16	ng	0.00
44) 2-Fluorobiphenyl	13.40	172	91837	77.47	ng	0.00
78) Terphenyl-d14	20.10	244	183008	85.48	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	14.22	163	15431	11.80	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022716\
Data File : BG021101.D
Acq On : 27 Feb 2016 5:58
Operator : UM/SJ
Sample : H1558-09
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
S11

Quant Time: Feb 29 00:24:32 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 26 16:39:44 2016
Response via : Initial Calibration

