

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022716\
 Data File : BG021108.D
 Acq On : 27 Feb 2016 10:32
 Operator : UM/SJ
 Sample : H1558-18
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 S-20

Quant Time: Feb 29 00:58:15 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 26 16:39:44 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	5930	20.00	ng	-0.01
21) Naphthalene-d8	10.98	136	24231	20.00	ng	-0.01
38) Acenaphthene-d10	14.78	164	15259	20.00	ng	0.00
63) Phenanthrene-d10	17.51	188	40099	20.00	ng	-0.01
75) Chrysene-d12	21.77	240	49847	20.00	ng	-0.02
86) Perylene-d12	25.06	264	45628	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	57561	169.94	ng	0.00
7) Phenol-d6	7.31	99	86724	172.06	ng	-0.01
23) Nitrobenzene-d5	9.33	82	58735	115.49	ng	-0.02
41) 2,4,6-Tribromophenol	16.26	330	33733	168.61	ng	0.00
44) 2-Fluorobiphenyl	13.40	172	121021	101.49	ng	-0.01
78) Terphenyl-d14	20.10	244	244778	114.76	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	14.22	163	24281	18.46	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022716\
Data File : BG021108.D
Acq On : 27 Feb 2016 10:32
Operator : UM/SJ
Sample : H1558-18
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
S-20

Quant Time: Feb 29 00:58:15 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 26 16:39:44 2016
Response via : Initial Calibration

