

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022717\
 Data File : BG025994.D
 Acq On : 28 Feb 2017 11:36
 Operator : SJ/MA
 Sample : I2011-01DL 5X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 FRAC-TANK-A155DL

Manual Integrations
 APPROVED

mohammad
 3/1/2017 12:38:10 PM

Quant Time: Feb 28 15:22:04 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 17 15:48:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.08	152	107622	20.00	ng	-0.01
21) Naphthalene-d8	10.88	136	488279	20.00	ng	-0.01
38) Acenaphthene-d10	14.68	164	337084	20.00	ng	-0.01
63) Phenanthrene-d10	17.41	188	755518	20.00	ng	-0.01
75) Chrysene-d12	21.65	240	780176	20.00	ng	-0.02
86) Perylene-d12	24.85	264	775256	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.64	112	78341	12.67	ng	-0.01
7) Phenol-d6	7.23	99	66354	7.25	ng	0.00
23) Nitrobenzene-d5	9.23	82	173893	22.47	ng	-0.02
41) 2,4,6-Tribromophenol	16.15	330	115044	36.94	ng	-0.01
44) 2-Fluorobiphenyl	13.31	172	490642	25.43	ng	-0.01
78) Terphenyl-d14	20.00	244	593901	18.51	ng	-0.02
Target Compounds						
31) Naphthalene	10.93	128	344461	13.64	ng	# 96
37) 2-Methylnaphthalene	12.52	142	559811	29.12	ng	# 92
45) 1,1'-Biphenyl	13.52	154	96593m	3.95	ng	
49) Dimethylphthalate	14.12	163	50694	2.19	ng	# 79
83) Bis(2-ethylhexyl)phthalate	21.54	149	84920	3.14	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG022717\
Data File : BG025994.D
Acq On : 28 Feb 2017 11:36
Operator : SJ/MA
Sample : I2011-01DL 5X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
FRAC-TANK-A155DL

Manual Integrations
APPROVED

mohammad
3/1/2017 12:38:10 PM

Quant Time: Feb 28 15:22:04 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG021717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 15:48:21 2017
Response via : Initial Calibration

