

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG022818\
 Data File : BG032905.D
 Acq On : 1 Mar 2018 2:07
 Operator : SJ/JU
 Sample : PB106927TB
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB106927TB

Quant Time: Mar 01 02:55:36 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 17:50:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.92	152	73028	20.00	ng	0.00
21) Naphthalene-d8	11.80	136	324346	20.00	ng	0.00
38) Acenaphthene-d10	15.53	164	180579	20.00	ng	0.00
63) Phenanthrene-d10	18.26	188	391850	20.00	ng	0.00
75) Chrysene-d12	22.62	240	352478	20.00	ng	-0.01
86) Perylene-d12	26.42	264	347841	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	6.33	112	690549	151.28	ng	0.00
7) Phenol-d6	8.01	99	901851	141.74	ng	0.00
23) Nitrobenzene-d5	10.12	82	516273	107.49	ng	0.00
41) 2,4,6-Tribromophenol	17.00	330	267293	140.77	ng	0.00
44) 2-Fluorobiphenyl	14.16	172	1168610	93.33	ng	0.00
78) Terphenyl-d14	20.78	244	1547365	98.63	ng	0.00


Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG022818\
Data File : BG032905.D
Acq On : 1 Mar 2018 2:07
Operator : SJ/JU
Sample : PB106927TB
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB106927TB

Quant Time: Mar 01 02:55:36 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 17:50:45 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.92 min Scan# 958
Delta R.T. -0.01 min
Lab File: BG032905.D
Acq: 1 Mar 2018 2:07

Instrument :
BNA_G
ClientSampleId :
PB106927TB

Tot Ion:152 Resp: 73028
Ion Ratio Lower Upper
152 100
150 159.9 126.6 189.8
115 60.3 53.0 79.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5

2-Fluorophenol
Concen: 151.281 ng
RT: 6.33 min Scan# 518
Delta R.T. -0.00 min
Lab File: BG032905.D
Acq: 1 Mar 2018 2:07

Tgt Ion:112 Resp: 690549
Ion Ratio Lower Upper
112 100
64 64.8 56.3 84.5
63 33.4 30.1 45.1



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 141.745 ng

RT: 8.01 min Scan# 804

Delta R.T. -0.00 min

Lab File: BG032905.D

Acq: 1 Mar 2018 2:07

Instrument :

BNA_G

ClientSampleId :

PB106927TB


Tot Ion: 99 Resp: 901851

Ion Ratio Lower Upper



99 100

42 19.9 14.1 21.1

71 32.7 26.1 39.1



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 11.80 min Scan# 1449

Delta R.T. -0.01 min

Lab File: BG032905.D

Acq: 1 Mar 2018 2:07

Tgt Ion: 136 Resp: 324346


Ion Ratio Lower Upper

136 100



137 10.7 9.2 13.8


54 9.8 10.3 15.5#

68 6.3 4.5 6.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC





#23

Nitrobenzene-d5

Concen: 107.490 ng

RT: 10.12 min Scan# 1162

Delta R.T. -0.01 min

Lab File: BG032905.D

Acq: 1 Mar 2018 2:07

Instrument :

BNA_G

ClientSampleId :

PB106927TB


Tot Ion: 82 Resp: 516273

Ion Ratio Lower Upper




82 100

128 46.0 29.7 44.5#

54 54.6 43.1 64.7



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 128.00 (127.70 to 128.70): BGC
Ion 54.00 (53.70 to 54.70): BGC



#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 15.53 min Scan# 2084

Delta R.T. -0.01 min

Lab File: BG032905.D

Acq: 1 Mar 2018 2:07


Tgt Ion:164 Resp: 180579

Ion Ratio Lower Upper



164 100


162 104.5 78.8 118.2

160 49.9 35.4 53.0



Abundance Ion 164.00 (163.70 to 164.70): BGC
Ion 162.00 (161.70 to 162.70): BGC
Ion 160.00 (159.70 to 160.70): BGC








#41
 2,4,6-Tribromophenol
 Concen: 140.775 ng
 RT: 17.00 min Scan# 2334
 Delta R.T. -0.01 min
 Lab File: BG032905.D
 Acq: 1 Mar 2018 2:07

Instrument :
 BNA_G
 ClientSampleId :
 PB106927TB

Tot Ion:330 Resp: 267293
 Ion Ratio Lower Upper
 330 100
 332 95.5 77.0 115.6
 141 44.9 25.2 37.8#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 93.335 ng
 RT: 14.16 min Scan# 1851
 Delta R.T. -0.01 min
 Lab File: BG032905.D
 Acq: 1 Mar 2018 2:07

Tgt Ion:172 Resp: 1168610
 Ion Ratio Lower Upper
 172 100
 171 38.7 30.0 45.0
 170 25.6 19.5 29.3



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 18.26 min Scan# 2548
Delta R.T. -0.01 min
Lab File: BG032905.D
Acq: 1 Mar 2018 2:07




Instrument :
BNA_G
ClientSampleId :
PB106927TB

Tot Ion:188 Resp: 391850
Ion Ratio Lower Upper
188 100
94 10.3 6.9 10.3
80 12.1 7.7 11.5#



#75
Chrysene-d12
Concen: 20.000 ng
RT: 22.62 min Scan# 3290
Delta R.T. -0.01 min
Lab File: BG032905.D
Acq: 1 Mar 2018 2:07

Tgt Ion:240 Resp: 352478
Ion Ratio Lower Upper
240 100
120 13.2 6.8 10.2#
236 26.4 20.9 31.3





#78

Terphenyl-d14

Concen: 98.631 ng

RT: 20.78 min Scan# 2977

Delta R.T. -0.01 min

Lab File: BG032905.D

Acq: 1 Mar 2018 2:07

Instrument :

BNA_G

ClientSampleId :

PB106927TB


Tot Ion:244 Resp: 1547365

Ion Ratio Lower Upper

244 100

212 8.9 5.8 8.8#




122 14.5 7.0 10.4#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 26.42 min Scan# 3938

Delta R.T. -0.01 min

Lab File: BG032905.D

Acq: 1 Mar 2018 2:07


Tgt Ion:264 Resp: 347841

Ion Ratio Lower Upper

264 100

260 23.5 17.4 26.0

265 21.1 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

