

Data Path : Z:\HPCHEM1\BNA G\DATA\BG030116\
 Data File : BG021192.D
 Acq On : 1 Mar 2016 20:25
 Operator : UM/SJ
 Sample : H1578-02
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CEB-1(10-12)20160224

Manual Integrations
 APPROVED

Sohil
 3/2/2016 12:51:36 PM

Quant Time: Mar 02 11:22:45 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 29 17:40:25 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	10919	20.00	ng	-0.02
21) Naphthalene-d8	10.96	136	54091	20.00	ng	-0.01
38) Acenaphthene-d10	14.75	164	33144	20.00	ng	-0.01
63) Phenanthrene-d10	17.49	188	75701	20.00	ng	-0.01
75) Chrysene-d12	21.76	240	77702	20.00	ng	-0.01
86) Perylene-d12	25.02	264	71924	20.00	ng	-0.03

System Monitoring Compounds						
5) 2-Fluorophenol	5.70	112	98136	142.88	ng	-0.01
7) Phenol-d6	7.30	99	176675	156.43	ng	0.00
23) Nitrobenzene-d5	9.31	82	129347	101.19	ng	-0.01
41) 2,4,6-Tribromophenol	16.23	330	42815	124.16	ng	-0.01
44) 2-Fluorobiphenyl	13.38	172	195946	83.95	ng	-0.01
78) Terphenyl-d14	20.08	244	261969	81.68	ng	-0.01

Target Compounds				Qvalue		
31) Naphthalene	11.00	128	15675	5.56	ng	97
49) Dimethylphthalate	14.19	163	40853	13.83	ng	99
51) Acenaphthene	14.82	154	7293	3.32	ng	96
54) Dibenzofuran	15.15	168	12106	4.12	ng	91
57) Fluorene	15.79	166	14703	5.25	ng	94
70) Phenanthrene	17.53	178	180771	44.25	ng	99
71) Anthracene	17.62	178	35693	8.54	ng	97
72) Carbazole	17.88	167	18359	4.95	ng	# 94
73) Di-n-butylphthalate	18.43	149	46678	8.96	ng	# 98
74) Fluoranthene	19.53	202	206787	43.08	ng	99
77) Pyrene	19.89	202	169050	36.47	ng	99
79) Butylbenzylphthalate	20.80	149	2724220	1151.59	ng	# 84
80) Benzo(a)anthracene	21.73	228	87644	19.05	ng	99
82) Chrysene	21.80	228	79060	18.13	ng	99
83) Bis(2-ethylhexyl)phthalate	21.63	149	761307	219.57	ng	# 90
85) Indeno(1,2,3-cd)pyrene	28.74	276	46819	9.37	ng	# 100
87) Benzo(b)fluoranthene	23.98	252	96881	21.42	ng	# 95
88) Benzo(k)fluoranthene	24.04	252	32608m	7.20	ng	
89) Benzo(a)pyrene	24.86	252	72158	16.50	ng	# 92
90) Dibenzo(a,h)anthracene	28.79	278	13209m	3.05	ng	
91) Benzo(g,h,i)perylene	29.90	276	43734	10.46	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA G\DATA\BG030116\
Data File : BG021192.D
Acq On : 1 Mar 2016 20:25
Operator : UM/SJ
Sample : H1578-02
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
CEB-1(10-12)20160224

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM

Quant Time: Mar 02 11:22:45 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 29 17:40:25 2016
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.14 min Scan# 903
Delta R.T. -0.02 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Instrument :
BNA_G
ClientSampleId :
CEB-1(10-12)20160224




Tot Ion	Ratio	Lower	Upper
152	100		
150	157.0	123.4	185.2
115	60.2	43.3	64.9

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5

2-Fluorophenol
Concen: 142.88 ng
RT: 5.70 min Scan# 487
Delta R.T. -0.01 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Lower	Upper
112	100		
64	79.5	42.6	63.8#
63	42.6	21.7	32.5#



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 156.43 ng

RT: 7.30 min Scan# 759

Delta R.T. -0.00 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G

ClientSampleId :

CEB-1(10-12)20160224

Tot Ion: 99 Resp: 176675

Ion Ratio Lower Upper

99 100


42 20.3 13.2 19.8#

71 43.4 25.2 37.8#

Manual Integrations
APPROVED

Sohil




3/2/2016 12:51:36 PM



Abundance Ion 99.00 (98.70 to 99.70): BGC

Ion 42.00 (41.70 to 42.70): BGC

Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 10.96 min Scan# 1382

Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Tgt Ion: 136 Resp: 54091


Ion Ratio Lower Upper

136 100

137 10.1 8.8 13.2

54 10.6 6.9 10.3#

68 7.9 5.8 8.6






Abundance Ion 136.00 (135.70 to 136.70): BGC

Ion 137.00 (136.70 to 137.70): BGC

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC






#23
Nitrobenzene-d5
Concen: 101.19 ng
RT: 9.31 min Scan# 1102
Delta R.T. -0.01 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Instrument :
BNA_G
ClientSampleId :
CEB-1(10-12)20160224




Tot Ion	Ratio	Lower	Upper
82	100		
128	35.8	36.3	54.5#
54	50.1	37.6	56.4

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM






Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 128.00 (127.70 to 128.70): BGC
Ion 54.00 (53.70 to 54.70): BGC




#31
Naphthalene
Concen: 5.56 ng
RT: 11.00 min Scan# 1390
Delta R.T. -0.01 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.5	9.4	14.2
127	12.3	10.8	16.2

Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 129.00 (128.70 to 129.70): BGC
Ion 127.00 (126.70 to 127.70): BGC





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.75 min Scan# 2028

Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G





ClientSampleId :

CEB-1(10-12)20160224

Tot Ion:	164	Resp:	33144
Ion Ratio	Lower	Upper	
164	100		
162	103.2	80.6	120.8
160	46.4	36.5	54.7

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM



#41

2,4,6-Tribromophenol

Concen: 124.16 ng




RT: 16.23 min Scan# 2280


Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Tgt Ion:	330	Resp:	42815
Ion Ratio	Lower	Upper	
330	100		
332	95.5	77.4	116.0
141	60.3	35.3	52.9#









#44
2-Fluorobiphenyl
Concen: 83.95 ng
RT: 13.38 min Scan# 1795
Delta R.T. -0.01 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Instrument :
BNA_G
ClientSampleId :
CEB-1(10-12)20160224

Tot Ion	Ratio	Resp	Lower	Upper
172	100	195946		
171	36.5		29.4	44.0
170	25.2		19.8	29.8




Manual Integrations
APPROVED


Sohil
3/2/2016 12:51:36 PM



#49
Dimethylphthalate
Concen: 13.83 ng
RT: 14.19 min Scan# 1933
Delta R.T. -0.02 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	40853		
194	4.1		2.8	4.2
164	10.1		7.9	11.9





#51

Acenaphthene

Concen: 3.32 ng

RT: 14.82 min Scan# 2039

Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G





ClientSampleId :

CEB-1(10-12)20160224

Tot Ion:	154	Resp:	7293
Ion Ratio	Lower	Upper	
154	100		
153	118.0	94.2	141.2
152	61.5	42.9	64.3

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM



#54

Dibenzofuran

Concen: 4.12 ng




RT: 15.15 min Scan# 2095


Delta R.T. -0.02 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Tgt Ion:	168	Resp:	12106
Ion Ratio	Lower	Upper	
168	100		
139	43.8	29.8	44.8
169	15.6	11.4	17.2









#57
Fluorene
 Concen: 5.25 ng
 RT: 15.79 min Scan# 2205
 Delta R.T. -0.02 min
 Lab File: BG021192.D
 Acq: 1 Mar 2016 20:25

Instrument :
 BNA_G
 ClientSampleId :
 CEB-1(10-12)20160224

Tot Ion	Ratio	Lower	Upper
166	100		
165	95.4	81.3	121.9
167	14.1	10.6	16.0




Manual Integrations
 APPROVED


Sohil
 3/2/2016 12:51:36 PM



#63
Phenanthrene-d10
 Concen: 20.00 ng
 RT: 17.49 min Scan# 2494
 Delta R.T. -0.01 min
 Lab File: BG021192.D
 Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Lower	Upper
188	100		
94	10.0	8.2	12.4
80	12.3	9.2	13.8









#70
Phenanthrene
Concen: 44.25 ng
RT: 17.53 min Scan# 2501
Delta R.T. -0.01 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Instrument :
BNA_G
ClientSampleId :
CEB-1(10-12)20160224

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.8	16.2	24.4
179	15.5	12.9	19.3




Manual Integrations
APPROVED


Sohil
3/2/2016 12:51:36 PM



#71
Anthracene
Concen: 8.54 ng
RT: 17.62 min Scan# 2516
Delta R.T. -0.02 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.7	14.8	22.2
179	14.5	12.6	19.0





#72

Carbazole

Concen: 4.95 ng

RT: 17.88 min Scan# 2561

Delta R.T. -0.02 min

Lab File: BG021192.D


Acq: 1 Mar 2016 20:25

Instrument :

BNA_G

ClientSampleId :




CEB-1(10-12)20160224



Tot Ion:	167	Resp:	18359
Ion Ratio		Lower	Upper
167	100		
166	24.1	17.5	26.3
139	16.6	10.3	15.5#

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM



#73

Di-n-butylphthalate


Concen: 8.96 ng

RT: 18.43 min Scan# 2654



Delta R.T. -0.02 min


Lab File: BG021192.D

Acq: 1 Mar 2016 20:25



Tgt Ion:	149	Resp:	46678
Ion Ratio		Lower	Upper
149	100		
150	9.6	7.6	11.4
104	6.4	4.1	6.1#





#74

Fluoranthene

Concen: 43.08 ng

RT: 19.53 min Scan# 2841

Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G





ClientSampleId :

CEB-1(10-12)20160224

Tot Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	11.3	0.0	32.5
203	17.7	0.0	37.6

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM



#75

Chrysene-d12

Concen: 20.00 ng




RT: 21.76 min Scan# 3220


Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Resp Lower	Resp Upper
240	100		
120	10.6	7.5	11.3
236	26.4	20.5	30.7





#77

Pvrene

Concen: 36.47 ng

RT: 19.89 min Scan# 2902

Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G





ClientSampleId :

CEB-1(10-12)20160224

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	21.4	16.6	25.0
203	17.6	13.8	20.8

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM



#78

Terphenyl-d14

Concen: 81.68 ng




RT: 20.08 min Scan# 2935


Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	8.8	5.8	8.6#
122	10.8	8.3	12.5





#79

Butylbenzylphthalate

Concen: 1151.59 ng

RT: 20.80 min Scan# 3057

Delta R.T. 0.03 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G

ClientSampleId :

CEB-1(10-12)20160224

Tot Ion:149 Resp: 2724220

Ion Ratio Lower Upper

149 100


91 80.3 58.6 87.8

206 38.5 14.2 21.4#

Manual Integrations
APPROVED

Sohil




3/2/2016 12:51:36 PM



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 19.05 ng

RT: 21.73 min Scan# 3216

Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25


Tgt Ion:228 Resp: 87644

Ion Ratio Lower Upper

228 100

226 28.0 22.7 34.1



229 20.1 15.5 23.3




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 18.13 ng

RT: 21.80 min Scan# 3228

Delta R.T. -0.01 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G




ClientSampleId :

CEB-1(10-12)20160224

Tot Ion:	228	Resp:	79060
Ion Ratio	Lower	Upper	
228	100		
226	30.7	23.7	35.5
229	20.8	16.7	25.1

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM



#83

Bis(2-ethylhexyl)phthalate

Concen: 219.57 ng





RT: 21.63 min Scan# 3199


Delta R.T. -0.00 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Tgt Ion:	149	Resp:	761307
Ion Ratio	Lower	Upper	
149	100		
167	28.8	18.9	28.3#
279	5.1	3.0	4.6#









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 9.37 ng
 RT: 28.74 min Scan# 4408
 Delta R.T. -0.05 min
 Lab File: BG021192.D
 Acq: 1 Mar 2016 20:25

Instrument :
 BNA_G
 ClientSampleId :
 CEB-1(10-12)20160224

Tot Ion	Ratio	Resp Lower	Upper
276	100		
138	24.1	0.0	0.0#
277	26.4	0.0	0.0#




Manual Integrations
 APPROVED


Sohil
 3/2/2016 12:51:36 PM



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 25.02 min Scan# 3775
 Delta R.T. -0.03 min
 Lab File: BG021192.D
 Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Resp Lower	Upper
264	100		
260	24.4	18.5	27.7
265	21.9	18.6	28.0





#87

Benzo(b)fluoranthene

Concen: 21.42 ng

RT: 23.98 min Scan# 3598

Delta R.T. -0.02 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Instrument :

BNA_G





ClientSampleId :

CEB-1(10-12)20160224

Tot Ion:	252	Resp:	96881
Ion Ratio	Lower	Upper	
252	100		
253	23.0	16.8	25.2
125	12.8	8.2	12.2

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM



#88

Benzo(k)fluoranthene

Concen: 7.20 ng m




RT: 24.04 min Scan# 3609


Delta R.T. -0.03 min

Lab File: BG021192.D

Acq: 1 Mar 2016 20:25

Tgt Ion:	252	Resp:	32608
Ion Ratio	Lower	Upper	
252	100		
253	25.7	16.4	24.6
125	10.5	8.2	12.2









#89
Benzo(a)pyrene
Concen: 16.50 ng
RT: 24.86 min Scan# 3749
Delta R.T. -0.03 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Instrument :
BNA_G
ClientSampleId :
CEB-1(10-12)20160224

Tot Ion	Ratio	Lower	Upper
252	100		
253	25.4	15.8	23.8#
125	10.5	8.4	12.6




Manual Integrations
APPROVED


Sohil
3/2/2016 12:51:36 PM



#90
Dibenzo(a,h)anthracene
Concen: 3.05 ng m
RT: 28.79 min Scan# 4418
Delta R.T. -0.06 min
Lab File: BG021192.D
Acq: 1 Mar 2016 20:25

Tgt Ion	Ratio	Lower	Upper
278	100		
139	22.1	10.6	15.8#
279	29.1	19.7	29.5





#91

Benzo(a,h,i)perylene

Concen: 10.46 ng

RT: 29.90 min Scan# 4606

Delta R.T. -0.06 min

Lab File: BG021192.D


Acq: 1 Mar 2016 20:25

Instrument :

BNA_G

ClientSampleId :

CEB-1(10-12)20160224



Tot Ion:	276	Resp:	43734
Ion Ratio		Lower	Upper
276	100		
277	25.4	16.6	25.0#
138	21.0	14.6	21.8

Manual Integrations
APPROVED

Sohil
3/2/2016 12:51:36 PM

