

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030222\
 Data File : BG052546.D
 Acq On : 2 Mar 2022 20:58
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 03 02:27:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	41.096	-2.7	121	0.00
3	Pyridine	40.000	37.945	5.1	104	0.00
4	n-Nitrosodimethylamine	40.000	35.425	11.4	100	0.00
5 S	2-Fluorophenol	80.000	81.780	-2.2	115	0.00
6	Aniline	40.000	37.505	6.2	102	0.00
7 S	Phenol-d6	80.000	77.825	2.7	107	0.00
8	2-Chlorophenol	40.000	40.043	-0.1	115	0.00
9	Benzaldehyde	40.000	36.277	9.3	102	0.00
10 C	Phenol	40.000	37.907	5.2	105	0.00
11	bis(2-Chloroethyl)ether	40.000	36.929	7.7	106	0.00
12	1,3-Dichlorobenzene	40.000	40.337	-0.8	114	0.00
13 C	1,4-Dichlorobenzene	40.000	40.005	-0.0	113	0.00
14	1,2-Dichlorobenzene	40.000	39.109	2.2	111	0.00
15	Benzyl Alcohol	40.000	36.827	7.9	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.878	15.3	98	0.00
17	2-Methylphenol	40.000	39.188	2.0	109	0.00
18	Hexachloroethane	40.000	38.429	3.9	112	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.788	13.0	98	0.00
20	3+4-Methylphenols	40.000	38.846	2.9	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	37.382	6.5	100	0.00
23 S	Nitrobenzene-d5	80.000	75.955	5.1	100	0.00
24	Nitrobenzene	40.000	37.676	5.8	98	0.00
25	Isophorone	40.000	37.715	5.7	99	0.00
26 C	2-Nitrophenol	40.000	41.887	-4.7	106	0.00
27	2,4-Dimethylphenol	40.000	39.663	0.8	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.359	4.1	101	0.00
29 C	2,4-Dichlorophenol	40.000	40.065	-0.2	105	0.00
30	1,2,4-Trichlorobenzene	40.000	39.516	1.2	106	0.00
31	Naphthalene	40.000	38.940	2.7	103	0.00
32	Benzoic acid	40.000	42.494	-6.2	118	0.01
33	4-Chloroaniline	40.000	39.271	1.8	101	0.00
34 C	Hexachlorobutadiene	40.000	38.082	4.8	103	0.00
35	Caprolactam	40.000	40.764	-1.9	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.545	-1.4	104	0.00
37	2-Methylnaphthalene	40.000	38.348	4.1	102	0.00
38	1-Methylnaphthalene	40.000	38.542	3.6	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.776	3.1	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.961	7.6	98	0.00
42 S	2,4,6-Tribromophenol	80.000	83.237	-4.0	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.424	-3.6	105	0.00
44	2,4,5-Trichlorophenol	40.000	40.232	-0.6	103	0.00
45 S	2-Fluorobiphenyl	80.000	77.779	2.8	101	0.00
46	1,1'-Biphenyl	40.000	39.359	1.6	102	0.00
47	2-Chloronaphthalene	40.000	39.554	1.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030222\
 Data File : BG052546.D
 Acq On : 2 Mar 2022 20:58
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 03 02:27:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 08 01:21:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.303	1.7	100	0.00
49	Acenaphthylene	40.000	39.353	1.6	102	0.00
50	Dimethylphthalate	40.000	39.447	1.4	103	0.00
51	2,6-Dinitrotoluene	40.000	39.863	0.3	101	0.00
52 C	Acenaphthene	40.000	39.471	1.3	101	0.00
53	3-Nitroaniline	40.000	41.945	-4.9	105	0.00
54 P	2,4-Dinitrophenol	40.000	42.218	-5.5	111	0.00
55	Dibenzofuran	40.000	39.174	2.1	103	0.00
56 P	4-Nitrophenol	40.000	40.858	-2.1	101	0.00
57	2,4-Dinitrotoluene	40.000	40.842	-2.1	103	0.00
58	Fluorene	40.000	39.725	0.7	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.168	-0.4	103	0.00
60	Diethylphthalate	40.000	39.628	0.9	103	0.00
61	4-Chlorophenyl-phenylether	40.000	38.769	3.1	102	0.00
62	4-Nitroaniline	40.000	40.672	-1.7	103	0.00
63	Azobenzene	40.000	37.323	6.7	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.268	-10.7	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.026	-0.1	103	0.00
67	4-Bromophenyl-phenylether	40.000	39.170	2.1	102	0.00
68	Hexachlorobenzene	40.000	38.273	4.3	100	0.00
69	Atrazine	40.000	39.602	1.0	101	0.00
70 C	Pentachlorophenol	40.000	41.525	-3.8	103	0.00
71	Phenanthrene	40.000	39.363	1.6	103	0.00
72	Anthracene	40.000	39.770	0.6	104	0.00
73	Carbazole	40.000	40.366	-0.9	106	0.00
74	Di-n-butylphthalate	40.000	40.585	-1.5	105	0.00
75 C	Fluoranthene	40.000	39.022	2.4	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	44.677	-11.7	102	0.00
78	Pyrene	40.000	40.491	-1.2	104	0.00
79 S	Terphenyl-d14	80.000	79.002	1.2	104	0.00
80	Butylbenzylphthalate	40.000	42.935	-7.3	109	0.00
81	Benzo(a)anthracene	40.000	40.162	-0.4	104	0.00
82	3,3'-Dichlorobenzidine	40.000	40.815	-2.0	104	0.00
83	Chrysene	40.000	39.119	2.2	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.636	-6.6	105	-0.01
85 c	Di-n-octyl phthalate	40.000	42.179	-5.4	106	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.146	2.1	102	0.01
88	Benzo(b)fluoranthene	40.000	39.607	1.0	103	0.00
89	Benzo(k)fluoranthene	40.000	39.040	2.4	100	0.00
90 C	Benzo(a)pyrene	40.000	39.681	0.8	103	0.00
91	Dibenzo(a,h)anthracene	40.000	39.090	2.3	102	-0.01
92	Benzo(g,h,i)perylene	40.000	39.100	2.2	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030222\
Data File : BG052546.D
Acq On : 2 Mar 2022 20:58
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 03 02:27:17 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 08 01:21:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0