


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG030320\
 Data File : BG044662.D
 Acq On : 3 Mar 2020 14:01
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG022420.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Feb 24 16:11:23 2020



AutoFind: Scans 2479, 2480, 2481; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.7	19359	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.3	19772	PASS
70	69	0.00	2	0.7	129	PASS
127	198	10	80	41.5	27053	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	65178	PASS
199	198	5	9	6.7	4396	PASS
275	198	10	60	27.5	17895	PASS
365	198	1	100	3.3	2162	PASS
441	198	0.01	100	13.3	8666	PASS
442	198	50	100	81.4	53072	PASS
443	442	15	24	19.4	10297	PASS