

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG030518\  
 Data File : BG033007.D  
 Acq On : 6 Mar 2018 3:23  
 Operator : SJ/JU  
 Sample : J1699-84  
 Misc :  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 B22

Quant Time: Mar 06 07:06:04 2018  
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 01 17:38:52 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.91  | 152  | 60775    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 11.80 | 136  | 268719   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 15.53 | 164  | 150262   | 20.00  | ng    | 0.00     |
| 63) Phenanthrene-d10        | 18.25 | 188  | 341119   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 22.61 | 240  | 315865   | 20.00  | ng    | -0.01    |
| 86) Perylene-d12            | 26.41 | 264  | 289135   | 20.00  | ng    | -0.02    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 6.33  | 112  | 518856   | 136.58 | ng    | 0.00     |
| 7) Phenol-d6                | 8.01  | 99   | 628190   | 118.64 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 10.11 | 82   | 468728   | 116.84 | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 17.00 | 330  | 240159   | 152.00 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 14.15 | 172  | 983443   | 94.39  | ng    | -0.01    |
| 78) Terphenyl-d14           | 20.77 | 244  | 1477990  | 105.13 | ng    | 0.00     |

Target Compounds Qvalue

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG030518\  
Data File : BG033007.D  
Acq On : 6 Mar 2018 3:23  
Operator : SJ/JU  
Sample : J1699-84  
Misc :  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
B22

Quant Time: Mar 06 07:06:04 2018  
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Mar 01 17:38:52 2018  
Response via : Initial Calibration

