


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG030518\
 Data File : BG032981.D
 Acq On : 5 Mar 2018 9:40
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG022118.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Mar 01 17:38:52 2018



AutoFind: Scans 2609, 2610, 2611; Background Corrected with Scan 2602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.6	39786	PASS
68	69	0.00	2	0.8	253	PASS
69	198	0.00	100	38.4	33466	PASS
70	69	0.00	2	0.6	206	PASS
127	198	10	80	52.2	45541	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	87216	PASS
199	198	5	9	6.8	5966	PASS
275	198	10	60	26.3	22952	PASS
365	198	1	100	3.5	3080	PASS
441	442	0.01	24	16.0	10288	PASS
442	198	50	100	73.7	64306	PASS
443	442	15	24	18.4	11804	PASS