

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG030617\
 Data File : BG026078.D
 Acq On : 6 Mar 2017 16:29
 Operator : SJ/MA
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02002

Quant Time: Mar 15 14:35:44 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG030317-PAH.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 15 14:13:45 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
3 S	2,4-Dichlorophenol-d3	0.296	0.297	-0.3	115	0.00
4	Naphthalene	0.978	1.008	-3.1	118	0.00
5	2-Methylnaphthalene	0.726	0.749	-3.2	118	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
7 S	Acenaphthylene-d8	1.705	1.759	-3.2	120	0.00
8	Acenaphthylene	1.742	1.755	-0.7	117	0.00
9 C	Acenaphthene	1.231	1.232	-0.1	118	0.00
10 S	Fluorene-d10	1.244	1.279	-2.8	120	0.00
11	Fluorene	1.390	1.420	-2.2	120	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
13	Pentachlorophenol	0.128	0.109	14.8	104	0.00
14	Phenanthrene	1.060	1.103	-4.1	120	0.00
15 S	Anthracene-d10	0.889	0.942	-6.0	121	0.00
16	Anthracene	1.071	1.116	-4.2	118	0.00
17 C	Fluoranthene	1.100	1.287	-17.0	123	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
19 S	Pyrene-d10	1.021	1.054	-3.2	126	0.00
20	Pyrene	1.273	1.300	-2.1	122	0.00
21	Benzo(a)anthracene	1.113	1.166	-4.8	128	0.00
22	Chrysene	1.039	1.134	-9.1	135	0.00
23 I	Perylene-d12	1.000	1.000	0.0	129	0.00
24	Benzo(b)fluoranthene	1.135	1.191	-4.9	134	0.00
25	Benzo(k)fluoranthene	1.084	1.114	-2.8	128	0.00
26 S	Benzo(a)pyrene-d12	0.863	0.907	-5.1	136	0.00
27 C	Benzo(a)pyrene	1.055	1.126	-6.7	135	0.00
28	Indeno(1,2,3-cd)pyrene	1.165	1.220	-4.7	134	0.00
29	Dibenzo(a,h)anthracene	0.924	0.971	-5.1	133	-0.01
30	Benzo(g,h,i)perylene	0.997	1.047	-5.0	132	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0