

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG030618\
 Data File : BG033020.D
 Acq On : 6 Mar 2018 13:43
 Operator : SJ/JU
 Sample : J1794-04
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 030118-JETT-F-U-M

Quant Time: Mar 06 17:52:11 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 06 17:38:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.91	152	57595	20.00	ng	0.00
21) Naphthalene-d8	11.80	136	257405	20.00	ng	0.00
38) Acenaphthene-d10	15.52	164	154038	20.00	ng	0.00
63) Phenanthrene-d10	18.26	188	358601	20.00	ng	0.00
75) Chrysene-d12	22.61	240	333952	20.00	ng	0.00
86) Perylene-d12	26.41	264	332761	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	6.32	112	145335	40.37	ng	0.00
7) Phenol-d6	8.00	99	122108	24.33	ng	0.00
23) Nitrobenzene-d5	10.11	82	413362	108.36	ng	0.00
41) 2,4,6-Tribromophenol	17.00	330	151889	93.78	ng	0.00
44) 2-Fluorobiphenyl	14.16	172	911857	85.38	ng	0.00
78) Terphenyl-d14	20.77	244	1265482	85.14	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG030618\
Data File : BG033020.D
Acq On : 6 Mar 2018 13:43
Operator : SJ/JU
Sample : J1794-04
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
030118-JETT-F-U-M

Quant Time: Mar 06 17:52:11 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 06 17:38:59 2018
Response via : Initial Calibration

