


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030623\
Data File : BG056834.D
Acq On : 6 Mar 2023 9:44
Operator : CG/JU
Sample : DFTPP499
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP499

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG022323.M
Title : SVOA CALIBRATION
Last Update : Thu Mar 02 00:12:28 2023



AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2479

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	18457	PASS
68	69	0.00	2	0.9	176	PASS
69	198	0.00	100	41.6	19970	PASS
70	69	0.00	2	0.3	53	PASS
127	198	10	80	41.3	19827	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	47987	PASS
199	198	5	9	6.8	3271	PASS
275	198	10	60	31.2	14994	PASS
365	198	1	100	5.4	2568	PASS
441	198	0.01	100	13.0	6217	PASS
442	198	50	100	76.4	36675	PASS
443	442	15	24	19.8	7275	PASS

m/z	Abundance
44.00	1230.0

Instrument :

BNA_G

ClientSampleId :

DFTPP499

m/z	Abundance
39.00	1274.0
39.90	257.0
44.00	800.0
50.00	5078.0
51.00	19416.0
52.00	909.0
56.00	628.0
57.00	1098.0
61.00	238.0
62.00	288.0
62.90	1120.0
65.10	455.0
69.00	20480.0
69.90	158.0
73.00	197.0
74.00	2472.0
75.00	3528.0
76.00	1319.0
77.00	22568.0
78.00	1313.0
79.00	1623.0
79.90	1117.0
81.00	1654.0
82.00	444.0
82.90	509.0
85.10	233.0
86.00	686.0
87.00	333.0
90.80	486.0
92.00	674.0
93.00	3073.0
95.90	156.0
98.00	2523.0
99.00	2067.0
101.00	1138.0
102.80	302.0
103.90	544.0
104.90	559.0
107.00	7294.0
108.00	1096.0
109.00	250.0
110.00	9899.0
111.00	2145.0
111.90	189.0
115.90	516.0
117.00	7008.0
118.00	519.0
122.00	464.0
122.90	662.0
124.00	516.0
125.00	306.0
127.00	19544.0
127.90	1291.0
129.00	8866.0
130.00	728.0
131.10	177.0
133.90	344.0
135.10	910.0
136.00	432.0
136.90	553.0
138.00	211.0
140.90	1193.0
141.90	310.0
143.10	287.0
146.00	175.0
147.00	536.0
148.00	1842.0
149.00	478.0
152.90	440.0
153.90	346.0
155.00	569.0
156.10	942.0
156.90	180.0
157.90	268.0
159.00	222.0
159.90	481.0
161.00	664.0
162.10	191.0
164.90	481.0
165.90	368.0
167.00	2437.0
168.00	1531.0
169.00	226.0
170.00	158.0
170.90	199.0
171.90	229.0
173.00	223.0
174.00	603.0
175.00	1282.0
176.00	239.0
177.00	701.0
178.90	2398.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

180.00	1551.0
181.00	845.0
181.90	202.0
184.00	222.0
185.00	751.0
186.00	6946.0
187.00	2126.0
188.00	194.0
189.00	570.0
191.00	203.0
191.90	813.0
193.00	719.0
193.90	177.0
196.00	1332.0
197.90	43912.0
199.00	3306.0
200.00	291.0
203.00	519.0
204.00	2278.0
205.00	3754.0
206.00	13243.0
207.00	1986.0
207.90	469.0
209.00	221.0
210.20	275.0
211.00	658.0
212.00	158.0
216.00	418.0
217.00	4274.0
218.00	598.0
221.00	1425.0
222.00	456.0
222.90	930.0
224.00	8030.0
225.00	2316.0
226.00	222.0
227.00	4017.0
228.00	487.0
228.90	777.0
231.00	297.0
234.00	271.0
235.00	160.0
237.00	206.0
239.00	178.0
240.90	375.0
242.00	589.0
243.00	373.0
244.00	5723.0
245.10	838.0
246.00	1461.0
246.90	408.0
248.80	240.0
253.60	165.0
255.00	32192.0
256.00	5015.0
257.00	609.0
258.00	2236.0
258.90	444.0
265.00	1055.0
266.00	159.0
272.90	953.0
274.00	2677.0
275.00	13031.0
276.00	2293.0
277.00	1343.0
278.10	196.0
283.10	261.0
285.00	210.0
293.00	230.0
296.00	4561.0
296.90	583.0
303.00	431.0
304.10	185.0
314.10	226.0
314.90	656.0
315.90	216.0
323.00	1167.0
324.00	270.0
326.90	346.0
334.00	796.0
335.20	348.0
345.90	286.0
352.00	454.0
353.10	191.0
354.00	411.0
365.00	2055.0
366.00	376.0
372.00	558.0
372.90	190.0
383.00	230.0
402.10	216.0
402.90	421.0
421.00	383.0
422.00	262.0
423.00	1401.0
424.00	284.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

441.10	4751.0
442.10	28496.0
443.10	5577.0
444.00	476.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

m/z	Abundance
38.00	230.0
39.00	1565.0
40.00	166.0
43.90	896.0
49.00	175.0
50.00	5243.0
51.00	21840.0
52.00	1096.0
54.90	157.0
56.00	900.0
56.90	1282.0
60.80	244.0
62.00	305.0
63.00	1106.0
64.00	176.0
65.00	695.0
68.00	527.0
69.00	23768.0
72.90	200.0
74.00	2643.0
75.00	4217.0
76.00	1306.0
77.00	27400.0
78.00	1576.0
79.00	1802.0
80.00	1997.0
81.00	2243.0
81.90	391.0
82.90	458.0
85.00	322.0
86.00	583.0
86.90	380.0
90.90	485.0
91.90	439.0
93.00	3671.0
94.00	202.0
98.00	2476.0
99.00	2143.0
100.00	179.0
101.00	828.0
103.00	207.0
103.90	740.0
105.00	586.0
106.10	324.0
107.00	7607.0
108.00	1442.0
110.00	12089.0
111.10	2080.0
111.90	286.0
115.90	473.0
117.00	7804.0
118.00	845.0
122.00	542.0
123.00	928.0
123.90	405.0
124.90	518.0
127.00	23208.0
128.00	1868.0
129.00	10333.0
129.90	969.0
131.00	238.0
134.10	284.0
135.00	1056.0
136.00	394.0
137.00	435.0
140.10	233.0
141.00	1901.0
141.90	348.0
143.00	346.0
146.00	220.0
147.00	911.0
148.00	1685.0
148.90	430.0
150.90	244.0
153.00	526.0
153.90	439.0
155.00	1067.0
156.00	1181.0
156.80	208.0
157.90	508.0
158.80	174.0
160.00	455.0
160.90	758.0
161.80	269.0
165.00	583.0
166.00	490.0
167.00	2950.0
168.00	1610.0
169.00	320.0
171.00	196.0
172.10	254.0
172.90	380.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

174.00	924.0
175.00	1126.0
175.90	246.0
177.00	840.0
178.00	274.0
178.90	2682.0
180.00	1671.0
181.00	610.0
182.90	184.0
183.70	167.0
185.00	1312.0
186.00	9760.0
187.00	2481.0
188.00	220.0
189.10	702.0
190.00	161.0
190.90	345.0
192.00	954.0
193.00	1264.0
193.90	286.0
196.00	1874.0
197.90	58616.0
198.90	3874.0
200.00	311.0
201.30	216.0
203.00	582.0
204.00	2452.0
205.00	4338.0
206.00	18712.0
207.00	2898.0
208.00	750.0
208.90	262.0
210.90	708.0
214.80	249.0
215.90	460.0
217.00	5590.0
218.00	678.0
221.00	2169.0
221.90	626.0
223.00	1175.0
224.00	10367.0
225.00	2579.0
226.10	344.0
227.00	5303.0
228.00	906.0
228.90	955.0
229.90	188.0
231.00	413.0
233.90	378.0
235.00	393.0
235.90	246.0
237.00	452.0
238.90	172.0
239.90	290.0
240.90	295.0
242.00	626.0
243.00	737.0
244.00	7960.0
245.00	1007.0
246.00	1942.0
246.90	342.0
248.90	250.0
251.80	156.0
252.90	255.0
255.00	44144.0
256.00	6965.0
257.10	494.0
258.00	3140.0
259.00	498.0
265.00	1146.0
265.90	168.0
270.80	223.0
272.00	176.0
272.90	1708.0
274.00	3868.0
275.00	18040.0
276.00	2561.0
277.00	2216.0
278.10	197.0
282.90	238.0
285.00	377.0
293.00	486.0
296.00	7127.0
297.00	1109.0
303.00	569.0
304.00	260.0
314.10	330.0
314.90	698.0
316.00	508.0
321.90	211.0
323.00	1858.0
324.00	213.0
326.90	373.0
328.00	198.0
331.90	226.0
332.90	360.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

334.00	1306.0
335.10	287.0
341.00	193.0
346.00	435.0
347.10	158.0
352.10	540.0
352.90	412.0
354.10	314.0
365.00	3176.0
365.90	456.0
370.90	181.0
371.90	1055.0
373.10	388.0
382.90	237.0
384.00	170.0
402.00	415.0
403.00	533.0
404.10	173.0
421.10	300.0
422.00	384.0
423.00	2766.0
424.00	595.0
441.10	7495.0
442.00	40816.0
443.10	7978.0
444.10	953.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

m/z	Abundance
38.10	232.0
39.10	1176.0
40.00	227.0
44.00	1088.0
50.00	4109.0
51.00	14115.0
52.00	803.0
56.00	312.0
57.00	1156.0
61.00	283.0
62.00	246.0
63.00	581.0
65.00	253.0
68.90	15661.0
74.00	1851.0
75.00	2869.0
76.00	863.0
77.00	17608.0
78.00	1286.0
79.00	1501.0
80.00	1127.0
81.00	1293.0
82.00	327.0
83.00	383.0
84.90	305.0
85.90	469.0
87.00	231.0
90.90	232.0
92.00	299.0
93.00	2381.0
98.00	1698.0
99.00	1851.0
99.90	190.0
101.00	822.0
103.00	219.0
104.00	559.0
105.00	369.0
107.00	5337.0
108.00	850.0
110.00	8523.0
110.90	1395.0
112.00	247.0
115.90	270.0
117.00	5610.0
117.90	450.0
121.90	367.0
122.90	489.0
123.90	359.0
124.90	322.0
127.00	16728.0
128.00	1066.0
129.00	7380.0
129.90	751.0
133.90	368.0
134.90	695.0
135.80	290.0
137.10	542.0
141.00	1189.0
142.00	304.0
143.00	176.0
145.80	224.0
146.90	616.0
148.00	1575.0
149.00	407.0
153.00	305.0
153.90	205.0
155.00	703.0
156.00	882.0
157.00	210.0
159.90	344.0
160.90	559.0
162.00	152.0
164.90	495.0
165.90	350.0
167.00	2307.0
168.00	1189.0
169.10	230.0
172.00	208.0
173.00	252.0
174.10	351.0
175.00	831.0
176.10	205.0
177.00	537.0
178.90	1830.0
180.00	1308.0
180.90	496.0
183.90	184.0
185.00	898.0
186.00	6280.0
187.00	1719.0
187.90	360.0
189.00	632.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

192.00	809.0
193.00	844.0
194.00	159.0
196.00	1166.0
197.90	41432.0
198.90	2633.0
200.00	425.0
202.90	552.0
204.10	2250.0
205.10	3343.0
206.00	13425.0
207.00	1947.0
208.00	667.0
208.90	217.0
210.20	166.0
211.00	257.0
214.90	206.0
216.10	313.0
216.90	3735.0
217.90	320.0
221.00	1691.0
221.90	589.0
223.00	1075.0
224.00	7905.0
225.00	2301.0
225.90	203.0
227.00	4310.0
228.00	598.0
228.90	789.0
230.00	216.0
231.10	258.0
233.90	208.0
235.00	255.0
236.00	220.0
236.90	251.0
238.80	217.0
240.00	175.0
241.10	185.0
242.00	408.0
243.00	453.0
244.00	6130.0
245.10	708.0
246.00	1433.0
246.80	264.0
249.10	336.0
252.70	191.0
253.00	181.0
255.00	33040.0
256.00	4952.0
257.00	469.0
258.00	2007.0
259.00	283.0
265.00	930.0
273.10	1104.0
274.00	2957.0
275.00	13911.0
276.00	1935.0
276.90	1424.0
282.90	169.0
283.10	171.0
285.00	217.0
292.90	340.0
296.00	5784.0
297.10	634.0
303.00	586.0
304.00	150.0
307.90	188.0
314.00	284.0
315.00	845.0
316.10	258.0
323.00	1655.0
324.00	342.0
327.00	352.0
327.90	202.0
332.10	188.0
334.00	855.0
335.00	223.0
341.00	453.0
345.90	361.0
352.10	441.0
353.00	321.0
354.00	496.0
365.00	2473.0
365.90	278.0
372.20	687.0
373.00	262.0
383.00	263.0
401.90	308.0
403.10	520.0
404.00	202.0
420.90	233.0
423.10	2508.0
424.10	571.0
441.10	6405.0
442.10	40712.0
443.00	8271.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499

444.00

633.0

Instrument :
BNA_G
ClientSampleId :
DFTPP499