

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030624\
 Data File : BG060573.D
 Acq On : 6 Mar 2024 15:51
 Operator : MA/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Quant Time: Mar 07 00:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 07 00:49:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.322	152	66440	20.000	ng	0.00
21) Naphthalene-d8	11.166	136	291140	20.000	ng	0.00
39) Acenaphthene-d10	14.950	164	199904	20.000	ng	0.00
64) Phenanthrene-d10	17.700	188	467464	20.000	ng	0.00
76) Chrysene-d12	22.024	240	418758	20.000	ng	0.00
86) Perylene-d12	25.584	264	471163	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	9.121	70	9767	2.449	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030624\
Data File : BG060573.D
Acq On : 6 Mar 2024 15:51
Operator : MA/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Quant Time: Mar 07 00:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 07 00:49:22 2024
Response via : Initial Calibration

