

m/z	Abundance
37.10	154.0
38.10	436.0
39.10	3187.0
40.10	1632.0
41.10	300.0
43.10	361.0
44.10	4105.0
45.00	151.0
49.10	561.0
50.10	10474.0
51.10	44320.0
52.10	2211.0
55.00	328.0
56.00	1596.0
57.10	3793.0
58.00	182.0
61.10	548.0
62.10	564.0
63.10	1893.0
64.00	239.0
65.10	990.0
66.00	173.0
69.00	44768.0
70.10	424.0
73.10	488.0
74.10	4730.0
75.10	6247.0
76.10	2843.0
77.10	54448.0
78.10	4005.0
79.10	3062.0
80.10	2253.0
81.00	3854.0
82.00	909.0
83.10	859.0
84.00	571.0
85.10	857.0
86.10	1560.0
87.10	692.0
87.80	247.0
91.00	804.0
92.10	1020.0
93.00	4882.0
94.00	459.0
96.20	415.0
98.10	4837.0
99.10	4296.0
100.00	355.0
101.00	2137.0
102.00	169.0
103.00	809.0
104.10	1208.0
105.10	1731.0
106.10	298.0
107.10	16068.0
108.10	2365.0
110.10	32552.0
111.10	5018.0
112.00	494.0
113.10	156.0
116.00	1168.0
117.10	11156.0
118.00	668.0
119.00	350.0
120.00	414.0
122.00	1121.0
123.00	1928.0
123.90	912.0
125.00	1156.0
127.10	65280.0
128.10	4018.0
129.00	22288.0
130.10	2297.0
131.00	849.0
132.20	313.0
132.90	315.0
134.10	933.0
135.10	1766.0
136.10	754.0
137.00	943.0
140.10	465.0
141.00	2651.0
142.10	977.0
143.00	669.0
144.00	279.0
144.90	200.0
146.10	596.0
147.10	1442.0
148.00	3391.0
149.10	925.0
150.10	413.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

151.00	297.0
151.90	276.0
153.00	869.0
154.00	671.0
155.10	1734.0
156.10	2743.0
157.10	669.0
158.00	433.0
159.00	318.0
160.00	1120.0
161.00	1476.0
162.10	565.0
163.00	159.0
165.10	939.0
166.10	987.0
167.10	5728.0
168.10	3805.0
169.00	825.0
170.00	224.0
171.10	552.0
172.10	395.0
173.00	651.0
174.00	1347.0
175.10	2861.0
176.00	813.0
177.10	1140.0
178.10	394.0
179.00	4657.0
180.10	3039.0
181.00	1634.0
182.10	264.0
183.20	169.0
184.20	315.0
185.10	1811.0
186.10	17472.0
187.10	5066.0
188.00	361.0
189.00	1528.0
190.10	224.0
191.10	774.0
192.00	1606.0
193.00	1394.0
194.00	440.0
195.20	428.0
196.10	3374.0
198.00	138112.0
199.00	8861.0
200.00	600.0
201.50	778.0
202.20	377.0
203.00	1015.0
204.10	4272.0
205.00	8007.0
206.10	34264.0
207.10	5628.0
208.00	1096.0
209.00	349.0
210.30	555.0
211.10	873.0
212.10	316.0
213.00	151.0
215.00	402.0
216.10	1171.0
217.00	8579.0
218.00	861.0
221.00	6493.0
221.80	1445.0
223.00	2111.0
224.10	18448.0
225.10	4787.0
226.00	689.0
227.00	7443.0
227.90	701.0
229.00	1769.0
230.10	248.0
231.10	517.0
232.00	171.0
234.00	502.0
235.00	354.0
236.00	421.0
237.10	577.0
239.10	217.0
240.00	175.0
241.00	406.0
242.10	1200.0
243.10	1137.0
244.10	15344.0
245.10	1419.0
246.00	2909.0
247.00	510.0
247.90	237.0
249.10	813.0
249.90	264.0
252.10	208.0
253.00	350.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

253.30	347.0
255.10	73584.0
256.10	9706.0
257.00	813.0
258.10	3999.0
259.00	542.0
260.10	229.0
261.10	252.0
265.00	1477.0
266.00	400.0
271.00	379.0
272.00	444.0
273.00	2509.0
274.10	5315.0
275.10	30288.0
276.10	5117.0
277.10	2072.0
278.00	298.0
278.90	226.0
283.20	590.0
284.00	497.0
285.00	725.0
289.00	171.0
289.40	154.0
292.00	201.0
293.00	554.0
293.90	213.0
295.00	255.0
296.10	8202.0
297.00	1538.0
297.90	166.0
301.00	186.0
302.00	395.0
303.10	1557.0
304.10	458.0
313.10	188.0
314.10	427.0
315.00	1024.0
316.10	423.0
320.90	258.0
322.00	267.0
323.10	2658.0
324.00	608.0
327.00	444.0
328.00	571.0
331.90	393.0
333.00	237.0
334.00	1148.0
335.00	534.0
336.10	385.0
341.10	269.0
346.10	476.0
352.10	694.0
353.00	500.0
354.10	1004.0
355.20	165.0
365.00	3018.0
366.10	576.0
369.90	298.0
371.00	377.0
372.10	1364.0
373.00	338.0
383.10	464.0
384.00	326.0
390.10	295.0
391.00	182.0
402.00	505.0
403.00	613.0
404.20	513.0
421.10	414.0
422.00	530.0
423.10	4999.0
424.10	1205.0
441.10	14486.0
442.10	84288.0
443.10	17360.0
444.10	1601.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

m/z	Abundance
37.00	226.0
38.00	476.0
39.10	3135.0
40.00	1511.0
41.10	292.0
44.10	3507.0
45.10	196.0
48.10	183.0
49.10	847.0
50.10	12136.0
51.10	47248.0
52.10	2405.0
53.00	164.0
55.10	382.0
56.00	1623.0
57.10	4689.0
58.10	268.0
61.10	774.0
62.10	955.0
63.10	2465.0
64.00	206.0
65.10	1533.0
67.20	176.0
69.00	48696.0
70.10	502.0
73.10	437.0
74.10	5215.0
75.10	8655.0
76.10	3207.0
77.10	60544.0
78.10	4350.0
79.10	3826.0
80.10	2796.0
81.10	4040.0
82.10	964.0
83.10	994.0
84.00	545.0
85.00	606.0
86.00	1269.0
87.10	609.0
88.00	437.0
89.00	234.0
91.10	1125.0
92.00	551.0
93.00	5982.0
94.10	269.0
95.10	177.0
96.10	331.0
97.00	206.0
98.10	4996.0
99.10	4122.0
100.00	459.0
101.00	2686.0
102.10	285.0
103.00	861.0
104.10	1221.0
105.10	1567.0
106.20	572.0
107.10	19616.0
108.10	3091.0
109.10	801.0
110.10	38008.0
111.10	5203.0
112.00	449.0
113.10	389.0
116.10	1115.0
117.00	12354.0
118.10	1479.0
119.10	292.0
120.10	167.0
122.00	1151.0
123.00	1926.0
124.00	686.0
125.00	1232.0
127.10	75256.0
128.10	5530.0
129.10	26504.0
130.00	2083.0
131.00	679.0
132.00	440.0
133.00	221.0
134.10	860.0
135.10	2565.0
136.00	741.0
137.00	1466.0
138.10	302.0
139.10	251.0
140.10	423.0
141.10	2730.0
142.10	1238.0
143.00	758.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

144.20	353.0
146.10	634.0
147.10	2172.0
148.00	4004.0
149.00	1049.0
150.10	326.0
151.10	388.0
151.60	190.0
152.30	252.0
153.00	1104.0
154.10	1010.0
155.00	2089.0
156.10	2953.0
157.10	663.0
158.00	818.0
159.00	435.0
160.00	1148.0
161.10	2853.0
161.90	594.0
163.30	198.0
163.90	442.0
165.10	1494.0
166.10	1048.0
167.00	7429.0
168.10	3844.0
169.10	762.0
169.90	350.0
170.60	227.0
171.00	190.0
172.00	644.0
173.10	742.0
174.10	1462.0
175.10	2826.0
176.00	968.0
177.00	1955.0
178.00	550.0
179.00	6720.0
180.10	4068.0
181.00	1461.0
182.00	272.0
183.30	192.0
184.00	442.0
185.10	2490.0
186.10	23752.0
187.10	5964.0
188.00	910.0
189.10	1446.0
189.90	232.0
191.00	949.0
192.10	2031.0
193.10	2079.0
194.00	419.0
194.90	322.0
196.10	6318.0
198.00	177664.0
199.00	11942.0
200.00	1037.0
201.50	768.0
203.00	1249.0
204.10	6209.0
205.10	10585.0
206.10	43168.0
207.10	5915.0
208.00	1683.0
209.10	792.0
210.20	1007.0
211.10	1741.0
215.10	430.0
216.10	852.0
217.00	10860.0
218.00	1309.0
221.00	6474.0
222.00	1378.0
223.10	3312.0
224.10	22928.0
225.10	5707.0
226.10	383.0
227.00	8360.0
228.00	1325.0
229.10	2723.0
230.00	650.0
231.10	718.0
232.10	318.0
233.00	214.0
234.00	489.0
235.10	763.0
236.00	808.0
237.00	1304.0
237.90	288.0
239.10	393.0
240.10	490.0
241.10	660.0
242.10	1509.0
243.00	1243.0
244.10	19296.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

245.10	2559.0
246.00	3097.0
247.10	869.0
248.00	358.0
248.90	660.0
251.00	248.0
252.00	254.0
253.20	557.0
255.10	97056.0
256.00	14453.0
257.10	1783.0
258.00	5463.0
259.10	919.0
260.00	245.0
261.00	209.0
265.00	2562.0
265.90	267.0
269.20	174.0
271.10	467.0
272.10	343.0
273.00	2811.0
274.00	8343.0
275.10	45016.0
276.10	5364.0
277.00	3879.0
278.00	661.0
279.10	220.0
283.10	617.0
284.00	350.0
285.10	497.0
289.10	363.0
291.20	171.0
292.00	171.0
293.10	903.0
293.90	174.0
295.10	333.0
296.00	10778.0
297.10	2075.0
298.10	163.0
301.10	173.0
302.30	226.0
303.10	1158.0
304.00	415.0
310.00	179.0
312.90	164.0
314.10	684.0
315.00	1207.0
316.00	654.0
318.00	158.0
321.00	545.0
321.90	364.0
323.10	3979.0
324.10	616.0
327.00	503.0
328.20	394.0
328.90	151.0
332.00	194.0
333.00	347.0
334.10	2559.0
335.00	776.0
341.10	457.0
346.00	765.0
347.00	343.0
351.10	153.0
352.10	863.0
353.00	810.0
354.10	1270.0
355.10	366.0
359.00	173.0
365.00	4735.0
366.10	832.0
371.00	315.0
372.10	1999.0
373.00	555.0
374.00	158.0
378.00	163.0
383.10	1003.0
384.20	155.0
390.00	260.0
392.10	246.0
402.10	824.0
403.10	1129.0
404.10	533.0
421.10	991.0
422.00	1151.0
423.10	7569.0
424.00	1561.0
425.00	323.0
438.20	234.0
438.80	195.0
441.10	20696.0
442.10	134656.0
443.10	26312.0
444.10	2353.0
445.00	212.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

m/z	Abundance
37.00	252.0
38.10	382.0
39.10	2208.0
40.00	1477.0
41.10	307.0
42.10	156.0
43.20	249.0
44.10	3945.0
45.00	155.0
49.10	616.0
50.10	7372.0
51.10	31472.0
52.10	1933.0
55.10	298.0
56.00	1091.0
57.00	2202.0
61.10	723.0
62.10	268.0
63.00	1518.0
64.00	316.0
65.10	592.0
69.00	31264.0
70.20	360.0
73.20	464.0
74.10	3874.0
75.10	5981.0
76.10	2132.0
77.10	41160.0
78.10	2993.0
79.10	1620.0
80.10	2315.0
81.10	2333.0
82.10	811.0
83.00	1068.0
84.00	556.0
85.00	615.0
86.00	737.0
87.00	287.0
89.10	295.0
91.10	600.0
92.00	797.0
93.10	4453.0
94.00	261.0
96.10	224.0
97.00	207.0
98.00	3287.0
99.00	3040.0
100.00	367.0
101.00	1397.0
103.00	462.0
104.00	1409.0
105.10	1417.0
106.10	387.0
107.10	13350.0
108.10	2455.0
110.10	25864.0
111.00	4282.0
112.10	427.0
116.10	829.0
117.10	8962.0
118.00	858.0
122.00	1298.0
123.00	1125.0
124.00	544.0
125.10	642.0
127.10	49712.0
128.00	3947.0
129.10	19536.0
130.00	1736.0
131.00	435.0
132.10	212.0
134.10	531.0
135.10	1779.0
136.00	645.0
137.00	1056.0
140.00	531.0
141.00	2930.0
142.20	472.0
143.00	477.0
145.10	236.0
146.00	652.0
147.10	1178.0
148.00	3077.0
149.10	558.0
149.90	258.0
151.10	392.0
152.10	343.0
153.00	969.0
154.10	658.0
155.10	1243.0
156.10	2100.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

157.00	345.0
157.90	483.0
159.00	521.0
160.10	705.0
161.10	1131.0
162.00	480.0
162.90	181.0
165.00	897.0
166.10	685.0
167.10	4912.0
168.10	2998.0
169.10	495.0
170.90	191.0
171.90	397.0
173.10	462.0
174.00	834.0
175.10	2497.0
176.10	930.0
177.00	1004.0
178.00	572.0
179.00	3975.0
180.00	2961.0
181.10	1171.0
182.00	218.0
182.80	193.0
183.90	258.0
185.10	2219.0
186.10	17368.0
187.10	4846.0
188.20	669.0
189.00	733.0
190.00	200.0
191.10	450.0
192.00	1328.0
193.00	1424.0
194.00	669.0
195.20	693.0
196.00	3512.0
198.00	130776.0
199.00	8746.0
200.00	531.0
201.50	544.0
203.10	868.0
204.10	5323.0
205.00	8430.0
206.10	32512.0
207.10	3468.0
208.10	1246.0
209.10	454.0
210.00	565.0
211.00	867.0
212.10	348.0
213.10	239.0
215.10	435.0
216.10	860.0
217.00	7852.0
218.00	991.0
220.10	266.0
221.00	5684.0
221.90	1408.0
223.00	1989.0
224.10	20528.0
225.10	4249.0
226.00	593.0
227.00	7195.0
228.00	1137.0
229.00	1597.0
230.20	235.0
231.10	915.0
233.10	198.0
234.00	514.0
235.00	513.0
236.10	555.0
237.10	552.0
237.90	182.0
239.10	297.0
239.90	213.0
241.00	444.0
242.10	952.0
243.20	1068.0
244.10	14507.0
245.10	2152.0
246.00	3068.0
247.10	412.0
249.00	655.0
251.00	361.0
252.00	157.0
252.80	369.0
253.50	388.0
255.10	78696.0
256.00	12253.0
257.10	1333.0
258.10	3379.0
259.00	828.0
261.00	235.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

263.90	224.0
265.00	2022.0
266.00	291.0
271.00	425.0
272.00	318.0
273.00	2159.0
274.10	6658.0
275.10	37016.0
276.10	4762.0
277.00	2633.0
277.90	388.0
283.10	319.0
284.20	255.0
285.00	306.0
289.00	233.0
292.20	244.0
293.10	475.0
294.10	194.0
295.10	242.0
296.10	8985.0
297.00	1910.0
302.30	241.0
303.10	1352.0
304.10	188.0
308.10	156.0
310.90	175.0
314.00	695.0
315.00	1014.0
316.20	453.0
321.10	314.0
323.00	3386.0
324.00	734.0
327.10	663.0
328.00	302.0
332.10	230.0
333.00	368.0
334.10	2749.0
335.10	557.0
341.10	467.0
346.10	633.0
347.00	187.0
352.00	862.0
353.00	809.0
354.00	865.0
355.10	323.0
357.30	160.0
365.00	4296.0
366.10	602.0
371.10	294.0
372.10	1491.0
373.00	380.0
383.00	782.0
384.10	298.0
390.10	200.0
391.20	186.0
402.00	900.0
403.10	602.0
403.90	270.0
421.10	1004.0
422.10	925.0
423.10	6105.0
424.10	1291.0
439.50	187.0
441.10	20960.0
442.10	129464.0
443.10	25176.0
444.10	2818.0
445.00	216.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

m/z	Abundance
40.00	1170.0
41.10	251.0
43.10	154.0
44.10	3350.0
45.10	206.0
47.10	273.0
49.10	536.0
57.20	167.0
58.20	298.0
75.90	203.0
78.30	220.0
84.10	500.0
85.10	155.0
86.00	250.0
135.10	172.0
193.10	151.0
207.10	353.0


Instrument :
BNA_G
ClientSampleId :
DFTPP10

Data Path : Z:\HPCHEM1\BNA_G\Data\BG030717\
 Data File : BG026088.D
 Acq On : 7 Mar 2017 9:33
 Operator : SJ/MA
 Sample : DFTPP10
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP10

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG030717MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Mar 07 16:02:09 2017



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2460

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.6	41013	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	27.9	41576	PASS
70	69	0.00	2	1.0	428	PASS
127	198	10	80	42.6	63416	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	148850	PASS
199	198	5	9	6.6	9849	PASS
275	198	10	60	25.2	37440	PASS
365	198	1	100	2.7	4016	PASS
441	442	0.01	100	16.1	18714	PASS
442	198	50	100	78.0	116136	PASS
443	442	15	24	19.8	22949	PASS