

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030722\
 Data File : BG052591.D
 Acq On : 7 Mar 2022 15:16
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG030722

Quant Time: Mar 08 01:12:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 08 01:09:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.01
2	1,4-Dioxane	40.000	40.153	-0.4	99	0.00
3	Pyridine	40.000	41.060	-2.7	97	0.00
4	n-Nitrosodimethylamine	40.000	39.852	0.4	101	0.00
5 S	2-Fluorophenol	80.000	81.838	-2.3	102	0.01
6	Aniline	40.000	41.958	-4.9	103	0.01
7 S	Phenol-d6	80.000	83.142	-3.9	103	0.00
8	2-Chlorophenol	40.000	41.168	-2.9	102	0.01
9	Benzaldehyde	40.000	38.956	2.6	106	0.01
10 C	Phenol	40.000	41.342	-3.4	102	0.01
11	bis(2-Chloroethyl)ether	40.000	41.137	-2.8	103	0.00
12	1,3-Dichlorobenzene	40.000	40.614	-1.5	103	0.00
13 C	1,4-Dichlorobenzene	40.000	41.070	-2.7	102	0.01
14	1,2-Dichlorobenzene	40.000	40.154	-0.4	100	0.01
15	Benzyl Alcohol	40.000	42.232	-5.6	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.297	-0.7	101	0.00
17	2-Methylphenol	40.000	42.040	-5.1	106	0.01
18	Hexachloroethane	40.000	39.632	0.9	101	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	42.299	-5.7	105	0.00
20	3+4-Methylphenols	40.000	42.435	-6.1	106	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	40.080	-0.2	105	0.00
23 S	Nitrobenzene-d5	80.000	79.617	0.5	103	0.00
24	Nitrobenzene	40.000	39.875	0.3	104	0.00
25	Isophorone	40.000	40.814	-2.0	106	0.00
26 C	2-Nitrophenol	40.000	41.313	-3.3	102	0.00
27	2,4-Dimethylphenol	40.000	42.134	-5.3	109	0.01
28	bis(2-Chloroethoxy)methane	40.000	39.446	1.4	105	0.00
29 C	2,4-Dichlorophenol	40.000	42.536	-6.3	105	0.00
30	1,2,4-Trichlorobenzene	40.000	41.007	-2.5	107	0.00
31	Naphthalene	40.000	40.092	-0.2	105	0.00
32	Benzoic acid	40.000	40.472	-1.2	110	0.01
33	4-Chloroaniline	40.000	42.618	-6.5	108	0.00
34 C	Hexachlorobutadiene	40.000	38.899	2.8	103	0.00
35	Caprolactam	40.000	43.928	-9.8	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.974	-4.9	107	0.00
37	2-Methylnaphthalene	40.000	40.285	-0.7	105	0.00
38	1-Methylnaphthalene	40.000	40.231	-0.6	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.188	-0.5	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.665	3.3	106	0.00
42 S	2,4,6-Tribromophenol	80.000	83.472	-4.3	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.048	-5.1	107	0.01
44	2,4,5-Trichlorophenol	40.000	41.687	-4.2	106	0.00
45 S	2-Fluorobiphenyl	80.000	79.975	0.0	106	0.00
46	1,1'-Biphenyl	40.000	40.662	-1.7	107	0.00
47	2-Chloronaphthalene	40.000	40.388	-1.0	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030722\
 Data File : BG052591.D
 Acq On : 7 Mar 2022 15:16
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG030722

Quant Time: Mar 08 01:12:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 08 01:09:10 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.570	-3.9	105	0.01
49	Acenaphthylene	40.000	40.761	-1.9	107	0.00
50	Dimethylphthalate	40.000	40.566	-1.4	106	0.01
51	2,6-Dinitrotoluene	40.000	41.080	-2.7	108	0.00
52 C	Acenaphthene	40.000	40.756	-1.9	106	0.00
53	3-Nitroaniline	40.000	41.967	-4.9	105	0.00
54 P	2,4-Dinitrophenol	40.000	40.843	-2.1	110	0.00
55	Dibenzofuran	40.000	40.092	-0.2	105	0.00
56 P	4-Nitrophenol	40.000	44.590	-11.5	106	0.00
57	2,4-Dinitrotoluene	40.000	41.672	-4.2	106	0.00
58	Fluorene	40.000	40.910	-2.3	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.174	-7.9	107	0.00
60	Diethylphthalate	40.000	40.705	-1.8	105	0.00
61	4-Chlorophenyl-phenylether	40.000	40.163	-0.4	107	0.00
62	4-Nitroaniline	40.000	42.153	-5.4	105	0.00
63	Azobenzene	40.000	40.146	-0.4	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.963	-9.9	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.139	-0.3	106	0.00
67	4-Bromophenyl-phenylether	40.000	40.436	-1.1	108	0.00
68	Hexachlorobenzene	40.000	39.538	1.2	105	0.00
69	Atrazine	40.000	41.461	-3.7	105	0.00
70 C	Pentachlorophenol	40.000	40.728	-1.8	108	0.00
71	Phenanthrene	40.000	39.918	0.2	106	0.00
72	Anthracene	40.000	39.655	0.9	105	0.00
73	Carbazole	40.000	40.195	-0.5	105	0.00
74	Di-n-butylphthalate	40.000	40.233	-0.6	105	0.00
75 C	Fluoranthene	40.000	39.415	1.5	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	42.411	-6.0	95	0.00
78	Pyrene	40.000	41.589	-4.0	103	0.00
79 S	Terphenyl-d14	80.000	82.337	-2.9	103	0.00
80	Butylbenzylphthalate	40.000	41.629	-4.1	101	0.00
81	Benzo(a)anthracene	40.000	40.806	-2.0	101	0.00
82	3,3'-Dichlorobenzidine	40.000	41.035	-2.6	99	0.00
83	Chrysene	40.000	40.371	-0.9	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.698	-4.2	101	0.00
85 c	Di-n-octyl phthalate	40.000	41.649	-4.1	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.246	-0.6	102	0.01
88	Benzo(b)fluoranthene	40.000	39.547	1.1	100	0.00
89	Benzo(k)fluoranthene	40.000	40.054	-0.1	101	0.00
90 C	Benzo(a)pyrene	40.000	40.372	-0.9	101	0.00
91	Dibenzo(a,h)anthracene	40.000	40.351	-0.9	101	0.03
92	Benzo(g,h,i)perylene	40.000	40.249	-0.6	101	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030722\
Data File : BG052591.D
Acq On : 7 Mar 2022 15:16
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG030722

Quant Time: Mar 08 01:12:35 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 08 01:09:10 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0