


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030723\
Data File : BG056865.D
Acq On : 7 Mar 2023 10:28
Operator : CG/JU
Sample : DFTPP500
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP500

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG030723.M
Title : SVOA CALIBRATION
Last Update : Tue Mar 07 23:54:44 2023



AutoFind: Scans 2490, 2491, 2492; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.3	13905	PASS
68	69	0.00	2	0.7	99	PASS
69	198	0.00	100	42.7	15120	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	42.9	15174	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	35381	PASS
199	198	5	9	7.0	2486	PASS
275	198	10	60	32.6	11532	PASS
365	198	1	100	5.4	1928	PASS
441	198	0.01	100	12.4	4377	PASS
442	198	50	100	76.2	26944	PASS
443	442	15	24	19.7	5321	PASS

m/z	Abundance
44.00	873.0

Instrument :

BNA_G

ClientSampleId :

DFTPP500

m/z	Abundance
37.90	203.0
39.00	905.0
40.10	164.0
44.00	612.0
49.10	214.0
50.10	3633.0
51.00	13748.0
52.00	748.0
54.90	196.0
56.00	442.0
57.00	1047.0
61.00	254.0
62.00	155.0
63.00	564.0
65.00	204.0
68.00	296.0
68.90	14646.0
73.00	293.0
74.00	1376.0
75.00	2664.0
76.10	1068.0
77.00	15738.0
78.00	1005.0
79.00	1422.0
80.00	765.0
81.00	1504.0
82.10	217.0
83.00	361.0
86.00	476.0
90.80	194.0
91.90	287.0
93.00	2313.0
94.10	153.0
98.00	1854.0
99.00	1164.0
101.00	627.0
102.90	222.0
103.90	628.0
104.90	427.0
107.00	5046.0
107.90	662.0
110.00	6742.0
111.00	1261.0
111.90	215.0
116.00	189.0
117.00	4896.0
117.90	437.0
122.00	241.0
123.00	673.0
124.00	178.0
125.00	184.0
127.00	14138.0
127.90	942.0
129.00	6320.0
130.00	562.0
132.00	168.0
134.10	177.0
134.90	582.0
136.00	177.0
137.10	284.0
140.90	1126.0
141.90	296.0
143.10	204.0
146.00	249.0
146.90	608.0
147.90	1078.0
148.90	322.0
151.00	151.0
152.90	365.0
155.10	708.0
156.00	797.0
158.00	175.0
160.00	265.0
161.00	408.0
165.00	511.0
166.00	360.0
167.00	1897.0
168.00	630.0
172.00	184.0
172.90	351.0
174.00	419.0
175.00	941.0
176.90	322.0
179.00	1338.0
179.90	823.0
181.00	381.0
184.10	172.0
184.90	826.0
186.00	5398.0
187.00	1528.0
187.90	151.0
188.90	204.0

Instrument :
BNA_G
ClientSampleId :
DFTPP500

191.20	179.0
192.00	533.0
193.00	543.0
196.00	780.0
197.90	29216.0
198.90	2057.0
200.00	157.0
201.30	227.0
203.10	393.0
204.00	1355.0
205.00	2341.0
206.00	9276.0
207.00	1484.0
207.90	496.0
208.90	192.0
211.10	363.0
216.00	208.0
217.00	2977.0
217.80	444.0
221.00	1069.0
221.80	183.0
222.10	186.0
222.90	457.0
224.00	5349.0
225.00	1569.0
225.90	224.0
227.00	3241.0
227.80	339.0
228.80	526.0
231.10	242.0
236.00	193.0
241.10	154.0
241.90	335.0
243.00	362.0
244.00	4175.0
245.00	466.0
246.00	935.0
247.00	263.0
253.00	159.0
253.20	153.0
255.00	22248.0
256.00	3344.0
257.00	266.0
258.00	1718.0
259.00	292.0
264.90	825.0
273.00	532.0
274.00	1676.0
275.00	8347.0
276.00	1436.0
277.00	908.0
278.00	170.0
283.00	154.0
284.90	215.0
292.90	222.0
296.00	2810.0
296.90	469.0
303.00	265.0
314.00	180.0
315.00	445.0
323.00	672.0
326.90	190.0
333.90	666.0
335.00	270.0
346.00	251.0
351.90	370.0
353.00	249.0
354.00	249.0
365.00	1262.0
366.00	205.0
372.00	545.0
402.90	151.0
421.30	210.0
421.90	166.0
423.00	1151.0
424.10	257.0
441.10	3125.0
442.10	17664.0
443.10	3735.0
444.10	333.0

Instrument :
BNA_G
ClientSampleId :
DFTPP500

m/z	Abundance
38.00	231.0
39.00	1144.0
40.00	184.0
44.00	714.0
49.10	199.0
50.00	4849.0
51.00	16624.0
52.10	993.0
56.00	499.0
57.00	1097.0
61.00	201.0
62.00	296.0
63.00	803.0
65.00	426.0
69.00	18392.0
73.20	213.0
74.00	2326.0
75.00	3089.0
76.10	1024.0
77.00	19904.0
78.10	1473.0
79.00	1632.0
80.00	1447.0
80.90	1617.0
82.00	384.0
83.00	394.0
85.00	381.0
86.00	395.0
86.90	217.0
88.10	151.0
91.00	332.0
92.00	441.0
93.00	2681.0
93.90	195.0
98.00	1849.0
99.00	1293.0
100.90	744.0
103.00	384.0
103.90	646.0
105.00	686.0
107.00	7644.0
108.00	1062.0
108.90	171.0
110.00	9088.0
111.00	1699.0
111.90	318.0
116.00	425.0
117.00	5977.0
118.00	357.0
122.10	424.0
123.00	696.0
123.90	270.0
125.00	299.0
127.00	18216.0
128.00	1466.0
129.00	8211.0
130.00	904.0
134.00	308.0
134.90	696.0
136.00	321.0
136.90	474.0
141.00	998.0
142.00	442.0
142.90	356.0
146.00	252.0
147.00	566.0
148.00	1666.0
149.00	301.0
151.00	166.0
152.10	188.0
153.00	272.0
154.10	250.0
155.10	820.0
156.00	984.0
157.90	296.0
158.90	202.0
160.00	421.0
160.80	473.0
165.00	434.0
165.90	434.0
167.00	2684.0
168.00	1220.0
169.10	245.0
172.00	291.0
172.90	428.0
174.00	648.0
175.00	1194.0
175.80	202.0
177.00	612.0
178.90	2097.0
180.00	1409.0
181.00	710.0

Instrument :
BNA_G
ClientSampleId :
DFTPP500

183.80	185.0
185.10	953.0
186.00	6535.0
187.00	1937.0
187.90	162.0
189.00	628.0
191.00	347.0
192.00	491.0
193.00	853.0
194.00	269.0
196.00	1407.0
197.90	43728.0
199.00	3247.0
200.00	275.0
201.60	291.0
203.00	425.0
204.00	2065.0
205.00	3298.0
206.00	14306.0
207.00	1663.0
208.00	518.0
210.00	433.0
211.00	870.0
214.90	373.0
215.90	340.0
217.00	4159.0
217.90	661.0
219.90	223.0
221.00	1536.0
221.70	414.0
223.00	936.0
224.00	7939.0
225.10	2070.0
225.90	238.0
227.00	3879.0
228.00	540.0
229.00	879.0
230.90	292.0
234.00	352.0
235.00	253.0
236.90	378.0
238.90	197.0
239.90	274.0
241.00	256.0
242.00	612.0
243.10	491.0
244.00	5637.0
245.00	1038.0
246.00	1613.0
247.00	343.0
248.90	189.0
250.00	160.0
253.00	200.0
255.00	32960.0
256.00	5783.0
257.00	324.0
258.00	2757.0
259.10	424.0
265.00	945.0
265.90	165.0
272.00	164.0
272.20	164.0
273.00	966.0
274.00	2408.0
275.00	14018.0
276.00	1899.0
276.90	1499.0
277.90	262.0
282.90	269.0
284.00	228.0
284.90	196.0
293.00	260.0
293.90	166.0
296.00	4874.0
297.00	720.0
303.00	592.0
304.00	185.0
314.00	342.0
315.10	714.0
316.10	312.0
323.00	1589.0
323.90	182.0
327.00	325.0
328.00	161.0
332.80	160.0
333.90	1005.0
335.00	240.0
346.00	338.0
351.90	546.0
353.10	156.0
354.00	524.0
365.00	2287.0
366.00	503.0
370.90	160.0
372.00	435.0
373.00	197.0

Instrument :
BNA_G
ClientSampleId :
DFTPP500

391.10	159.0
401.90	164.0
403.10	400.0
404.10	176.0
420.90	253.0
422.10	343.0
423.00	1786.0
424.10	419.0
441.00	5591.0
442.00	32000.0
443.00	6398.0
443.90	617.0

Instrument :
BNA_G
ClientSampleId :
DFTPP500

m/z	Abundance
39.00	962.0
40.90	183.0
43.00	178.0
44.00	565.0
50.00	3012.0
51.00	11342.0
52.00	615.0
55.90	344.0
57.00	1082.0
61.00	293.0
61.90	158.0
63.00	620.0
65.10	160.0
69.00	12321.0
73.00	230.0
74.00	1350.0
75.00	2639.0
76.10	789.0
77.00	14135.0
78.00	1006.0
79.00	1018.0
80.00	812.0
80.90	999.0
83.00	325.0
85.10	218.0
86.00	389.0
90.80	247.0
92.00	315.0
92.90	1914.0
98.00	1352.0
98.90	1364.0
101.00	554.0
103.00	172.0
103.90	375.0
104.90	403.0
107.00	5820.0
108.00	544.0
108.80	163.0
110.00	6117.0
110.90	1201.0
116.00	248.0
117.00	4780.0
117.90	324.0
121.90	317.0
123.00	449.0
124.00	253.0
125.00	213.0
127.00	13169.0
128.00	960.0
129.00	6100.0
129.80	538.0
131.00	204.0
134.00	187.0
135.00	421.0
136.00	257.0
137.00	303.0
141.00	1129.0
142.00	250.0
142.80	185.0
147.00	403.0
148.00	1027.0
149.00	189.0
153.00	333.0
153.80	244.0
154.90	618.0
156.10	728.0
158.00	325.0
159.00	192.0
160.00	372.0
160.90	407.0
165.00	491.0
166.10	285.0
167.00	2192.0
168.00	852.0
172.00	332.0
173.00	223.0
174.10	428.0
175.00	837.0
176.00	211.0
177.10	498.0
177.80	218.0
179.00	1578.0
180.00	1047.0
181.00	368.0
185.00	794.0
186.00	5732.0
187.00	1622.0
187.80	169.0
189.00	388.0
191.00	264.0
192.00	616.0
193.00	535.0

Instrument :
BNA_G
ClientSampleId :
DFTPP500

196.10	1053.0
198.00	33200.0
199.00	2153.0
199.90	310.0
201.50	164.0
203.10	565.0
204.00	1522.0
205.00	3008.0
206.00	11770.0
207.00	1636.0
208.10	464.0
209.00	281.0
209.90	446.0
211.10	566.0
215.10	173.0
216.00	405.0
216.90	4110.0
217.90	246.0
221.00	984.0
222.00	537.0
223.00	610.0
224.00	6844.0
225.00	1914.0
225.90	212.0
227.00	3471.0
228.00	459.0
229.00	624.0
234.10	190.0
235.10	219.0
235.80	231.0
236.00	231.0
237.00	206.0
240.00	157.0
241.10	198.0
242.00	444.0
243.00	298.0
244.00	5565.0
244.90	405.0
246.00	1478.0
246.90	377.0
249.00	327.0
252.10	235.0
252.80	152.0
253.20	158.0
255.00	28168.0
256.00	4533.0
257.00	271.0
258.00	1992.0
258.90	298.0
265.00	925.0
273.00	1040.0
274.00	2335.0
275.00	12232.0
276.00	1580.0
277.00	909.0
277.90	159.0
282.90	151.0
284.00	204.0
285.00	184.0
292.90	407.0
296.00	4333.0
296.90	688.0
303.00	596.0
313.90	278.0
315.00	648.0
316.00	244.0
323.00	1175.0
326.90	291.0
328.00	227.0
331.90	186.0
333.00	189.0
334.10	975.0
340.90	225.0
345.90	411.0
352.20	293.0
352.90	259.0
354.10	444.0
355.00	159.0
364.90	2236.0
365.90	272.0
371.10	195.0
372.00	638.0
373.10	157.0
383.00	228.0
402.10	307.0
402.90	496.0
421.10	222.0
421.90	248.0
423.10	1791.0
424.00	333.0
441.00	4414.0
442.00	31168.0
443.10	5829.0
444.10	704.0

Instrument :
BNA_G
ClientSampleId :
DFTPP500