

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030724\
 Data File : BG060592.D
 Acq On : 7 Mar 2024 17:02
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG030724

Quant Time: Mar 08 03:37:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 08 03:35:45 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	39.893	0.3	97	0.00
3	Pyridine	40.000	38.429	3.9	91	0.00
4	n-Nitrosodimethylamine	40.000	38.555	3.6	91	0.00
5 S	2-Fluorophenol	80.000	81.884	-2.4	101	0.00
6	Aniline	40.000	40.247	-0.6	95	0.00
7 S	Phenol-d6	80.000	80.573	-0.7	95	0.00
8	2-Chlorophenol	40.000	40.984	-2.5	95	0.00
9	Benzaldehyde	40.000	39.726	0.7	92	0.00
10 C	Phenol	40.000	40.447	-1.1	94	0.00
11	bis(2-Chloroethyl)ether	40.000	39.755	0.6	95	0.00
12	1,3-Dichlorobenzene	40.000	39.154	2.1	95	0.00
13 C	1,4-Dichlorobenzene	40.000	39.118	2.2	94	0.00
14	1,2-Dichlorobenzene	40.000	40.501	-1.3	97	0.00
15	Benzyl Alcohol	40.000	41.141	-2.9	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.912	0.2	93	0.00
17	2-Methylphenol	40.000	40.389	-1.0	94	0.00
18	Hexachloroethane	40.000	40.941	-2.4	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.097	-2.7	95	0.00
20	3+4-Methylphenols	40.000	41.291	-3.2	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	39.331	1.7	95	0.00
23 S	Nitrobenzene-d5	80.000	91.167	-14.0	104	0.00
24	Nitrobenzene	40.000	44.549	-11.4	101	0.00
25	Isophorone	40.000	39.812	0.5	96	0.00
26 C	2-Nitrophenol	40.000	41.512	-3.8	105	0.00
27	2,4-Dimethylphenol	40.000	38.100	4.7	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.183	2.0	94	0.00
29 C	2,4-Dichlorophenol	40.000	39.803	0.5	95	0.00
30	1,2,4-Trichlorobenzene	40.000	40.293	-0.7	100	0.00
31	Naphthalene	40.000	40.268	-0.7	97	0.00
32	Benzoic acid	40.000	41.683	-4.2	107	0.00
33	4-Chloroaniline	40.000	40.022	-0.1	96	0.00
34 C	Hexachlorobutadiene	40.000	41.197	-3.0	99	0.00
35	Caprolactam	40.000	38.911	2.7	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.934	0.2	95	0.00
37	2-Methylnaphthalene	40.000	39.558	1.1	96	0.00
38	1-Methylnaphthalene	40.000	39.264	1.8	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.067	-2.7	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.917	-9.8	101	0.00
42 S	2,4,6-Tribromophenol	80.000	87.183	-9.0	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.882	-7.2	100	0.00
44	2,4,5-Trichlorophenol	40.000	41.680	-4.2	98	0.00
45 S	2-Fluorobiphenyl	80.000	80.615	-0.8	97	0.00
46	1,1'-Biphenyl	40.000	40.468	-1.2	97	0.00
47	2-Chloronaphthalene	40.000	40.219	-0.5	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030724\
 Data File : BG060592.D
 Acq On : 7 Mar 2024 17:02
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG030724

Quant Time: Mar 08 03:37:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 08 03:35:45 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.700	-4.3	106	0.00
49	Acenaphthylene	40.000	40.295	-0.7	96	0.00
50	Dimethylphthalate	40.000	40.542	-1.4	97	0.00
51	2,6-Dinitrotoluene	40.000	42.432	-6.1	106	0.00
52 C	Acenaphthene	40.000	39.806	0.5	95	0.00
53	3-Nitroaniline	40.000	41.897	-4.7	106	0.00
54 P	2,4-Dinitrophenol	40.000	42.033	-5.1	112	0.00
55	Dibenzofuran	40.000	40.236	-0.6	96	0.00
56 P	4-Nitrophenol	40.000	43.696	-9.2	105	0.00
57	2,4-Dinitrotoluene	40.000	43.059	-7.6	108	0.00
58	Fluorene	40.000	39.851	0.4	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.234	-5.6	99	0.00
60	Diethylphthalate	40.000	40.572	-1.4	99	0.00
61	4-Chlorophenyl-phenylether	40.000	39.841	0.4	96	0.00
62	4-Nitroaniline	40.000	42.138	-5.3	104	0.00
63	Azobenzene	40.000	40.028	-0.1	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.379	-5.9	113	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.534	-1.3	96	0.00
67	4-Bromophenyl-phenylether	40.000	40.915	-2.3	98	0.00
68	Hexachlorobenzene	40.000	40.433	-1.1	96	0.00
69	Atrazine	40.000	39.920	0.2	96	0.00
70 C	Pentachlorophenol	40.000	43.533	-8.8	101	0.00
71	Phenanthrene	40.000	40.302	-0.8	96	0.00
72	Anthracene	40.000	41.042	-2.6	97	0.00
73	Carbazole	40.000	40.716	-1.8	97	0.00
74	Di-n-butylphthalate	40.000	41.095	-2.7	97	0.00
75 C	Fluoranthene	40.000	40.575	-1.4	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	42.224	-5.6	98	0.00
78	Pyrene	40.000	40.466	-1.2	99	0.00
79 S	Terphenyl-d14	80.000	80.256	-0.3	99	0.00
80	Butylbenzylphthalate	40.000	42.746	-6.9	100	0.00
81	Benzo(a)anthracene	40.000	40.911	-2.3	102	0.00
82	3,3'-Dichlorobenzidine	40.000	41.707	-4.3	101	0.00
83	Chrysene	40.000	40.134	-0.3	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.226	-8.1	103	0.00
85 c	Di-n-octyl phthalate	40.000	42.241	-5.6	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.273	-3.2	102	0.00
88	Benzo(b)fluoranthene	40.000	41.445	-3.6	101	0.00
89	Benzo(k)fluoranthene	40.000	40.437	-1.1	101	0.00
90 C	Benzo(a)pyrene	40.000	40.998	-2.5	100	0.00
91	Dibenzo(a,h)anthracene	40.000	41.599	-4.0	101	0.00
92	Benzo(g,h,i)perylene	40.000	41.436	-3.6	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030724\
Data File : BG060592.D
Acq On : 7 Mar 2024 17:02
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG030724

Quant Time: Mar 08 03:37:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 08 03:35:45 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0