



2,4-Dinitrophenol



Response = 2.254e-001 * Amt - 9.622e-002
 Coef of Det (r^2) = 0.998256 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG030922.M
 Calibration Table Last Updated: Thu Mar 10 06:37:54 2022

Pentachlorophenol



Response = 1.478e-001 * Amt - 4.248e-002
 Coef of Det (r^2) = 0.998588 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG030922.M
 Calibration Table Last Updated: Thu Mar 10 06:37:54 2022