

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030922\
 Data File : BG052623.D
 Acq On : 9 Mar 2022 22:16
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG030922

Quant Time: Mar 10 06:39:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 06:37:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.533	0.493	7.5	111	0.00
3	Pyridine	1.409	1.388	1.5	113	0.00
4	n-Nitrosodimethylamine	0.671	0.643	4.2	110	0.00
5 S	2-Fluorophenol	1.158	1.133	2.2	114	0.00
6	Aniline	1.891	1.887	0.2	113	0.00
7 S	Phenol-d6	1.676	1.663	0.8	115	0.00
8	2-Chlorophenol	1.222	1.200	1.8	113	0.00
9	Benzaldehyde	0.959	0.905	5.6	113	0.00
10 C	Phenol	1.657	1.631	1.6	115	0.00
11	bis(2-Chloroethyl)ether	1.273	1.204	5.4	110	0.00
12	1,3-Dichlorobenzene	1.456	1.399	3.9	113	0.00
13 C	1,4-Dichlorobenzene	1.483	1.420	4.2	113	0.00
14	1,2-Dichlorobenzene	1.436	1.350	6.0	110	0.00
15	Benzyl Alcohol	1.296	1.304	-0.6	116	0.00
16	2,2'-oxybis(1-Chloropropane	1.717	1.624	5.4	111	0.00
17	2-Methylphenol	1.096	1.065	2.8	112	0.00
18	Hexachloroethane	0.502	0.478	4.8	112	0.00
19 P	n-Nitroso-di-n-propylamine	1.134	1.113	1.9	115	0.00
20	3+4-Methylphenols	1.516	1.515	0.1	114	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
22	Acetophenone	0.517	0.514	0.6	115	0.00
23 S	Nitrobenzene-d5	0.441	0.443	-0.5	114	0.00
24	Nitrobenzene	0.413	0.414	-0.2	113	0.00
25	Isophorone	0.734	0.752	-2.5	115	0.00
26 C	2-Nitrophenol	0.187	0.194	-3.7	118	0.00
27	2,4-Dimethylphenol	0.270	0.286	-5.9	117	0.00
28	bis(2-Chloroethoxy)methane	0.442	0.443	-0.2	113	0.00
29 C	2,4-Dichlorophenol	0.321	0.334	-4.0	114	0.00
30	1,2,4-Trichlorobenzene	0.397	0.400	-0.8	117	0.00
31	Naphthalene	1.052	1.055	-0.3	114	0.00
32	Benzoic acid	0.157	0.115	26.8#	130	0.00
33	4-Chloroaniline	0.426	0.448	-5.2	118	0.00
34 C	Hexachlorobutadiene	0.259	0.258	0.4	114	0.00
35	Caprolactam	0.108	0.118	-9.3	123	0.00
36 C	4-Chloro-3-methylphenol	0.351	0.364	-3.7	117	0.00
37	2-Methylnaphthalene	0.760	0.777	-2.2	115	0.00
38	1-Methylnaphthalene	0.747	0.762	-2.0	115	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
40	1,2,4,5-Tetrachlorobenzene	0.669	0.662	1.0	116	0.00
41 P	Hexachlorocyclopentadiene	0.300	0.314	-4.7	115	0.00
42 S	2,4,6-Tribromophenol	0.284	0.294	-3.5	114	0.00
43 C	2,4,6-Trichlorophenol	0.412	0.442	-7.3	118	0.00
44	2,4,5-Trichlorophenol	0.447	0.454	-1.6	114	0.00
45 S	2-Fluorobiphenyl	1.482	1.463	1.3	114	0.00
46	1,1'-Biphenyl	1.496	1.487	0.6	113	0.00
47	2-Chloronaphthalene	1.164	1.151	1.1	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030922\
 Data File : BG052623.D
 Acq On : 9 Mar 2022 22:16
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG030922

Quant Time: Mar 10 06:39:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 06:37:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.375	0.380	-1.3	113	0.00
49	Acenaphthylene	1.843	1.844	-0.1	114	0.00
50	Dimethylphthalate	1.519	1.516	0.2	114	0.00
51	2,6-Dinitrotoluene	0.332	0.332	0.0	117	0.00
52 C	Acenaphthene	1.117	1.115	0.2	114	0.00
53	3-Nitroaniline	0.342	0.351	-2.6	113	0.00
54 P	2,4-Dinitrophenol	0.157	0.177	-12.7	119	0.00
55	Dibenzofuran	1.895	1.885	0.5	115	0.00
56 P	4-Nitrophenol	0.238	0.258	-8.4	114	0.00
57	2,4-Dinitrotoluene	0.467	0.482	-3.2	118	0.00
58	Fluorene	1.539	1.515	1.6	113	0.00
59	2,3,4,6-Tetrachlorophenol	0.410	0.446	-8.8	119	0.00
60	Diethylphthalate	1.525	1.517	0.5	114	0.00
61	4-Chlorophenyl-phenylether	0.858	0.850	0.9	114	0.00
62	4-Nitroaniline	0.361	0.362	-0.3	113	0.00
63	Azobenzene	1.506	1.492	0.9	114	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
65	4,6-Dinitro-2-methylphenol	0.127	0.135	-6.3	114	0.00
66 c	n-Nitrosodiphenylamine	0.562	0.571	-1.6	112	0.00
67	4-Bromophenyl-phenylether	0.248	0.248	0.0	114	0.00
68	Hexachlorobenzene	0.265	0.267	-0.8	115	0.00
69	Atrazine	0.207	0.222	-7.2	117	0.00
70 C	Pentachlorophenol	0.115	0.132	-14.8	114	0.00
71	Phenanthrene	1.049	1.044	0.5	112	0.00
72	Anthracene	1.027	1.015	1.2	110	0.00
73	Carbazole	1.005	1.007	-0.2	113	0.00
74	Di-n-butylphthalate	1.097	1.110	-1.2	112	0.00
75 C	Fluoranthene	1.310	1.273	2.8	108	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
77	Benzidine	0.415	0.398	4.1	97	0.00
78	Pyrene	1.337	1.347	-0.7	111	0.00
79 S	Terphenyl-d14	1.130	1.126	0.4	110	0.00
80	Butylbenzylphthalate	0.486	0.502	-3.3	110	0.00
81	Benzo(a)anthracene	1.340	1.345	-0.4	109	0.00
82	3,3'-Dichlorobenzidine	0.465	0.460	1.1	104	0.00
83	Chrysene	1.253	1.250	0.2	109	0.00
84	Bis(2-ethylhexyl)phthalate	0.680	0.700	-2.9	110	0.00
85 c	Di-n-octyl phthalate	1.127	1.176	-4.3	111	0.00
86 I	Perylene-d12	1.000	1.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	1.433	1.450	-1.2	109	0.00
88	Benzo(b)fluoranthene	1.235	1.249	-1.1	109	0.00
89	Benzo(k)fluoranthene	1.242	1.251	-0.7	110	0.00
90 C	Benzo(a)pyrene	1.045	1.067	-2.1	110	0.00
91	Dibenzo(a,h)anthracene	1.189	1.199	-0.8	109	0.01
92	Benzo(g,h,i)perylene	1.161	1.176	-1.3	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030922\
Data File : BG052623.D
Acq On : 9 Mar 2022 22:16
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG030922

Quant Time: Mar 10 06:39:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 10 06:37:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0