


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG031017\
 Data File : BG026164.D
 Acq On : 10 Mar 2017 8:35
 Operator : SJ/MA
 Sample : DFTPP13
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP13

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG030717MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Mar 13 18:09:30 2017



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2460

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.6	44357	PASS
68	69	0.00	2	1.7	741	PASS
69	198	0.00	100	30.5	42877	PASS
70	69	0.00	2	0.4	187	PASS
127	198	10	80	44.4	62381	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	140592	PASS
199	198	5	9	6.7	9358	PASS
275	198	10	60	24.2	34074	PASS
365	198	1	100	3.2	4438	PASS
441	442	0.01	100	17.0	16266	PASS
442	198	50	100	68.2	95917	PASS
443	442	15	24	20.3	19518	PASS

m/z	Abundance
38.10	268.0
39.10	2139.0
40.10	2069.0
44.10	3770.0
49.10	283.0
50.10	9047.0
51.10	38016.0
52.10	2403.0
55.10	366.0
56.10	1401.0
57.10	3487.0
58.10	221.0
61.10	557.0
62.10	630.0
63.10	1433.0
64.00	156.0
65.10	1002.0
66.10	152.0
68.10	1050.0
69.00	37472.0
70.10	330.0
73.20	339.0
74.10	3839.0
75.10	6099.0
76.10	1803.0
77.10	46920.0
78.10	3366.0
79.10	2564.0
80.00	1953.0
81.10	2852.0
82.10	656.0
83.10	802.0
83.90	260.0
85.00	360.0
86.10	755.0
87.00	245.0
88.00	385.0
91.10	668.0
92.10	708.0
93.10	4345.0
94.00	282.0
94.90	264.0
96.00	298.0
98.10	3599.0
99.00	3367.0
100.00	168.0
101.00	1571.0
103.10	504.0
104.10	1079.0
105.00	889.0
107.10	12978.0
108.10	1975.0
110.00	25064.0
111.00	4331.0
112.00	360.0
116.10	992.0
117.10	11137.0
118.10	956.0
119.10	216.0
119.80	267.0
121.20	176.0
122.10	869.0
123.10	2203.0
124.10	710.0
125.00	628.0
127.10	48936.0
128.10	4323.0
129.00	18656.0
130.10	1212.0
131.00	294.0
134.00	567.0
135.00	1968.0
136.10	401.0
137.00	927.0
138.00	284.0
140.00	166.0
141.00	2215.0
142.10	997.0
142.90	584.0
145.10	183.0
146.00	390.0
147.10	1283.0
148.00	2094.0
149.10	977.0
150.10	269.0
151.40	384.0
152.00	207.0
153.00	878.0
154.00	581.0
155.10	1291.0
156.00	1679.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

157.00	327.0
158.00	323.0
159.10	423.0
160.00	855.0
161.00	1315.0
162.00	354.0
164.00	167.0
165.00	1241.0
166.00	491.0
167.10	5602.0
168.00	3737.0
169.00	381.0
170.10	169.0
170.90	377.0
172.00	579.0
173.00	609.0
174.10	767.0
175.10	1950.0
176.10	310.0
177.10	868.0
178.00	377.0
179.00	3494.0
180.10	2825.0
181.00	787.0
181.90	247.0
182.90	189.0
183.90	321.0
185.10	1311.0
186.10	14505.0
187.10	4361.0
188.00	379.0
189.00	985.0
191.00	433.0
192.10	940.0
193.10	1389.0
194.00	392.0
195.00	185.0
196.10	2856.0
198.00	105680.0
199.00	6989.0
200.00	483.0
201.40	441.0
202.00	302.0
203.00	621.0
204.10	3446.0
205.10	5694.0
206.10	27112.0
207.10	3868.0
208.00	874.0
209.00	487.0
211.10	966.0
212.10	208.0
215.00	375.0
216.10	604.0
217.00	6799.0
218.00	868.0
221.00	3986.0
222.00	983.0
223.00	1561.0
224.10	14677.0
225.00	3660.0
226.20	727.0
227.00	5024.0
228.10	1354.0
229.00	1355.0
231.00	624.0
233.90	288.0
235.10	417.0
235.90	242.0
237.10	490.0
238.80	215.0
240.10	199.0
241.00	382.0
242.00	766.0
243.00	1160.0
244.10	11047.0
245.00	1351.0
246.10	1984.0
247.10	213.0
249.00	527.0
250.90	172.0
252.10	293.0
252.90	382.0
255.00	55808.0
256.00	8359.0
257.00	902.0
258.00	2764.0
259.00	305.0
260.10	216.0
261.10	155.0
265.00	1276.0
266.10	283.0
271.10	334.0
272.30	295.0
273.10	1844.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

274.00	4250.0
275.10	23976.0
276.00	3627.0
277.00	1624.0
278.00	575.0
283.10	208.0
283.80	165.0
285.10	303.0
286.10	230.0
288.10	209.0
289.00	167.0
293.00	611.0
295.10	188.0
296.00	6071.0
297.00	722.0
297.90	226.0
303.00	844.0
304.10	168.0
309.10	234.0
313.90	285.0
315.00	782.0
316.00	565.0
321.10	274.0
323.10	1688.0
324.00	201.0
327.00	307.0
328.10	187.0
332.10	207.0
333.00	166.0
334.10	1121.0
335.00	360.0
341.10	406.0
342.00	152.0
346.10	413.0
352.00	546.0
352.90	161.0
354.00	555.0
365.00	2621.0
366.10	187.0
372.10	1434.0
373.00	174.0
390.10	159.0
401.90	448.0
403.10	524.0
421.00	561.0
422.00	317.0
423.10	2307.0
424.10	555.0
441.10	8823.0
442.10	52656.0
443.10	12043.0
444.10	878.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

m/z	Abundance
37.10	205.0
38.10	445.0
39.10	3464.0
40.10	1847.0
44.10	3669.0
47.00	163.0
49.10	718.0
50.10	12274.0
51.10	54664.0
52.10	2659.0
55.10	284.0
56.10	1799.0
57.10	3496.0
61.10	902.0
62.10	960.0
63.10	2262.0
64.10	216.0
65.10	1259.0
67.20	156.0
68.10	1174.0
69.00	51784.0
70.10	232.0
73.10	396.0
74.10	5947.0
75.10	8043.0
76.10	3640.0
77.10	67944.0
78.10	4229.0
79.00	3978.0
80.00	2834.0
81.10	4236.0
82.00	1328.0
83.00	1528.0
84.00	238.0
85.00	816.0
86.10	1615.0
87.10	652.0
88.10	195.0
91.00	901.0
92.00	1106.0
93.00	7066.0
94.10	517.0
96.10	203.0
98.00	5400.0
99.10	5279.0
100.00	600.0
101.10	2655.0
102.00	331.0
103.00	670.0
104.10	1706.0
105.10	1713.0
106.10	687.0
107.10	18880.0
108.10	3258.0
110.00	37680.0
111.00	6144.0
112.00	665.0
113.10	262.0
116.00	1017.0
117.10	13839.0
118.10	1137.0
119.10	262.0
120.00	183.0
122.00	1079.0
123.10	2094.0
124.10	967.0
125.00	783.0
127.10	75904.0
128.10	6459.0
129.10	28160.0
130.00	2425.0
131.10	587.0
132.00	327.0
132.90	164.0
134.00	705.0
135.00	2666.0
136.10	896.0
137.00	1274.0
141.10	3275.0
142.00	1191.0
143.10	935.0
145.00	455.0
146.00	565.0
147.00	1601.0
148.00	4314.0
149.00	957.0
150.00	343.0
151.10	322.0
153.00	1377.0
154.10	1038.0
155.00	2636.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

156.10	2770.0
157.10	587.0
158.10	734.0
159.00	834.0
160.00	1263.0
161.10	1935.0
162.00	309.0
163.00	203.0
164.00	258.0
165.10	1445.0
166.00	930.0
167.10	8432.0
168.00	4719.0
169.00	640.0
170.00	276.0
170.50	215.0
171.00	434.0
172.10	670.0
173.00	904.0
174.00	1886.0
175.10	3146.0
176.00	994.0
177.00	818.0
178.00	471.0
179.00	6371.0
180.00	3645.0
181.00	1788.0
182.10	318.0
183.10	316.0
184.00	593.0
185.10	2831.0
186.10	19624.0
187.10	6328.0
188.10	591.0
189.00	1507.0
190.20	320.0
191.00	1046.0
192.00	1929.0
193.10	1518.0
194.00	419.0
195.00	292.0
196.10	4564.0
198.00	170432.0
199.00	11674.0
200.10	1054.0
201.30	691.0
203.00	1136.0
204.10	6361.0
205.10	11034.0
206.10	43464.0
207.10	5365.0
208.10	1430.0
209.00	444.0
210.00	715.0
210.50	706.0
211.10	1448.0
212.20	198.0
213.10	204.0
215.00	831.0
216.00	1076.0
217.00	9049.0
218.00	1343.0
221.00	7241.0
223.00	3130.0
224.10	25496.0
225.00	6342.0
226.10	478.0
227.00	9311.0
228.00	1037.0
229.00	2065.0
230.00	315.0
231.00	719.0
232.00	399.0
234.00	799.0
235.10	623.0
236.10	585.0
237.10	489.0
238.10	164.0
239.00	378.0
240.00	287.0
241.30	451.0
242.10	1224.0
243.10	1718.0
244.10	18792.0
245.10	2411.0
246.00	3515.0
247.00	325.0
248.00	210.0
249.00	529.0
251.10	174.0
252.10	342.0
253.20	595.0
255.00	93336.0
256.00	13990.0
257.10	1146.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

258.00	4573.0
259.10	1051.0
260.00	211.0
261.00	189.0
264.00	199.0
265.00	2151.0
266.00	322.0
267.00	158.0
270.00	212.0
271.10	541.0
272.10	300.0
273.00	2826.0
274.00	7443.0
275.10	39800.0
276.00	5946.0
277.00	3025.0
278.00	406.0
283.00	511.0
284.10	266.0
285.00	650.0
286.10	227.0
291.10	162.0
292.10	166.0
292.90	568.0
293.90	166.0
295.00	361.0
296.00	10966.0
297.10	1368.0
302.00	185.0
303.10	1456.0
304.00	301.0
309.90	223.0
314.10	501.0
315.10	926.0
316.00	432.0
317.20	168.0
320.90	332.0
322.10	263.0
323.10	3960.0
324.10	1064.0
327.10	648.0
328.10	412.0
332.00	400.0
333.10	283.0
334.10	2529.0
335.10	451.0
341.10	505.0
342.00	181.0
346.00	521.0
347.00	246.0
352.00	1022.0
353.00	629.0
354.10	900.0
355.00	280.0
362.90	156.0
365.00	5387.0
366.00	793.0
371.10	391.0
372.10	2008.0
373.20	481.0
383.00	481.0
390.10	216.0
401.10	262.0
402.10	521.0
403.10	893.0
404.00	440.0
421.00	794.0
422.10	865.0
423.10	5596.0
424.10	1195.0
438.00	345.0
441.10	18160.0
442.10	112688.0
443.10	22240.0
444.10	1988.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

m/z	Abundance
37.20	218.0
38.00	451.0
39.10	2361.0
40.00	2538.0
41.10	167.0
44.10	3438.0
49.00	190.0
50.10	9136.0
51.10	40392.0
52.10	2136.0
56.00	1302.0
57.10	3235.0
57.90	186.0
61.10	551.0
62.10	976.0
63.10	1693.0
64.00	357.0
65.10	1045.0
67.20	214.0
69.00	39376.0
72.20	240.0
73.00	245.0
74.10	4283.0
75.10	7281.0
76.10	3024.0
77.10	50704.0
78.10	3516.0
79.00	2933.0
80.00	2150.0
81.00	3570.0
82.00	796.0
83.00	1082.0
84.00	281.0
85.00	557.0
86.10	1245.0
87.20	336.0
88.00	225.0
91.00	785.0
92.00	920.0
93.00	5327.0
94.10	232.0
96.00	226.0
98.10	4057.0
99.00	3625.0
100.00	502.0
101.00	2381.0
103.00	556.0
104.10	837.0
105.00	1099.0
107.10	15399.0
108.10	2253.0
110.00	29824.0
111.00	3981.0
112.10	569.0
113.10	262.0
116.10	1060.0
117.10	11739.0
118.10	868.0
120.00	251.0
122.10	1350.0
123.00	1511.0
124.10	427.0
125.00	664.0
127.10	62304.0
128.10	5144.0
129.00	23104.0
130.00	1802.0
131.00	489.0
132.10	199.0
132.90	230.0
134.10	512.0
135.00	1913.0
136.00	757.0
137.00	920.0
138.90	176.0
140.00	206.0
141.00	3225.0
142.10	964.0
143.00	777.0
144.10	353.0
146.20	517.0
147.00	1239.0
148.00	3144.0
149.00	1122.0
150.00	269.0
151.00	327.0
151.90	231.0
153.10	897.0
154.00	863.0
155.00	2072.0
156.10	2780.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

157.10	835.0
158.10	750.0
159.00	412.0
160.00	643.0
161.00	1314.0
162.00	547.0
165.10	1299.0
166.00	606.0
167.10	5785.0
168.10	3799.0
169.10	457.0
170.20	228.0
171.00	212.0
172.00	732.0
173.00	960.0
174.10	1543.0
175.10	2470.0
176.00	826.0
177.00	1133.0
178.10	489.0
179.00	4339.0
180.10	3423.0
181.10	1883.0
181.90	223.0
182.90	223.0
184.10	577.0
185.00	2679.0
186.10	18936.0
187.00	5619.0
188.10	388.0
189.00	1187.0
192.00	1643.0
193.10	1819.0
194.10	653.0
196.00	4555.0
198.00	145664.0
199.00	9412.0
200.10	728.0
201.50	815.0
203.10	959.0
204.10	5083.0
205.10	9510.0
206.10	36928.0
207.10	5335.0
208.10	1099.0
209.00	303.0
210.00	640.0
211.00	1557.0
211.70	420.0
215.00	285.0
216.00	730.0
217.00	9250.0
218.10	1336.0
219.20	228.0
221.10	6356.0
223.00	2200.0
224.10	20312.0
225.10	5565.0
226.10	458.0
227.00	7739.0
228.00	1255.0
229.00	2314.0
231.10	960.0
232.10	218.0
233.10	287.0
234.00	535.0
235.10	503.0
236.00	389.0
237.10	744.0
239.00	507.0
241.10	385.0
242.00	1014.0
243.10	1328.0
244.10	17376.0
245.10	1198.0
246.00	3113.0
247.10	814.0
249.10	656.0
250.10	214.0
251.10	263.0
252.00	170.0
253.10	514.0
255.00	80440.0
256.10	11746.0
257.00	1102.0
258.10	5003.0
259.10	801.0
261.00	203.0
265.00	2023.0
266.10	436.0
269.30	152.0
270.00	195.0
271.10	556.0
272.00	323.0
273.00	2864.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

274.10	6528.0
275.10	38448.0
276.00	5002.0
277.00	3114.0
278.10	388.0
283.00	387.0
284.00	279.0
285.10	618.0
289.00	153.0
290.00	247.0
291.00	392.0
291.90	189.0
293.00	512.0
294.00	338.0
294.30	419.0
295.00	249.0
296.00	9807.0
297.10	1491.0
301.00	177.0
302.10	177.0
303.10	1642.0
304.00	397.0
309.00	160.0
310.00	184.0
313.10	171.0
314.00	593.0
315.00	1367.0
316.00	468.0
321.10	250.0
323.10	3892.0
324.10	610.0
327.00	756.0
327.90	239.0
332.10	295.0
333.00	308.0
334.10	2198.0
335.00	765.0
341.10	337.0
342.10	224.0
346.00	890.0
352.00	1034.0
353.10	1001.0
354.00	1385.0
355.00	257.0
359.10	160.0
365.00	5306.0
366.00	590.0
371.00	252.0
372.10	2519.0
373.00	303.0
377.00	172.0
383.10	332.0
384.00	227.0
390.00	320.0
402.10	728.0
403.00	1094.0
404.10	253.0
414.90	182.0
421.00	918.0
422.10	730.0
423.10	6273.0
424.10	1181.0
438.80	163.0
441.10	21816.0
442.10	122408.0
443.10	24272.0
444.10	2553.0
445.10	285.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13

Scan 2460 (17.791 min): BG026164.D
DFTPP13

m/z	Abundance
40.10	2297.0
44.10	2895.0
49.20	162.0
83.90	169.0
147.10	231.0
167.00	219.0
207.00	460.0
458.80	211.0

Instrument :
BNA_G
ClientSampleId :
DFTPP13