

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031022\
 Data File : BG052628.D
 Acq On : 10 Mar 2022 11:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 10 13:03:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 06:37:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	37.752	5.6	88	0.00
3	Pyridine	40.000	41.104	-2.8	92	0.00
4	n-Nitrosodimethylamine	40.000	46.052	-15.1	103	0.00
5 S	2-Fluorophenol	80.000	83.697	-4.6	95	0.00
6	Aniline	40.000	42.380	-6.0	94	0.00
7 S	Phenol-d6	80.000	85.423	-6.8	97	0.00
8	2-Chlorophenol	40.000	41.672	-4.2	94	0.00
9	Benzaldehyde	40.000	42.431	-6.1	99	0.00
10 C	Phenol	40.000	42.711	-6.8	98	0.00
11	bis(2-Chloroethyl)ether	40.000	41.339	-3.3	94	0.00
12	1,3-Dichlorobenzene	40.000	41.328	-3.3	95	0.00
13 C	1,4-Dichlorobenzene	40.000	41.453	-3.6	95	0.00
14	1,2-Dichlorobenzene	40.000	40.948	-2.4	94	0.00
15	Benzyl Alcohol	40.000	44.587	-11.5	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	44.655	-11.6	103	0.00
17	2-Methylphenol	40.000	42.364	-5.9	96	0.00
18	Hexachloroethane	40.000	41.778	-4.4	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.949	-7.4	98	0.00
20	3+4-Methylphenols	40.000	42.811	-7.0	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	39.766	0.6	96	0.00
23 S	Nitrobenzene-d5	80.000	83.000	-3.8	99	0.00
24	Nitrobenzene	40.000	41.179	-2.9	97	0.00
25	Isophorone	40.000	41.052	-2.6	96	0.00
26 C	2-Nitrophenol	40.000	41.344	-3.4	98	0.00
27	2,4-Dimethylphenol	40.000	41.070	-2.7	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.003	-0.0	95	0.00
29 C	2,4-Dichlorophenol	40.000	41.423	-3.6	95	0.00
30	1,2,4-Trichlorobenzene	40.000	40.324	-0.8	99	0.00
31	Naphthalene	40.000	40.326	-0.8	96	0.00
32	Benzoic acid	40.000	26.293	34.3#	97	0.00
33	4-Chloroaniline	40.000	41.611	-4.0	98	0.00
34 C	Hexachlorobutadiene	40.000	41.129	-2.8	99	0.00
35	Caprolactam	40.000	41.613	-4.0	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.147	-5.4	99	0.00
37	2-Methylnaphthalene	40.000	41.125	-2.8	97	0.00
38	1-Methylnaphthalene	40.000	40.450	-1.1	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.898	-2.2	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.200	-5.5	96	0.00
42 S	2,4,6-Tribromophenol	80.000	86.965	-8.7	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.855	-7.1	99	0.00
44	2,4,5-Trichlorophenol	40.000	42.573	-6.4	100	0.00
45 S	2-Fluorobiphenyl	80.000	81.052	-1.3	98	0.00
46	1,1'-Biphenyl	40.000	40.850	-2.1	97	0.00
47	2-Chloronaphthalene	40.000	39.829	0.4	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031022\
 Data File : BG052628.D
 Acq On : 10 Mar 2022 11:05
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 10 13:03:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 06:37:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.289	-8.2	101	0.00
49	Acenaphthylene	40.000	40.746	-1.9	97	0.00
50	Dimethylphthalate	40.000	40.678	-1.7	97	0.00
51	2,6-Dinitrotoluene	40.000	40.382	-1.0	99	0.00
52 C	Acenaphthene	40.000	41.403	-3.5	99	0.00
53	3-Nitroaniline	40.000	40.959	-2.4	94	0.00
54 P	2,4-Dinitrophenol	40.000	40.136	-0.3	100	0.00
55	Dibenzofuran	40.000	40.599	-1.5	98	0.00
56 P	4-Nitrophenol	40.000	42.795	-7.0	95	0.00
57	2,4-Dinitrotoluene	40.000	42.850	-7.1	102	0.00
58	Fluorene	40.000	40.766	-1.9	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	45.089	-12.7	103	0.00
60	Diethylphthalate	40.000	41.627	-4.1	99	0.00
61	4-Chlorophenyl-phenylether	40.000	41.134	-2.8	99	0.00
62	4-Nitroaniline	40.000	41.753	-4.4	98	0.00
63	Azobenzene	40.000	41.971	-4.9	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.864	-2.2	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.698	0.8	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.779	0.6	100	0.00
68	Hexachlorobenzene	40.000	39.644	0.9	100	0.00
69	Atrazine	40.000	42.170	-5.4	101	0.00
70 C	Pentachlorophenol	40.000	40.834	-2.1	98	0.00
71	Phenanthrene	40.000	39.650	0.9	99	0.00
72	Anthracene	40.000	39.638	0.9	98	0.00
73	Carbazole	40.000	40.193	-0.5	101	0.00
74	Di-n-butylphthalate	40.000	40.394	-1.0	99	0.00
75 C	Fluoranthene	40.000	40.875	-2.2	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	43.521	-8.8	101	0.00
78	Pyrene	40.000	40.208	-0.5	102	0.00
79 S	Terphenyl-d14	80.000	79.878	0.2	101	0.00
80	Butylbenzylphthalate	40.000	40.931	-2.3	101	0.00
81	Benzo(a)anthracene	40.000	40.270	-0.7	101	0.00
82	3,3'-Dichlorobenzidine	40.000	41.179	-2.9	100	0.00
83	Chrysene	40.000	40.917	-2.3	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.476	-1.2	100	0.00
85 c	Di-n-octyl phthalate	40.000	41.413	-3.5	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.438	-3.6	105	0.00
88	Benzo(b)fluoranthene	40.000	40.367	-0.9	103	0.00
89	Benzo(k)fluoranthene	40.000	40.555	-1.4	105	-0.01
90 C	Benzo(a)pyrene	40.000	40.956	-2.4	104	0.00
91	Dibenzo(a,h)anthracene	40.000	41.199	-3.0	105	0.00
92	Benzo(g,h,i)perylene	40.000	41.056	-2.6	104	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031022\
Data File : BG052628.D
Acq On : 10 Mar 2022 11:05
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 10 13:03:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 10 06:37:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0