


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031121\
 Data File : BG048354.D
 Acq On : 11 Mar 2021 10:04
 Operator : CG/JU
 Sample : DFTPP88
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP088

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG031121.M
 Title : SVOA CALIBRATION
 Last Update : Thu Mar 11 15:12:16 2021



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	45240	PASS
68	69	0.00	2	0.5	267	PASS
69	198	0.00	100	43.0	53298	PASS
70	69	0.00	2	0.5	279	PASS
127	198	10	80	42.7	52925	PASS
197	198	0.00	2	0.3	326	PASS
198	198	100	100	100.0	124069	PASS
199	198	5	9	6.9	8521	PASS
275	198	10	60	31.5	39109	PASS
365	198	1	100	4.3	5319	PASS
441	198	0.01	100	12.3	15240	PASS
442	198	50	100	73.8	91610	PASS
443	442	15	24	19.6	17992	PASS

m/z	Abundance
39.00	302.0
39.90	160.0
41.00	424.0
42.00	177.0
43.10	627.0
44.00	3374.0
55.10	392.0
57.10	285.0
58.30	168.0
59.10	559.0
62.80	165.0
67.20	265.0
70.90	252.0
71.10	253.0
73.10	200.0
81.10	227.0
82.10	338.0
82.90	193.0
84.90	209.0
96.00	225.0
97.10	338.0
99.20	154.0
114.90	273.0
117.10	155.0
163.10	165.0
166.80	157.0
193.20	223.0
207.00	909.0
208.10	180.0
208.90	313.0
233.10	302.0
273.80	191.0
281.10	730.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

m/z	Abundance
37.00	329.0
38.10	572.0
39.10	4260.0
40.00	392.0
41.10	405.0
42.10	188.0
43.10	347.0
44.10	3728.0
50.10	16093.0
51.10	57832.0
52.10	2693.0
54.10	186.0
55.10	625.0
56.00	2281.0
57.00	4225.0
58.10	512.0
58.90	365.0
61.00	1238.0
62.00	583.0
63.00	2345.0
64.10	787.0
65.10	1356.0
65.90	246.0
67.10	502.0
69.00	68008.0
70.00	358.0
70.90	229.0
73.20	770.0
74.00	8336.0
75.00	10241.0
76.10	3538.0
77.10	67712.0
78.00	4227.0
79.00	5890.0
80.00	3855.0
81.00	6158.0
82.00	1393.0
83.10	1353.0
83.90	229.0
85.00	918.0
86.00	1867.0
87.10	682.0
87.90	585.0
88.70	194.0
90.90	1117.0
91.90	1186.0
93.10	9083.0
94.10	506.0
95.20	180.0
95.90	492.0
98.10	8131.0
99.00	5638.0
100.00	363.0
101.00	3193.0
101.70	211.0
102.90	722.0
103.20	696.0
104.00	1932.0
105.00	2417.0
107.00	22488.0
108.00	3162.0
110.00	34224.0
111.10	6104.0
112.00	755.0
113.10	299.0
113.90	236.0
116.10	1562.0
117.00	20168.0
118.00	1527.0
119.00	295.0
119.90	257.0
120.20	152.0
122.00	1987.0
123.00	1938.0
124.00	1206.0
125.00	1394.0
127.00	64984.0
128.00	5417.0
129.00	29576.0
129.90	1897.0
131.20	528.0
131.90	289.0
133.00	322.0
134.00	928.0
134.90	2026.0
136.00	815.0
137.00	960.0
137.90	209.0
139.00	246.0
139.90	365.0
141.00	4313.0
142.00	1018.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

143.00	1571.0
144.10	170.0
145.10	279.0
146.00	901.0
147.10	1780.0
148.00	5167.0
149.00	1049.0
150.10	242.0
150.70	283.0
152.10	254.0
152.90	1357.0
154.00	1443.0
155.00	2138.0
156.00	3050.0
157.10	1215.0
158.00	783.0
159.10	550.0
160.00	1359.0
161.00	1473.0
162.00	605.0
163.00	180.0
165.00	1882.0
166.00	1404.0
167.00	6951.0
168.10	3515.0
169.00	642.0
169.80	165.0
170.90	401.0
172.00	812.0
173.00	644.0
174.00	1821.0
175.00	3243.0
175.90	495.0
177.00	1135.0
178.10	824.0
178.90	6871.0
180.00	5148.0
181.00	2634.0
181.90	527.0
182.90	308.0
184.00	572.0
185.00	3229.0
186.00	22928.0
187.00	6681.0
188.00	735.0
189.00	1939.0
190.10	278.0
190.90	600.0
191.90	1979.0
193.00	2613.0
194.10	858.0
195.10	221.0
196.00	3520.0
197.90	142592.0
199.00	10555.0
200.00	634.0
200.70	206.0
201.60	631.0
203.00	1187.0
204.00	7189.0
205.00	12232.0
206.00	43584.0
207.00	9078.0
208.00	1998.0
208.90	921.0
210.20	831.0
211.10	1800.0
211.70	339.0
214.20	286.0
215.00	508.0
216.00	1315.0
217.00	13707.0
217.90	1457.0
219.00	455.0
221.00	5952.0
222.00	1651.0
223.00	3247.0
224.00	25304.0
225.00	7409.0
226.00	883.0
227.00	10738.0
227.90	1602.0
228.90	2563.0
230.00	385.0
231.00	911.0
232.00	295.0
234.00	706.0
235.00	1063.0
236.10	756.0
236.90	891.0
237.90	182.0
238.90	495.0
240.20	334.0
241.00	691.0
242.00	1316.0
242.90	1225.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

244.00	20656.0
245.10	2518.0
246.00	4441.0
246.90	776.0
247.80	489.0
249.10	439.0
250.00	309.0
252.00	323.0
252.90	717.0
255.00	102096.0
256.00	16712.0
257.00	1437.0
258.00	8340.0
258.90	1098.0
260.00	199.0
263.90	310.0
265.00	3465.0
266.00	357.0
267.10	351.0
270.30	173.0
271.10	155.0
272.00	310.0
273.00	3204.0
274.00	7814.0
275.00	40640.0
276.00	4397.0
277.00	3664.0
277.80	490.0
281.00	287.0
282.00	215.0
282.80	441.0
283.20	648.0
284.10	239.0
285.00	1004.0
287.90	182.0
289.10	159.0
289.90	287.0
292.10	199.0
293.00	1105.0
293.90	169.0
295.00	552.0
296.00	13783.0
297.00	2341.0
300.90	365.0
301.90	257.0
303.00	1393.0
304.10	238.0
305.10	156.0
307.90	174.0
311.90	155.0
313.90	491.0
315.00	1441.0
316.20	845.0
316.90	210.0
321.00	630.0
323.00	3591.0
323.90	669.0
326.90	1174.0
327.90	629.0
332.10	378.0
332.90	310.0
334.00	2417.0
334.90	679.0
335.90	288.0
340.90	304.0
341.30	191.0
342.10	200.0
345.90	1024.0
346.70	293.0
352.10	1273.0
353.00	939.0
354.00	1247.0
355.20	539.0
358.80	169.0
365.00	5216.0
365.90	740.0
367.10	255.0
370.90	455.0
372.00	2071.0
372.90	489.0
382.90	382.0
390.00	382.0
391.00	335.0
400.90	242.0
402.10	441.0
403.00	1153.0
404.00	460.0
421.00	785.0
422.00	530.0
423.00	4952.0
424.10	1026.0
424.90	249.0
437.40	236.0
438.50	156.0
441.00	13859.0
442.00	84744.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

443.00	16880.0
444.00	1632.0
445.30	184.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

m/z	Abundance
35.90	196.0
37.10	223.0
38.10	870.0
39.10	4262.0
40.20	448.0
41.10	526.0
43.10	529.0
44.10	3939.0
49.10	499.0
50.10	14633.0
51.10	51408.0
52.10	2927.0
54.90	679.0
56.00	1952.0
57.10	3697.0
58.10	389.0
59.10	278.0
61.10	624.0
62.00	764.0
63.10	2305.0
64.00	439.0
65.00	846.0
65.90	215.0
67.40	258.0
69.00	61552.0
70.10	300.0
71.00	378.0
72.40	304.0
73.00	751.0
74.10	5674.0
75.00	11047.0
76.10	3253.0
77.10	64040.0
78.10	3818.0
79.00	4200.0
80.00	4268.0
81.00	5170.0
81.90	1083.0
83.00	1202.0
84.30	160.0
85.00	779.0
86.10	1358.0
87.00	837.0
88.10	226.0
88.70	172.0
91.10	1265.0
92.10	1555.0
93.00	8364.0
94.10	504.0
96.10	495.0
98.00	6957.0
99.00	5176.0
99.90	299.0
100.90	2623.0
103.10	934.0
104.00	1805.0
105.00	2608.0
106.10	775.0
107.00	21864.0
108.10	3338.0
110.00	30424.0
111.10	5462.0
111.90	599.0
113.20	237.0
114.10	178.0
116.10	1592.0
117.00	17976.0
118.00	1214.0
119.00	473.0
120.10	282.0
120.90	200.0
122.00	1502.0
123.00	2185.0
124.00	1016.0
125.00	1261.0
127.00	60832.0
128.00	5579.0
129.00	29744.0
130.00	2411.0
131.10	453.0
132.60	165.0
133.00	337.0
134.00	902.0
135.00	2158.0
135.90	932.0
137.10	1840.0
138.20	402.0
139.90	580.0
141.00	3565.0
142.00	1132.0
143.00	863.0
144.00	332.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

145.00	380.0
145.90	652.0
147.00	2163.0
148.00	5249.0
148.90	926.0
150.00	413.0
150.90	528.0
151.70	267.0
153.00	1300.0
154.10	887.0
155.00	2190.0
156.10	2828.0
157.00	785.0
158.00	819.0
159.10	533.0
159.90	1261.0
161.00	2267.0
162.00	515.0
162.90	202.0
165.00	1509.0
166.00	1346.0
167.00	7721.0
168.10	2369.0
169.10	955.0
170.00	227.0
171.00	427.0
171.90	523.0
172.90	1034.0
174.00	1692.0
175.00	2696.0
175.90	313.0
176.90	2156.0
178.00	743.0
179.00	5837.0
180.00	4430.0
180.90	1931.0
182.00	263.0
182.90	285.0
183.90	701.0
185.00	3145.0
186.00	22832.0
187.10	7265.0
188.00	388.0
189.00	1664.0
190.00	494.0
191.00	673.0
192.00	1720.0
193.10	2661.0
194.10	421.0
194.90	201.0
196.00	3814.0
197.90	147200.0
199.00	9587.0
200.10	706.0
201.50	671.0
201.90	665.0
202.90	1559.0
204.00	6571.0
205.00	10154.0
206.00	46008.0
207.00	6799.0
208.00	2048.0
209.00	690.0
210.00	613.0
210.90	2027.0
215.00	823.0
215.90	1098.0
217.00	14283.0
218.00	1613.0
219.90	193.0
220.90	6302.0
222.00	1864.0
223.00	3235.0
224.00	28200.0
225.00	6676.0
226.00	830.0
227.00	14253.0
228.00	1582.0
229.00	2984.0
230.00	375.0
230.90	928.0
231.80	371.0
233.10	198.0
234.00	1302.0
234.90	1103.0
236.10	485.0
237.00	1230.0
239.00	562.0
240.00	489.0
241.00	797.0
242.10	1558.0
243.10	2498.0
244.00	19336.0
245.10	2431.0
246.00	5011.0
246.90	1094.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

247.90	268.0
248.90	581.0
250.10	165.0
251.00	374.0
251.90	400.0
253.00	670.0
255.00	112192.0
256.00	14746.0
257.00	897.0
258.00	7788.0
259.00	983.0
261.00	249.0
264.90	3339.0
265.90	243.0
267.30	165.0
269.30	175.0
270.90	289.0
271.70	704.0
273.00	3408.0
274.00	8508.0
275.00	48480.0
276.00	7271.0
276.90	5200.0
277.90	743.0
278.80	375.0
280.90	295.0
283.00	570.0
283.80	318.0
284.10	344.0
285.00	592.0
288.90	324.0
289.90	269.0
292.00	357.0
293.00	973.0
294.00	210.0
294.90	401.0
296.00	16816.0
297.00	1805.0
301.00	235.0
303.10	1405.0
304.00	456.0
308.00	241.0
309.20	217.0
313.00	274.0
314.90	1670.0
316.00	775.0
320.90	462.0
322.00	266.0
323.00	4369.0
324.00	733.0
326.90	1033.0
328.10	419.0
331.80	274.0
332.90	532.0
334.00	4091.0
335.10	869.0
341.00	476.0
345.90	565.0
347.00	185.0
351.00	211.0
352.10	1382.0
353.00	1010.0
354.00	1018.0
355.10	487.0
363.60	203.0
365.00	6532.0
365.90	1274.0
370.10	241.0
370.90	497.0
372.10	2557.0
372.90	483.0
383.00	431.0
384.10	200.0
389.80	296.0
390.10	308.0
390.90	252.0
402.00	859.0
403.10	1067.0
404.00	428.0
421.10	732.0
422.00	1288.0
423.00	7573.0
423.90	1862.0
437.10	162.0
439.50	166.0
441.00	19088.0
442.00	110304.0
443.00	21888.0
444.00	2212.0
444.70	160.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

m/z	Abundance
38.00	246.0
39.00	1908.0
40.00	355.0
41.00	512.0
43.10	650.0
44.00	2851.0
50.10	6784.0
51.10	26480.0
52.10	1280.0
53.20	216.0
55.00	457.0
55.90	1370.0
57.00	2092.0
58.00	256.0
58.90	166.0
61.10	478.0
62.00	160.0
63.00	1164.0
64.10	279.0
65.00	632.0
67.10	255.0
68.00	802.0
69.00	30336.0
69.90	180.0
71.00	331.0
73.10	454.0
74.10	3367.0
75.00	4961.0
76.10	1869.0
77.10	33104.0
78.10	1595.0
79.10	2707.0
80.00	1626.0
81.00	2381.0
82.00	674.0
83.10	488.0
84.40	194.0
85.20	616.0
86.00	691.0
87.00	173.0
91.10	1031.0
92.00	985.0
93.00	4884.0
95.30	205.0
95.60	225.0
95.90	332.0
98.00	3467.0
99.00	3640.0
101.00	1583.0
103.00	726.0
104.00	1544.0
105.10	887.0
106.10	631.0
107.00	11197.0
108.10	1802.0
109.10	537.0
110.00	17608.0
111.10	2688.0
112.00	201.0
113.10	200.0
114.70	280.0
115.90	686.0
116.20	813.0
117.00	10481.0
118.00	779.0
118.80	301.0
121.90	1392.0
123.00	1265.0
123.90	637.0
124.90	498.0
127.00	32960.0
128.00	2205.0
129.00	13319.0
130.10	1529.0
132.00	347.0
132.90	326.0
134.10	435.0
135.00	946.0
136.10	517.0
137.00	663.0
138.80	180.0
139.80	428.0
141.00	2168.0
142.20	479.0
142.90	327.0
143.80	167.0
144.30	204.0
145.90	491.0
147.00	1099.0
148.00	2273.0
149.00	540.0
151.10	311.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

151.60	254.0
153.10	680.0
154.00	528.0
155.10	1126.0
156.10	1017.0
158.10	463.0
159.10	254.0
159.90	734.0
161.00	1071.0
161.80	309.0
162.90	278.0
164.10	209.0
164.90	974.0
166.10	692.0
167.00	4392.0
168.10	1126.0
169.00	507.0
171.00	329.0
171.90	538.0
173.00	588.0
174.00	1175.0
175.00	2332.0
176.10	561.0
177.00	657.0
177.90	587.0
179.00	4020.0
180.00	2262.0
180.90	1348.0
182.00	198.0
184.00	317.0
185.00	1951.0
186.00	13989.0
187.00	4181.0
188.00	400.0
188.90	1220.0
190.80	494.0
191.90	882.0
193.00	1252.0
194.10	208.0
195.20	295.0
196.00	1995.0
196.70	980.0
198.00	82416.0
198.90	5421.0
199.80	430.0
200.80	231.0
201.30	677.0
203.00	977.0
203.90	3213.0
205.00	7844.0
206.00	26776.0
207.00	4424.0
208.00	1087.0
209.00	381.0
209.70	246.0
211.00	1271.0
211.70	155.0
215.00	421.0
216.00	820.0
217.00	8991.0
218.00	580.0
220.90	3031.0
221.90	1249.0
223.10	2269.0
224.00	16720.0
225.10	4226.0
227.00	8885.0
228.00	839.0
229.00	1579.0
230.00	296.0
230.90	602.0
231.90	228.0
232.60	187.0
233.90	399.0
235.10	452.0
236.10	396.0
237.10	547.0
239.00	552.0
240.00	342.0
240.80	445.0
241.90	751.0
243.00	945.0
244.00	11679.0
245.00	1875.0
246.00	2033.0
247.00	452.0
249.00	434.0
251.20	191.0
253.00	337.0
253.90	1100.0
255.00	67648.0
256.00	10119.0
256.90	713.0
257.90	4601.0
259.10	445.0
260.00	189.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088

260.70	192.0
264.10	187.0
265.00	1937.0
267.10	181.0
270.00	181.0
271.00	268.0
272.20	192.0
273.10	2529.0
274.10	4317.0
275.00	28208.0
275.90	3552.0
277.00	3104.0
278.00	370.0
281.80	158.0
282.10	172.0
282.40	195.0
283.10	384.0
283.90	453.0
285.00	464.0
288.90	282.0
289.90	312.0
292.00	244.0
293.10	958.0
294.10	174.0
295.00	717.0
296.00	8998.0
297.00	1161.0
301.00	193.0
302.00	247.0
303.00	1372.0
304.10	445.0
314.20	443.0
315.00	1533.0
316.00	354.0
317.00	212.0
321.00	445.0
323.00	2566.0
324.20	351.0
325.70	172.0
327.00	567.0
327.80	169.0
331.70	388.0
333.00	469.0
334.00	2099.0
335.00	555.0
341.00	300.0
346.00	751.0
351.90	1019.0
353.00	466.0
354.00	1296.0
354.80	346.0
365.00	4210.0
365.80	410.0
366.10	302.0
372.00	1973.0
373.00	289.0
382.90	444.0
385.00	151.0
389.80	201.0
401.30	155.0
402.00	493.0
402.90	766.0
404.20	355.0
421.00	625.0
422.00	423.0
423.00	4684.0
424.10	773.0
437.50	210.0
439.40	151.0
441.00	12774.0
442.00	79784.0
443.00	15209.0
444.10	1132.0

Instrument :
BNA_G
ClientSampleId :
DFTPP088