

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031125\
 Data File : BG064100.D
 Acq On : 11 Mar 2025 17:59
 Operator : RC/JU
 Sample : Q1523-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-A2-02-C

Manual Integrations
 APPROVED

Reviewed By :Anahy Claudio 03/12/2025
 Supervised By :Jagrut Upadhyay 03/12/2025

Quant Time: Mar 12 06:34:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 15:39:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.865	152	37349	20.000	ng	0.00
21) Naphthalene-d8	10.650	136	156052	20.000	ng	0.00
39) Acenaphthene-d10	14.486	164	103773	20.000	ng	0.00
64) Phenanthrene-d10	17.224	188	216840	20.000	ng	0.00
76) Chrysene-d12	21.455	240	225564	20.000	ng	0.00
86) Perylene-d12	24.469	264	244851	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.456	112	311767	130.340	ng	0.00
7) Phenol-d6	7.036	99	379747	116.702	ng	0.00
23) Nitrobenzene-d5	9.016	82	311873	110.442	ng	0.00
42) 2,4,6-Tribromophenol	15.973	330	183855	159.387	ng	0.00
45) 2-Fluorobiphenyl	13.112	172	644323	94.245	ng	0.00
79) Terphenyl-d14	19.851	244	1092452	97.929	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	10.703	128	134959	16.038	ng	97
37) 2-Methylnaphthalene	12.307	142	51251	8.627	ng	97
38) 1-Methylnaphthalene	12.524	142	41327m	7.101	ng	
52) Acenaphthene	14.545	154	12450	2.051	ng	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031125\
Data File : BG064100.D
Acq On : 11 Mar 2025 17:59
Operator : RC/JU
Sample : Q1523-06
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-A2-02-C

Manual Integrations
APPROVED

Reviewed By :Anahy Claudio 03/12/2025
Supervised By :Jagrut Upadhyay 03/12/2025

Quant Time: Mar 12 06:34:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 05 15:39:19 2025
Response via : Initial Calibration

