

Data Path : Z:\HPCHEM1\BNA G\DATA\BG031218\
 Data File : BG033103.D
 Acq On : 13 Mar 2018 00:21
 Operator : SJ/JU
 Sample : J1867-10
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CPS073071

Quant Time: Mar 13 04:23:25 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 09 15:10:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.91	152	40276	20.00	ng	0.00
21) Naphthalene-d8	11.79	136	181872	20.00	ng	0.00
38) Acenaphthene-d10	15.52	164	105384	20.00	ng	0.00
63) Phenanthrene-d10	18.25	188	251541	20.00	ng	0.00
75) Chrysene-d12	22.60	240	241134	20.00	ng	0.00
86) Perylene-d12	26.40	264	229130	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	6.33	112	173711	69.00	ng	0.00
7) Phenol-d6	8.00	99	139591	39.78	ng	0.00
23) Nitrobenzene-d5	10.11	82	347721	127.12	ng	0.00
41) 2,4,6-Tribromophenol	16.99	330	212310	191.60	ng	0.00
44) 2-Fluorobiphenyl	14.15	172	768180	105.13	ng	0.00
78) Terphenyl-d14	20.77	244	1197762	111.60	ng	0.00
Target Compounds						
35) Caprolactam	12.65	113	8783	8.715	ng	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG031218\
Data File : BG033103.D
Acq On : 13 Mar 2018 00:21
Operator : SJ/JU
Sample : J1867-10
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
CPS073071

Quant Time: Mar 13 04:23:25 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 09 15:10:14 2018
Response via : Initial Calibration

